

## German Accreditation Body

### Annex to the partial accreditation certificate D-K-15070-01-01 according to DIN EN ISO/IEC 17025:2018

**Valid from:** 19.04.2023

**Date of issue:** 19.04.2023

This document attachment is part of the accreditation certificate D-K-15070-01-00.

Partial Accreditation Certificate Holder:

**Testo Industrial Services GmbH**  
**Gewerbestraße 3, 79199 Kirchzarten,**  
**Germany**

**The German original version**  
**„Anlage zur Akkreditierungsurkunde D-K-15070-01-01 nach DIN EN ISO/IEC 17025:2018“**  
**is valid.**

The calibration laboratory meets the requirements according to DIN EN ISO/IEC 17025:2018 to perform the conformity assessment activities listed in this annex. The calibration laboratory shall comply with additional legal and normative requirements, where applicable, including those in relevant sectoral programs, provided that they are explicitly confirmed below.

The requirements for the management system in DIN EN ISO/IEC 17025 are written in a language relevant for calibration laboratories and are overall in accordance with the principles of DIN EN ISO 9001.

Other locations:

**Calibration laboratory Kirchzarten, Erich-Rieder Straße 4, 79199 Kirchzarten**  
**Calibration laboratory Munich, Nikolaus-Otto-Straße 2, 85221 Dachau**  
**Calibration laboratory Essen, Alte Landstraße 3c, 45329 Essen**  
**Calibration laboratory Hamburg, Meiendorfer Straße 205, 22145 Hamburg**  
**Calibration laboratory Mörfelden-Walldorf, Kurhessenstraße 11, 64546 Mörfelden-Walldorf**  
**Calibration laboratory Winsen, Tönnhäuser Weg 100-106, 21423 Winsen (Luhe)**

*This deed annex is valid only together with the deed issued in writing and reflects the status at the date of issue. The current status of valid and monitored accreditation can be found in the database of accredited bodies of the German Accreditation Body ([www.dakks.de](http://www.dakks.de))*

**Electrical measurands**

**DC and low frequency measurands**

- DC voltage <sup>a), b)</sup>
- AC voltage <sup>a), b)</sup>
- DC current strength <sup>a), b)</sup>
- AC current strength <sup>a), b)</sup>
- AC/DC transfer
- Electrical power <sup>a), b)</sup>
- Phase angle
- DC resistance <sup>a), b)</sup>
- AC resistance
- Capacity <sup>a), b)</sup>
- Inductance
- Tension ratio <sup>a), b)</sup>
- High voltage measurands <sup>a)</sup>

**Time and frequency**

- Time interval <sup>a), b)</sup>
- Frequency and speed <sup>a), b)</sup>

**High-frequency measured variables**

- RF impedance (reflection coefficient) <sup>a), b)</sup>
- RF power <sup>a), b)</sup>
- RF attenuation <sup>a), b)</sup>
- RF noise
- Modulation measurands
- Oscilloscope measured variables <sup>a), b)</sup>
- Rise time <sup>a), b)</sup>
- Bandwidth <sup>a), b)</sup>
- Pulsed measurands <sup>a), b)</sup>

**Dimensional measurands**

**Length**

- Diameter <sup>a), b)</sup>
- Thread <sup>a), b)</sup>
- Parallel gauge blocks
- Linear encoders <sup>a), b)</sup>
- length measuring equipment
- Line dimensions, distances

**Angle**

- Inclinometers
- Angle of rotation

**Coordinate measuring technology**

- Coordinate measuring machines <sup>c)</sup>

**a) also on-site calibration**

**b) Mobile laboratory**

**c) On-site calibration only**

Within the measurands/calibration items marked with <sup>\*</sup>, the calibration laboratory is permitted to apply the standards/calibration guidelines listed here with different editions without requiring prior information and approval by DAkkS. The calibration laboratory has an up-to-date list of all standards/calibration guidelines in the flexible accreditation area.

Location Calibration laboratory Kirchzarten, Erich-Rieder Straße 4, 79199 Kirchzarten

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range		Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC voltage	0 V			0.1 $\mu$ V	$U =$ measured value
Measuring devices	1 $\mu$ V	to 220 mV		$7.5 \cdot 10^{-6} U$	
	> 220 mV	to 2,2 V		$5 \cdot 10^{-6} U$	
	> 2,2 V	to 11 V		$3.5 \cdot 10^{-6} U$	
	> 11 V	to 22 V		$3.5 \cdot 10^{-6} U$	
	> 22 V	to 220 V		$5 \cdot 10^{-6} U$	
	> 220 V	to 1000 V		$6.5 \cdot 10^{-6} U$	
Sources	0 V			0.1 $\mu$ V	
	1 $\mu$ V	to 200 mV		$5.8 \cdot 10^{-6} U$	
	> 200 mV	to 2 V		$2.9 \cdot 10^{-6} U$	
	> 2 V	to 20 V		$2.9 \cdot 10^{-6} U$	
	> 20 V	to 200 V		$4 \cdot 10^{-6} U$	
	> 200 V	to 1000 V		$4.5 \cdot 10^{-6} U$	
High Voltage	> 1 kV	to 50 kV		$0.4 \cdot 10^{-3} U$	
Direct current strength	0 A			0.2 nA	$I =$ measured value
Sources and	0.1 $\mu$ A	to < 1 $\mu$ A		$35 \cdot 10^{-6} I$	
Measuring devices	1 $\mu$ A	to < 10 $\mu$ A		$50 \cdot 10^{-6} I$	
	10 $\mu$ A	to < 100 $\mu$ A		$15 \cdot 10^{-6} I$	
	100 $\mu$ A	to 320 mA		$18 \cdot 10^{-6} I$	
	> 320 mA	to 1 A		$15 \cdot 10^{-6} I$	
	> 1 A	to 10 A		$0.2 \cdot 10^{-3} I$	
	> 10 A	to 1000 A		$0.3 \cdot 10^{-3} I$	
Current clamps	1 mA	to 2,2 A		$1 \cdot 10^{-3} I$	
	> 2,2 A	to 20 A		$2 \cdot 10^{-3} I$	
	> 20 A	to 1000 A		$3 \cdot 10^{-3} I$	
Current transformer	1 A to 120 A			$0.22 \cdot 10^{-3} I$	
	> 120 A to 1000 A			$0.25 \cdot 10^{-3} I$	
Direct current- resistance	0 $\Omega$	to < 1 m $\Omega$		1,3 $\mu\Omega$	$R =$ measured value
	1 m $\Omega$	to 10 m $\Omega$		$7 \cdot 10^{-6} R$	
	> 10 m $\Omega$	to 10 M $\Omega$		$4 \cdot 10^{-6} R$	
	> 10 M $\Omega$	to 100 M $\Omega$		$8 \cdot 10^{-6} R$	
	> 100 M $\Omega$	to 10 G $\Omega$		$0.17 \cdot 10^{-3} R$	
	> 10 G $\Omega$	to 100 G $\Omega$		$0.21 \cdot 10^{-3} R$	
	> 100 G $\Omega$	to 1 T $\Omega$		$0.35 \cdot 10^{-3} R$	
	> 1 T $\Omega$	to 10 T $\Omega$		$0.6 \cdot 10^{-3} R$	
	> 10 T $\Omega$	to 100 T $\Omega$		$29 \cdot 10^{-3} R$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Extended uncertainty of measurement	Remarks	
	Measuring range		Measuring conditions / Procedures			
DC power meters	1 mW	to	< 220 W	1 mV to 0.22 V 1 A to 1000 A	$0.6 \cdot 10^{-3} P$	
	1 mW	to	320 W	0.22 V to 1000 V 0.1 mA to 0.32 A	$20 \cdot 10^{-6} P$	
	> 70 mW	to	1 kW	0.22 V to 1000 V > 0.32 A to 1 A	$35 \cdot 10^{-6} P$	
	> 0,22 W	to	10 kW	0.22 V to 1000 V > 1 A to 10 A	$0.2 \cdot 10^{-3} P$	
	> 2,2 W	to	1 MW	0.22 V to 1000 V > 10 A to 1000 A	$0.3 \cdot 10^{-3} P$	
DC power sources	1 mW	to	< 200 W	1 mV to 0.2 V 1 A to 1000 A	$0.4 \cdot 10^{-3} P$	
	1 mW	to	320 W	0.2 V to 1000 V 0.1 mA to 0.32 A	$20 \cdot 10^{-6} P$	
	> 64 mW	to	1 kW	0.2 V to 1000 V > 0.32 A to 1 A	$35 \cdot 10^{-6} P$	
	> 0,2 W	to	10 kW	0.2 V to 1000 V > 1 A to 10 A	$0.2 \cdot 10^{-3} P$	
	> 2 W	to	1 MW	0.2 V to 1000 V > 10 A to 1000 A	$0.3 \cdot 10^{-3} P$	
AC voltage Measuring devices and Sources	1 mV	to	2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	<i>U</i> = measured value
				> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
				> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
				> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
				> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
				> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
				> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
				> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV	to	7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
				> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
			> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$		
			> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$		
			> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$		
			> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$		
			> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$		
			> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and Sources	> 7 mV to 22 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$80 \cdot 10^{-6} U$ $80 \cdot 10^{-6} U$ $65 \cdot 10^{-6} U$ $75 \cdot 10^{-6} U$ $75 \cdot 10^{-6} U$ $95 \cdot 10^{-6} U$ $0.19 \cdot 10^{-3} U$ $0.21 \cdot 10^{-3} U$	<i>U</i> = measured value
	> 22 mV to 70 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$70 \cdot 10^{-6} U$ $58 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $55 \cdot 10^{-6} U$ $0.11 \cdot 10^{-3} U$ $0.13 \cdot 10^{-3} U$	
	> 70 mV to 220 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$39 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $28 \cdot 10^{-6} U$ $42 \cdot 10^{-6} U$ $85 \cdot 10^{-6} U$ $0.1 \cdot 10^{-3} U$	
	> 220 mV to 700 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$25 \cdot 10^{-6} U$ $22 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $27 \cdot 10^{-6} U$ $40 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage Measuring devices and Sources	> 700 mV to 2,2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
> 2,2 V to 7 V	> 2,2 V to 7 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
> 7 V to 22 V	> 7 V to 22 V	10 Hz to 20 Hz	$17 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
> 22 V to 70 V	> 22 V to 70 V	> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
		> 500 kHz to 1MHz	$0.11 \cdot 10^{-3} U$	
		> 10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
> 70 V to 220 V	> 70 V to 220 V	> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$	
> 500 kHz to 1 MHz	> 500 kHz to 1 MHz	> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
		> 10 Hz to 20 Hz	$19 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$18 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$17 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$17 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz	> 50 kHz to 100 kHz	> 50 kHz to 100 kHz	$32 \cdot 10^{-6} U$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage Measuring devices and Sources	> 220 V to 1000 V	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	$U =$ measured value
		> 20 Hz to 40 Hz	$27 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$45 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$45 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$65 \cdot 10^{-6} U$	
High Voltage	> 1 kV to 30 kV > 30 kV to 50 kV	50 Hz	$0.5 \cdot 10^{-3} U$	
		50 Hz	$0.6 \cdot 10^{-3} U$	
AC power Sources and Measuring devices (areas)	100 $\mu$ A to 1 mA	10 Hz to 40 Hz	$120 \cdot 10^{-6} I$	$I =$ measured value
		> 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$160 \cdot 10^{-6} I$ $60 \cdot 10^{-6} I$	
	> 1 mA to 10 mA	10 Hz to 40 Hz	$46 \cdot 10^{-6} I$	
	> 10 mA to 1 A	10 Hz to 40 Hz	$17 \cdot 10^{-6} I$	
		> 40 Hz to 1 kHz; > 1 kHz to 10 kHz;		
	> 1 A to 10 A	10 Hz to 40 Hz	$32 \cdot 10^{-6} I$	
	> 10 A to 20 A	> 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$39 \cdot 10^{-6} I$	
10 Hz to 40 Hz		$69 \cdot 10^{-6} I$ $69 \cdot 10^{-6} I$ $0.17 \cdot 10^{-3} I$		
Current clamps	1 mA to 2,2 A	40 Hz to 5 kHz	$2 \cdot 10^{-3} I$	$I =$ measured value
	> 2,2 A to 20 A	40 Hz to 5 kHz	$3 \cdot 10^{-3} I$	
	> 20 A to 800 A	40 Hz to 65 Hz	$4 \cdot 10^{-3} I$	
Current transformer	1 A to 120 A	40 Hz to 850 Hz	$0.16 \cdot 10^{-3} I$	
	1 A to 120 A	> 850 Hz to 2 kHz	$0.47 \cdot 10^{-3} I$	
	> 120 A to 600 A	40 Hz to 400 Hz	$0.52 \cdot 10^{-3} I$	
	> 120 A to 1000 A	40 Hz to 65 Hz	$0.6 \cdot 10^{-3} I$	
Alternating current- resistance Areas	0,1 $\Omega$ to < 0,316 $\Omega$	10 Hz to < 20 Hz	$0,2 - 10^{-3} R$	$R =$ measured value
		20 Hz to 500 Hz	$50 - 10^{-6} R$	
	> 500 Hz to 1 kHz	$0,2 - 10^{-3} R$		
	0,316 $\Omega$ to 3,16 $\Omega$	10 Hz to 30 Hz	$30 - 10^{-6} R$	
> 30 Hz to 55 Hz		$20 - 10^{-6} R$		
> 55 Hz to 1 kHz		$30 - 10^{-6} R$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Extended uncertainty of measurement	Remarks
	Measuring range	Measuring conditions / Procedures			
Alternating current resistor Areas	> 3,16 Ω to 10 Ω	10 Hz to 30 Hz > 30 Hz to 55 Hz > 55 Hz to 1 kHz		25 - 10 <sup>-6</sup> R 40 μΩ 9 - 10 <sup>-6</sup> R 60 μΩ 25 - 10 <sup>-6</sup> R 40 μΩ	
Cargo Charge amplifiers, Charge meters	10 pC to 10 <sub>s</sub> pC	50 Hz to 10 kHz > 10 kHz to 20 kHz > 20 kHz to 50 kHz		0,4 % 0,6 % 1,0 %	
Phase angle  Between current and voltage measuring devices	-180° to 180°	10 V to 1000 V 0.1 A to 50 A 45 Hz to 65 Hz > 65 Hz to 180 Hz > 180 Hz to 450 Hz > 450 Hz to 850 Hz > 850 Hz to 3 kHz > 3 kHz to 6 kHz		0,0051° 0,0075° 0,018° 0,033° 0,12° 0,23°	
		10 V to 1000 V > 50 A to 80 A 45 Hz to 65 Hz > 65 Hz to 180 Hz > 180 Hz to 450 Hz > 450 Hz to 850 Hz > 850 Hz to 3 kHz		0,0052° 0,0083° 0,025° 0,05° 0,25°	
		10 V to 1000 V > 80 A to 120 A 45 Hz to 65 Hz > 65 Hz to 180 Hz > 180 Hz to 450 Hz > 450 Hz to 850 Hz > 850 Hz to 3 kHz > 3 kHz to 6 kHz		0,0055° 0,0091° 0,020° 0,035° 0,25° 0,5°	
Phase angle between voltages  Measuring devices	-180° to 180°	10 V to 1000 V 45 Hz to 65 Hz > 65 Hz to 180 Hz > 180 Hz to 450 Hz > 450 Hz to 850 Hz > 850 Hz to 3 kHz > 3 kHz to 6 kHz		0,006° 0,0075° 0,025° 0,043° 0,15° 0,3°	



Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between tensions Sources	-180° to 180	0.05 V to 2 V Measuring frequency: 40 Hz to 65 Hz 0.05 V to 0.2 V Measuring frequency: > 65 Hz to 1 kHz > 1 kHz to 2 kHz > 0.2 V to 2 V Measuring frequency: > 65 Hz to 1 kHz > 1 kHz to 2 kHz	0,015°  0,03° 0,055°  0,02° 0,035°	
Electrical power Alternating current effect-performance	5 mW up to 50 kW > 2.5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz $\pm 0.05 \leq \cos \phi_F \leq \pm 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$  However, not smaller than 80 - 10 <sup>-6</sup> 0,14 - 10 <sup>-3</sup>	w: Uncertainty of the amplitude of the voltage fundamental w(I <sub>F</sub> ): Uncertainty of the amplitude of the current fundamentals w(Φ <sub>F</sub> ): Uncertainty of the phase shift angle w(U <sub>rms</sub> ): Uncertainty of the stress rms value w(I <sub>rms</sub> ): Uncertainty of the current rms value
AC blind- performance	5 mvar to 50 kvar > 2.5 var up to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz $\pm 0.05 \leq \sin \phi_F \leq \pm 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$  However, not smaller than 80 - 10 <sup>-6</sup> 0,14 - 10 <sup>-3</sup>	
Apparent power	0.1 VA up to 50 kVA > 50 VA to 120 kVA	1 V to 1000 V 45 Hz to 65 Hz 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$  However, not smaller than 80 - 10 <sup>-6</sup> 0,14 - 10 <sup>-3</sup>	
Stress ratio	± 2 mV/V	Bridge voltage: 5 V  Measuring frequency 225 Hz Measuring frequency 600 Hz Measuring frequency 4.8 kHz	0.04 μV/V 0.05 μV/V 0.12 μV/V	Calibration of 350 Ω bridge standards and the associated indicators
	± 2 mV/V	Bridge voltage: 2,5 V  Measuring frequency 225 Hz Measuring frequency 600 Hz Measuring frequency 4.8 kHz	0.04 μV/V 0.04 μV/V 0.12 μV/V	at discrete points In 10 % steps
	± 5 mV/V	Bridge voltage: 5 V  Measuring frequency 225 Hz Measuring frequency 4.8 kHz	0.06 μV/V 0.22 μV/V	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
Stress ratio	± 10 mV/V	Bridge voltage: 5 V			
		Measuring frequency 225 Hz Measuring frequency 4.8 kHz	0.06 µV/V 0.45 µV/V		
	± 5 mV/V	Bridge voltage: 2,5 V			
		Measuring frequency 225 Hz Measuring frequency 600 Hz Measuring frequency 4.8 kHz	0.06 µV/V 0.06 µV/V 0.22 µV/V		
		Bridge voltage: 2,5 V			
	± 10 mV/V	Measuring frequency 225 Hz Measuring frequency 600 Hz Measuring frequency 4.8 kHz	0.06 µV/V 0.10 µV/V 0.45 µV/V		Calibrate 350 Ω Bridge standards and the associated display devices
		Bridge voltage: 1 V			at discrete points In 10 % steps
	± 10 mV/V	Measuring frequency 600 Hz	0.11 µV/V		
± 20 mV/V	Bridge voltage: 1 V				
	Measuring frequency 4.8 kHz	0.6 µV/V			
± 100 mV/V	Bridge voltage: 1 V				
	Measuring frequency 4.8 kHz	3.5 µV/V			
± 100 mV/V	Bridge voltage: 2,5 V				
	Measuring frequency 4.8 kHz	4.0 µV/V			
Stress ratio DC voltage, bridge standards	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:	2.0 µV/V		
			2.5 µV/V		
		0,5 V	2.5 µV/V		
			2.5 µV/V		
			2.5 µV/V		
			2.5 µV/V		
			2.5 µV/V		
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:	1.0 µV/V		
			2.0 µV/V		
		1 V	2.0 µV/V		
			2.0 µV/V		
			2.0 µV/V		
			2.0 µV/V		
			2.0 µV/V		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Stress ratio DC voltage, Bridge standards	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  2,5 V	0.5 µV/V 0.5 µV/V 0.5 µV/V 0.5 µV/V 0.5 µV/V 1.5 µV/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  5 V	0.30 µV/V 0.25 µV/V 0.25 µV/V 0.25 µV/V 0.35 µV/V 1.5 µV/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  7,5 V	0.20 µV/V 0.20 µV/V 0.20 µV/V 0.20 µV/V 0.3 µV/V 1.5 µV/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  10 V	0.10 µV/V 0.15 µV/V 0.15 µV/V 0.20 µV/V 0.3 µV/V 1.5 µV/V	
Stress ratio DC voltage Bridges, gauges, Measuring amplifier	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  0,5 V	0.35 µV/V 0.35 µV/V 0.40 µV/V 0.55 µV/V 2.5 µV/V	
	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  1 V	0.20 µV/V 0.20 µV/V 0.3 µV/V 0.5 µV/V 2.5 µV/V	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Stress ratio		Bridge voltage:		
DC voltage	-2 mV/V to 2 mV/V		0.10 $\mu$ V/V	
Bridges, gauges,	-5 mV/V to 5 mV/V		0.15 $\mu$ V/V	
Measuring amplifier	-10 mV/V to 10 mV/V	2.5 V; 5 V; 7.5 V; 10 V	0.25 $\mu$ V/V	
	-20 mV/V to 20 mV/V		0.45 $\mu$ V/V	
	-100 mV/V to 100 mV/V		2.5 $\mu$ V/V	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Frequency	10 MHz	Measuring time > 30 min	$1 \cdot 10^{-11} f$	$f$ = current measured value
Frequency measurement	1 mHz to 46 GHz	Measuring time > 5 min	$1 \cdot 10^{-10} f$	
Frequency synthesis	1 mHz to 50 GHz		$1 \cdot 10^{-10} f$	
Time interval	1 ns to 1000 s		$1 \cdot 10^{-10} f$ not smaller than 1 ns	Trigger uncertainties must be taken into account
Speed Optical	$1 \text{ min}^{-1}$ up to $2 \cdot 10^5 \text{ min}^{-1}$	with light pulse generator	$6 \cdot 10^{-6}$ not less than $0.001 \text{ min}^{-1}$	
Mechanical	$1 \text{ min}^{-1}$ up to $10000 \text{ min}^{-1}$		$4 \cdot 10^{-4}$ not less than $0.01 \text{ min}^{-1}$	
Capacity Gauges	1 nF to 100 nF	50 Hz to 10 kHz	$1.0 \cdot 10^{-3} C$	C: measured value with normal capacitances
	> 100 nF to 1000 nF	50 Hz to 1 kHz	$1.0 \cdot 10^{-3} C$	
		> 1 kHz to 10 kHz	$2.5 \cdot 10^{-3} C$	
	190 pF to < 400 pF	10 Hz to 10 kHz	$4 \cdot 10^{-3} C$ 8 pF	with 5520A / 5522A
	400 pF to < 1.1 nF	10 Hz to 10 kHz	$4.5 \cdot 10^{-3} C$ 8 pF	
	1.1 nF to < 3.3 nF	10 Hz to 3 kHz	$4.0 \cdot 10^{-3} C$ 8 pF	
	3.3 nF to < 11 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 8 pF	
	11 nF to < 33 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	33 nF to < 110 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	110 nF to < 330 nF	10 Hz to 1 kHz	$4.5 \cdot 10^{-3} C$	
	330 nF to < 1.1 µF	10 Hz to 600 Hz	$4.5 \cdot 10^{-3} C$	
	1.1 µF to < 3.3 µF	10 Hz to 300 Hz	$4.5 \cdot 10^{-3} C$	
	3.3 µF to < 11 µF	10 Hz to 150 Hz	$4.5 \cdot 10^{-3} C$	
	11 µF to < 33 µF	10 Hz to 120 Hz	$6.0 \cdot 10^{-3} C$	
	33 µF to < 110 µF	10 Hz to 80 Hz	$6.5 \cdot 10^{-3} C$	
	110 µF to < 330 µF	DC to 50 Hz	$6.0 \cdot 10^{-3} C$	
	330 µF to < 1.1 mF	DC to 20 Hz	$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	DC to 6 Hz	$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 2 Hz	$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF	DC to 0.6 Hz	$8.0 \cdot 10^{-3} C$	
	33 mF to 110 mF	DC to 0.2 Hz	$11 \cdot 10^{-3} C$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - High-frequency measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Oscilloscope measurement quantities		Square wave voltage 10 Hz to 10 kHz		
Deflection vertical	1 mV to 5V	$R_i = 50 \Omega$	0,35 %	$R_i$ = internal resistance
Deflection horizontal	1 mV to 120 V	$R_i = 1 M\Omega$ Time markers or sines	0,35 %	
	50 ps to < 1 $\mu$ s	< 1 V	6 ps	
	1 $\mu$ s to 5 s		$1.5 \cdot 10^{-3} t$	
Rise time $t_r$	25 ps to 245 hp	500 mV	9 ps	$t_r$ = natural rise time of the oscilloscope
	> 245 hp to 10 ms	250 mV > 250 mv to 2.5 V	$35 \cdot 10^{-3} t_r$	
	150 ps to 10 ms		$35 \cdot 10^{-3} t_r 5 \text{ ps}$	
	250 ps to 10 ms		$35 \cdot 10^{-3} t_r 8 \text{ ps}$	
Bandwidth $B$	$f_c$ 50 MHz to 26.5 GHz	0.2 V to 2 V $R_i = 50 \Omega$	$3 \% f_c$	$f_c$ = frequency -3dB point
RF attenuation	0 dB to 60 dB	300 kHz to 6 GHz	0.3 dB	Connector 50 $\Omega$ : N50
RF transmission Phase angle $\phi$	-180° to 180°	300 kHz to 6 GHz	$U_T - 180^\circ / \pi K - f$	$U_T = \arcsin\left(\frac{10U}{20 - 1}\right) K$ 0.1°/GHz $U$ : Uncertainty of the Attenuation in dB
RF impedance Reflection coefficient Amount $ \Gamma $	0 to 1	300 kHz to 2 GHz > 2 GHz to 6 GHz	$0,005 \quad 0,005 r^2$ $0,009 \quad 0,005 r^2$	Connector 50 $\Omega$ : N50
Phase angle $\phi$	-180° to 180°	$0,1 <  \Gamma  < 1$ 300 kHz to 6 GHz	$\arcsin \frac{U_T}{ T } \frac{180^\circ}{n}$	Connector 50 $\Omega$ : N50
HF power	0.1 mW to 10 mW	9 kHz to 50 MHz > 50 MHz to 5 GHz > 5 GHz to 18 GHz > 18 GHz to 26.5 GHz	$17 \cdot 10^{-3} P$ $22 \cdot 10^{-3} P$ $30 \cdot 10^{-3} P$ $40 \cdot 10^{-3} P$	Connector 50 $\Omega$ : N50; PC-3.5
Amplitude modulation Modulation depth $m$	0 to 1,0	$f_{MOD} < 1 \text{ MHz}$	0,025 $m$ 0,004	$f_{HF}$ = carrier frequency $f_{HF} < 4 \text{ GHz}$ $f_{MOD}$ = modulation frequency $\Delta f$ = frequency deviation
Frequency modulation Frequency deviation $\Delta f$	0 to 5 MHz		0.041 $\Delta f$ 25 Hz	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - High-frequency measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase modulation Phase deviation $\Delta\Phi$	0 to (4 MHz / $f_{MOD}$ ) rad		$0.041\Delta\Phi$ 0.025 rad	$\Delta\Phi$ = phase deviation
Total Harmonic Distortion <i>THD</i> / distortion factor	0 to 0,3	100 Hz to 50 kHz	0.0165 <i>THD</i> 0.0001	
<i>THDAudio</i>	0 to 0,3	100 kHz to 2 GHz	0.0675 <i>THD</i> 0.0001	
	0 to 0,3	100 Hz to 50 kHz	0.007 <i>THDAudio</i> 0.001	
Flicker*) $\Delta U / U$	0.4 to 5	DIN EN 61000-4-15:2011  (115 V, 60 Hz); (230 V 50 Hz)  (115 V, 60 Hz); (230 V 50 Hz)	$7 \cdot 10^{-3} \Delta U / U$	
Frequency	0.0083 Hz to 40 Hz		$3 \cdot 10^{-3} \Delta U / U$	
$P_{st}$ (Short Term)	10 minutes		0,5 %	
$P_{lt}$ (Long Term)	2 hours		1,7 %	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Dimensional measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Length *) cylindrical setting standards, ring gauges:				
Diameter	1 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	d = is the measured Diameter
Plug gauges: Diameter	1 mm to 200 mm	Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Test pins: Diameter	0.1 mm to 30 mm	VDI/VDE/DGQ 2618 Sheet 4.2:2007 Point 3.2.2 (Opt. 1)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Thread gauges (single and multiple cylindrical External and internal threads with straight flanks, symmetrical profile) Threaded mandrels: simple Flank diameter	1.4 mm to 200 mm nominal pitch: 0.3 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.8:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Three-wire method d = is the measured diameter
Threaded rings: simple pitch diameter	3 mm to 200 mm nominal pitch: 0.5 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.9:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Two-ball method d = is the measured diameter
Threaded mandrels: simple pitch diameter	1.4 mm to 200 mm Nominal diameter	VDI/VDE/DGQ 2618 Sheet 4.8:2006 Point 3.2.2 (Opt. 1) to Point 3.2.6 (Opt. 5)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Scanning process d = is the measured
Outer diameter			2 $\mu\text{m}$	Diameter
Core diameter / Piercing diameter			5 $\mu\text{m}$	
Gradient / pitch	0.5 mm to 8 mm		1.5 $\mu\text{m}$	
Thread profile angle $\alpha$	> 27°		$(3 \cdot 1 / l_f)^3$ , but not smaller than 6'	$l_f$ = flank length in mm
Threaded rings: simple Flank diameter	5 mm to 200 mm nominal diameter	VDI/VDE/DGQ 2618 Sheet 4.9:2006 Point 3.2.2 (Opt. 1) to Point 3.2.6 (Opt. 5)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Scanning process d = is the measured
Outer diameter			5 $\mu\text{m}$	Diameter
Core diameter / Piercing diameter			2 $\mu\text{m}$	
Gradient / pitch	0.5 mm to 8 mm		1,5 $\mu\text{m}$	
Thread profile angle $\alpha$	> 27°		$(3 \cdot 1 / l_f)^3$ , but not smaller than 6'	$l_f$ = flank length in mm

Valid 19.04.2023

Date of issue: 19.04.2023





Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Dimensional measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)				Remarks
	Measuring range		Measuring conditions / Procedures	Extended uncertainty of measurement	
Inside micrometers with 3-line contact on the calibration object	3 mm	to 150 mm	VDI/VDE/DGQ 2618 Sheet 10.8:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured diameter
Lever gauges (quick probe) for outdoor measurements		to 200 mm	VDI/VDE/DGQ 2618 Sheet 12.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length
Lever gauges (quick probe) for indoor measurements	2 mm	to 200 mm	VDI/VDE/DGQ 2618 Sheet 13.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Dial gauges	0 mm	to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.1:2021	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	mechanical dial gauges
			VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	electronic digital dial gauges
Fine pointer	0 mm	to 3 mm	VDI/VDE/DGQ 2618 Sheet 11.2:2002	0.6 $\mu\text{m}$	
Feeler lever gauges	0 mm	to 1,6 mm	VDI/VDE/DGQ 2618 Sheet 11.3:2002	1.0 $\mu\text{m}$	
electr. inductive Linear Encoders		to 100 mm	VDI/VDE/DGQ 2618 Sheet 14.1:2010	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	
electr. incremental Linear Encoders		to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	
Gauge blocks off  Steel after DIN EN ISO 3650	0,5 mm	to 150 mm	VDI/VDE/DGQ 2618  Sheet 3.1:2004  Deviation measurement  of the center dimension <i>l<sub>c</sub></i> from Nominal dimension <i>l<sub>n</sub></i> through Difference measurement	For the center dimension:  $0.08 \mu\text{m} \cdot 0.7 \cdot 10^{-6} \cdot l$  For <i>f<sub>0</sub></i> and <i>f<sub>u</sub></i> :  0,07 $\mu\text{m}$	<i>l</i> is the length of the dimension  For the smallest measuring Uncertainties are the pushability and Startup features of both
Gauge blocks off  Ceramics after DIN EN ISO 3650			Deviation measurement <i>f<sub>0</sub></i> and <i>f<sub>u</sub></i> from center dimension through 5-point  Difference measurement	For the center dimension:  $0.1 \mu\text{m} \cdot 0.8 \cdot 10^{-6} \cdot l$  For <i>f<sub>0</sub></i> and <i>f<sub>u</sub></i> :  0,07 $\mu\text{m}$	Measuring surfaces of the  Calibration object with a suitable  Flat glass plate to check
Gauge blocks off Tungsten carbide according to DIN EN ISO 3650					For the center dimension:  $0.1 \mu\text{m} \cdot 0.8 \cdot 10^{-6} \cdot l$  For <i>f<sub>0</sub></i> and <i>f<sub>u</sub></i> :  0,07 $\mu\text{m}$

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Dimensional measurands

Calibration and Measurement Capabilities (CMC)						
Measurand/ calibration item	Measuring range		Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
Angle Perpendicularity deviation Flatness and Straightness deviation	to	30 µm	VDI/VDE/DGQ/DKD 2618 Sheet 7.1:2019 (Opt. 2)	$2.5 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l_z$ $4 \mu\text{m} \cdot 5 \cdot 10^{-6} \cdot l_z$	$l_z$ = length of the forming or locating element up to 500 mm Leg length	
Protractor Graduation 1° Scale interval 5'	-180° 0°	to to	180° 360°	VDI/VDE/DGQ 2618 Sheet 7.2:2008	30' 1'	
Flat rulers Parallelism deviation Flatness deviation		to	500 mm	VDI/VDE/DGQ 2618 Sheet 5.1:2022	$4 \mu\text{m} \cdot 5 \cdot 10^{-6} \cdot l$ $2.2 \mu\text{m} \cdot 3.5 \cdot 10^{-6} \cdot l$	$l$ is the measured Length
Straight edge Straightness deviation		to	500 mm	VDI/VDE/DGQ 2618 Sheet 5.2:2013	$2.2 \mu\text{m} \cdot 3.5 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Tape measures and scales Tape measures Standards	0 m 0 m	until until	100 m 3 m	4_VB_00237_EN V3	$50 \mu\text{m} \cdot 20 \cdot 10^{-6} \cdot l$	
Inclinometers	-2000 µm/m (-412")	to	2000 µm/m (412")	4_VB_00244_EN V3	$1.7 \mu\text{m}/\text{m}$ (0,35")	max. leg length of the KG: 500 mm

**Location Calibration laboratory Kirchzarten, Gewerbestraße 3, 79199 Kirchzarten**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC voltage	0 V		0.1 $\mu$ V	<i>U</i> = measured value
	10 mV		$12 \cdot 10^{-6} U$	
	100 mV		$1.3 \cdot 10^{-6} U$	
	1 V		$0.35 \cdot 10^{-6} U$	
	10 V		$0.25 \cdot 10^{-6} U$	
	100 V		$0.4 \cdot 10^{-6} U$	
	1000 V		$0.9 \cdot 10^{-6} U$	
	1 $\mu$ V up to 100 mV > 100 mV up to 100 V > 100 V up to 1000 V		$1.4 \cdot 10^{-6} U$ 0.15 $\mu$ V $0.5 \cdot 10^{-6} U$ $1.4 \cdot 10^{-6} U$	
High Voltage	> 1 kV up to 50 kV		$0.4 \cdot 10^{-3} U$	
Direct current strength	0 A up to 10 pA		$1.6 \cdot 10^{-3} / 2$ fA	<i>I</i> = measured value
	> 10 pA up to 100 pA		$0.3 \cdot 10^{-3} /$	
	> 100 pA up to 1 nA		$0.2 \cdot 10^{-3} /$	
	> 1 nA up to 10 nA		$28 \cdot 10^{-6} /$	
	> 10 nA up to 100 nA		$10 \cdot 10^{-6} /$	
	> 100 nA up to 1 $\mu$ A		$1.0 \cdot 10^{-6} /$	
	> 1 $\mu$ A up to 100 mA		$0.8 \cdot 10^{-6} /$	
	> 100 mA up to 20 A > 20 A up to 1000 A		$1.5 \cdot 10^{-6} /$ $9 \cdot 10^{-6} /$	
Direct current strength  Current clamps	1 up to 2.2 A mA		$1 \cdot 10^{-3} /$	
	> 2,2 A up to 20 A > 20 A up to 1000 A		$2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	
Direct current strength  Current transformer	1 A to 120 A		$0.22 \cdot 10^{-3} /$	
	> 120 A to 1000 A		$0.25 \cdot 10^{-3} /$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC resistance (discrete values)	100 $\mu\Omega$ , 1 m $\Omega$ , 10 m $\Omega$		$1.2 \cdot 10^{-6} R$	$R =$ measured value
	100 m $\Omega$		$0.5 \cdot 10^{-6} R$	
	1 $\Omega$		$80 \cdot 10^{-9} R$	
	10 $\Omega$		$0.12 \cdot 10^{-6} R$	
	100 $\Omega$ , 1 k $\Omega$		$0.1 \cdot 10^{-6} R$	
	10 k $\Omega$		$50 \cdot 10^{-9} R$	
	100 k $\Omega$		$0.1 \cdot 10^{-6} R$	
	1 M $\Omega$		$0.25 \cdot 10^{-6} R$	
	10 M $\Omega$		$0.75 \cdot 10^{-6} R$	
	100 M $\Omega$		$2.5 \cdot 10^{-6} R$	
	1 G $\Omega$		$8 \cdot 10^{-6} R$	
	10 G $\Omega$		$28 \cdot 10^{-6} R$	
	100 G $\Omega$		$64 \cdot 10^{-6} R$	
DC resistance (discrete values)	1 T $\Omega$		$0.12 \cdot 10^{-3} R$	$R =$ measured value
	10 T $\Omega$		$0.32 \cdot 10^{-3} R$	
	100 T $\Omega$		$0.87 \cdot 10^{-3} R$	
DC resistance Areas	0 $\mu\Omega$ to < 1 m $\Omega$		$4 \cdot 10^{-6} R$ 1 n $\Omega$	$R =$ measured value
	1 m $\Omega$ to < 10 m $\Omega$		$3 \cdot 10^{-6} R$ 1 n $\Omega$	
	10 m $\Omega$ to < 100 m $\Omega$		$2 \cdot 10^{-6} R$ 1 n $\Omega$	
	0,1 $\Omega$ to < 1 $\Omega$		$0.5 \cdot 10^{-6} R$	
	1 $\Omega$ to 100 k $\Omega$		$0.2 \cdot 10^{-6} R$	
	> 100 k $\Omega$ to 1 M $\Omega$		$0.6 \cdot 10^{-6} R$	
	> 1 M $\Omega$ to 10 M $\Omega$		$1.1 \cdot 10^{-6} R$	
	> 10 M $\Omega$ to 100 M $\Omega$		$2.5 \cdot 10^{-6} R$	
	> 100 M $\Omega$ to 1 G $\Omega$		$8 \cdot 10^{-6} R$	
	> 1 G $\Omega$ to 10 G $\Omega$		$30 \cdot 10^{-6} R$	
	> 10 G $\Omega$ to 100 G $\Omega$		$82 \cdot 10^{-6} R$	
	> 100 G $\Omega$ to 1 T $\Omega$		$0.14 \cdot 10^{-3} R$	
	> 1 T $\Omega$ to 10 T $\Omega$		$0.35 \cdot 10^{-3} R$	
> 10 T $\Omega$ to 100 T $\Omega$		$1.2 \cdot 10^{-3} R$		
DC power	1 mW to 2 kW	Product of $U$ and $I$ ; $1 \text{ mV} \leq U \leq 1000 \text{ V}$ , $100 \mu\text{A} \leq I \leq 1000 \text{ A}$	$8 \cdot 10^{-6}$	
	> 2 kW to 1000 kW		$15 \cdot 10^{-6}$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC resistance Resistors (discrete values)	0,1 Ω	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz;	25 · 10 <sup>-6</sup> R 15 · 10 <sup>-6</sup> R 11 · 10 <sup>-6</sup> R 15 · 10 <sup>-6</sup> R	R = measured value
	1 Ω	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz;	25 · 10 <sup>-6</sup> R 11 · 10 <sup>-9</sup> R	
	10 Ω	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz;	25 · 10 <sup>-6</sup> R 10 · 10 <sup>-6</sup> R	
	100 Ω	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz 5 kHz; 10 kHz;	12 · 10 <sup>-6</sup> R 8 · 10 <sup>-6</sup> R 6 · 10 <sup>-6</sup> R	
	1 kΩ	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz;	35 · 10 <sup>-6</sup> R	
AC resistance Resistors (discrete values)	10 kΩ	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz;	85 · 10 <sup>-6</sup> R 45 · 10 <sup>-6</sup> R 110 · 10 <sup>-6</sup> R 65 · 10 <sup>-6</sup> R	
AC resistance (areas)	0,1 Ω up to 1 Ω	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 10 kHz	35 · 10 <sup>-6</sup> · R 30 · 10 <sup>-6</sup> · R 20 · 10 <sup>-6</sup> · R	
	1 Ω up to < 10 Ω	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 10 kHz	35 · 10 <sup>-6</sup> · R 30 · 10 <sup>-6</sup> · R 15 · 10 <sup>-6</sup> · R	
	10 Ω up to < 100 Ω	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 10 kHz	35 · 10 <sup>-6</sup> · R 30 · 10 <sup>-6</sup> · R 15 · 10 <sup>-6</sup> · R	
	100 Ω up to < 1 kΩ	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 10 kHz	40 · 10 <sup>-6</sup> · R 40 · 10 <sup>-6</sup> · R 35 · 10 <sup>-6</sup> · R	
	1 kΩ up to 10 kΩ	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to <400 Hz 400 Hz to 10 kHz	90 · 10 <sup>-6</sup> · R 45 · 10 <sup>-6</sup> · R 110 · 10 <sup>-6</sup> · R 65 · 10 <sup>-6</sup> · R	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Cargo Charge amplifier and Charge meters	1 pC to 10 000 pC	0.2 Hz to < 1 Hz	0,5 %	Calibration result: Amount of the Transmission- coefficients
		1 Hz to 10 kHz	0,4 %	
		> 10 kHz to 20 kHz	0,6 %	
		> 20 kHz to 50 kHz	1,0 %	
AC/DC Transfer- AC voltage-sources	1 mV	10 Hz; 20 Hz; 30 Hz	$0.23 \cdot 10^{-3} U$	<i>U</i> = measured value
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$0.11 \cdot 10^{-3} U$	
		300 Hz; 400 Hz; 500 Hz;	$0.11 \cdot 10^{-3} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$0.11 \cdot 10^{-3} U$	
		20 kHz; 30 kHz; 50 kHz	$0.11 \cdot 10^{-3} U$	
		70 kHz; 100 kHz	$0.16 \cdot 10^{-3} U$	
		200 kHz; 300 kHz	$0.17 \cdot 10^{-3} U$	
		500 kHz; 700 kHz; 800 kHz	$0.25 \cdot 10^{-3} U$	
	1 MHz	$0.27 \cdot 10^{-3} U$		
	2 mV	10 Hz, 20 Hz, 30 Hz	$0.13 \cdot 10^{-3} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$0.08 \cdot 10^{-3} U$	
		300 Hz; 400 Hz; 500 Hz;	$0.08 \cdot 10^{-3} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$0.08 \cdot 10^{-3} U$	
		20 kHz; 30 kHz; 50 kHz	$0.08 \cdot 10^{-3} U$	
		70 kHz; 100 kHz; 200 kHz;	$0.11 \cdot 10^{-3} U$	
		300 kHz	$0.11 \cdot 10^{-3} U$	
		500 kHz; 700 kHz; 800 kHz	$0.16 \cdot 10^{-3} U$	
	1 MHz	$0.18 \cdot 10^{-3} U$		
	6 mV	10 Hz; 20 Hz; 30 Hz	$60 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$45 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$45 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$45 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$45 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$55 \cdot 10^{-6} U$	
200 kHz; 300 kHz		$73 \cdot 10^{-6} U$		
500 kHz		$0.13 \cdot 10^{-3} U$		
700 kHz; 800 kHz; 1 MHz	$0.16 \cdot 10^{-3} U$			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage-sources	10 mV	10 Hz; 20 Hz; 30 Hz	$44 \cdot 10^{-6} U$	<i>U</i> = measured value
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$37 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$37 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$37 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$37 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$50 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$60 \cdot 10^{-6} U$	
		500 kHz; 700 kHz; 800 kHz;	$0.14 \cdot 10^{-3} U$	
		1 MHz	$0.14 \cdot 10^{-3} U$	



Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- sources	20 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$30 \cdot 10^{-6} U$ $26 \cdot 10^{-6} U$ $26 \cdot 10^{-6} U$ $26 \cdot 10^{-6} U$ $26 \cdot 10^{-6} U$ $38 \cdot 10^{-6} U$ $47 \cdot 10^{-6} U$ $98 \cdot 10^{-6} U$ $98 \cdot 10^{-6} U$	$U =$ measured value
	40 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$24 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $31 \cdot 10^{-6} U$ $47 \cdot 10^{-6} U$ $90 \cdot 10^{-6} U$ $90 \cdot 10^{-6} U$	
	60 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz 500 Hz; 1 kHz; 2 kHz; 5 kHz 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$29 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $20 \cdot 10^{-6} U$ $23 \cdot 10^{-6} U$ $27 \cdot 10^{-6} U$ $42 \cdot 10^{-6} U$ $86 \cdot 10^{-6} U$ $86 \cdot 10^{-6} U$	
	100 mV	10 Hz; 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz; 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$24 \cdot 10^{-6} U$ $18 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- sources	200 mV	10 Hz; 20 Hz	$18 \cdot 10^{-6} U$	<i>U</i> = measured value
		30 Hz	$11 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$6 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$6 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$6 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$6 \cdot 10^{-6} U$	
		70 kHz; 100 kHz;	$8 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$8 \cdot 10^{-6} U$	
		500 kHz	$17 \cdot 10^{-6} U$	
		700 kHz	$22 \cdot 10^{-6} U$	
	800 kHz	$27 \cdot 10^{-6} U$		
	1 MHz	$28 \cdot 10^{-6} U$		
	300 mV	10 Hz	$14 \cdot 10^{-6} U$	
		20 Hz	$11 \cdot 10^{-6} U$	
		30 Hz	$10 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$5 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$5 \cdot 10^{-6} U$	
		1 kHz; 2 kHz	$5 \cdot 10^{-6} U$	
		5 kHz; 10 kHz; 20 kHz;	$7 \cdot 10^{-6} U$	
		30 kHz; 50 kHz; 70 kHz	$7 \cdot 10^{-6} U$	
		100 kHz; 200 kHz; 300 kHz	$8 \cdot 10^{-6} U$	
		500 kHz; 700 kHz;	$18 \cdot 10^{-6} U$	
	800 kHz; 1 MHz	$18 \cdot 10^{-6} U$		
	400 mV	10 Hz; 20 Hz	$10 \cdot 10^{-6} U$	
		30 Hz	$7 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$5 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$5 \cdot 10^{-6} U$	
1 kHz; 2 kHz; 5 kHz; 10 kHz;		$5 \cdot 10^{-6} U$		
20 kHz; 30 kHz; 50 kHz;		$5 \cdot 10^{-6} U$		
70 kHz; 100 kHz		$5 \cdot 10^{-6} U$		
200 kHz; 300 kHz		$6 \cdot 10^{-6} U$		
500 kHz		$15 \cdot 10^{-6} U$		
700 kHz; 800 kHz; 1MHz		$17 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- sources	500 mV	10 Hz 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz, 50 kHz; 70 kHz; 100 kHz; 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$15 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$	$U =$ measured value
	600 mV	10 Hz 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1kHz; 2 kHz 5 kHz 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$14 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$	
	700 mV	10 Hz 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz; 70 kHz; 100 kHz 200 kHz; 300 kHz; 500 kHz; 700 kHz; 800 kHz; 1 MHz	$15 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage sources	1 V	10 Hz	$9 \cdot 10^{-6} U$	<i>U</i> = measured value
		20 Hz	$8 \cdot 10^{-6} U$	
		30 Hz	$7 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz;	$4 \cdot 10^{-6} U$	
		120 Hz; 300 Hz	$4 \cdot 10^{-6} U$	
		400 Hz; 500 Hz; 1 kHz	$2 \cdot 10^{-6} U$	
		2 kHz; 5 kHz	$3 \cdot 10^{-6} U$	
		10 kHz; 20 kHz;	$4 \cdot 10^{-6} U$	
		30 kHz; 50 kHz	$4 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$5 \cdot 10^{-6} U$	
	200 kHz; 300 kHz	$7 \cdot 10^{-6} U$		
	500 kHz	$11 \cdot 10^{-6} U$		
	700 kHz; 800 kHz; 1 MHz	$14 \cdot 10^{-6} U$		
	2 V	10 Hz; 20 Hz	$8 \cdot 10^{-6} U$	
		30 Hz	$5 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$2 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$2 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$2 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$2 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$5 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$6 \cdot 10^{-6} U$	
		500 kHz	$11 \cdot 10^{-6} U$	
		700 kHz; 800 kHz	$14 \cdot 10^{-6} U$	
	1 MHz	$16 \cdot 10^{-6} U$		
	3 V; 4 V; 5 V; 6 V; 7 V; 8 V	10 Hz	$10 \cdot 10^{-6} U$	
		20 Hz	$8 \cdot 10^{-6} U$	
		30 Hz	$5 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$3 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$3 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$3 \cdot 10^{-6} U$	
20 kHz; 30 kHz; 50 kHz		$3 \cdot 10^{-6} U$		
70 kHz		$4 \cdot 10^{-6} U$		
100 kHz		$5 \cdot 10^{-6} U$		
200 kHz; 300 kHz		$8 \cdot 10^{-6} U$		
500 kHz	$9 \cdot 10^{-6} U$			
700 kHz; 800 kHz	$12 \cdot 10^{-6} U$			
1 MHz	$15 \cdot 10^{-6} U$			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- sources	10 V	10 Hz	$10 \cdot 10^{-6} U$	<i>U</i> = measured value
		20 Hz	$8 \cdot 10^{-6} U$	
		30 Hz	$5 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz; 1 MHz	$4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$	
20 V	10 Hz; 20 Hz	$8 \cdot 10^{-6} U$		
	30 Hz	$5 \cdot 10^{-6} U$		
	40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz; 1 MHz	$3 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$		
	30 V; 40 V; 50 V; 60 V; 70 V	10 Hz; 20 Hz	$9 \cdot 10^{-6} U$	
30 V; 40 V; 50 V; 60 V; 70 V	30 Hz	$7 \cdot 10^{-6} U$		
	40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz	$6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$		
	100 V	10 Hz	$10 \cdot 10^{-6} U$	
		20 Hz	$9 \cdot 10^{-6} U$	
30 Hz		$7 \cdot 10^{-6} U$		
40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz		$6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / method	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- sources	200 V	10 Hz; 20 Hz 30 Hz; 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz	$10 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U$ = measured value
	300 V; 400 V; 500 V; 600 V; 700 V; 800 V; 1000 V	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz 2 kHz; 5 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz 70 kHz; 100 kHz	$9 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$	
AC/DC Transfer- AC voltage- measuring instruments	1 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz; 70 kHz; 100 kHz; 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz 1 MHz	$0.30 \cdot 10^{-3} U$ $0.22 \cdot 10^{-3} U$ $0.22 \cdot 10^{-3} U$ $0.22 \cdot 10^{-3} U$ $0.26 \cdot 10^{-3} U$ $0.26 \cdot 10^{-3} U$ $0.26 \cdot 10^{-3} U$ $0.32 \cdot 10^{-3} U$ $0.33 \cdot 10^{-3} U$	
	2 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz; 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$0.16 \cdot 10^{-3} U$ $0.13 \cdot 10^{-3} U$ $0.13 \cdot 10^{-3} U$ $0.13 \cdot 10^{-3} U$ $0.13 \cdot 10^{-3} U$ $0.14 \cdot 10^{-3} U$ $0.14 \cdot 10^{-3} U$ $0.20 \cdot 10^{-3} U$ $0.20 \cdot 10^{-3} U$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- measuring instruments	6 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz; 1 MHz	$67 \cdot 10^{-6} U$ $57 \cdot 10^{-6} U$ $57 \cdot 10^{-6} U$ $57 \cdot 10^{-6} U$ $57 \cdot 10^{-6} U$ $65 \cdot 10^{-6} U$ $80 \cdot 10^{-6} U$ $0.14 \cdot 10^{-3} U$ $0.16 \cdot 10^{-3} U$	$U =$ measured value
	10 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$50 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $55 \cdot 10^{-6} U$ $65 \cdot 10^{-6} U$ $0.15 \cdot 10^{-3} U$ $0.15 \cdot 10^{-3} U$	
	20 mV	10 Hz 20 Hz; 30 Hz; 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$32 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $40 \cdot 10^{-6} U$ $50 \cdot 10^{-6} U$ $0.1 \cdot 10^{-3} U$ $0.1 \cdot 10^{-3} U$	
	40 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz; 500 kHz; 700 kHz; 800 kHz; 1 MHz	$30 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $43 \cdot 10^{-6} U$ $86 \cdot 10^{-6} U$ $86 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / method	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- measuring instruments	60 mV	10 Hz; 20 Hz; 30 Hz	$30 \cdot 10^{-6} U$	<i>U</i> = measured value
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$25 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$25 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$25 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$25 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$28 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$43 \cdot 10^{-6} U$	
		500 kHz; 700 kHz; 800 kHz;	$86 \cdot 10^{-6} U$	
		1 MHz	$86 \cdot 10^{-6} U$	
	100 mV	10 Hz; 20 Hz	$24 \cdot 10^{-6} U$	
		30 Hz	$18 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$8 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$8 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$8 \cdot 10^{-6} U$	
		20 kHz; 300 kHz; 50 kHz	$8 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$9 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$10 \cdot 10^{-6} U$	
		500 kHz; 700 kHz; 800 kHz;	$30 \cdot 10^{-6} U$	
1 MHz	$30 \cdot 10^{-6} U$			
	200 mV	10 Hz; 20 Hz	$18 \cdot 10^{-6} U$	
		30 Hz	$11 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz,	$6 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$6 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$6 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$6 \cdot 10^{-6} U$	
		70 kHz; 100 kHz; 200 kHz;	$8 \cdot 10^{-6} U$	
		300 kHz	$8 \cdot 10^{-6} U$	
		500 kHz;	$17 \cdot 10^{-6} U$	
700 kHz	$22 \cdot 10^{-6} U$			
800 kHz; 1 MHz	$28 \cdot 10^{-6} U$			



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / method	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- measuring instruments	300 mV	10 Hz; 20 Hz 30 Hz 40 Hz, 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz, 2 kHz; 5 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz 70 kHz; 100 kHz; 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz; 1 MHz	$14 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $18 \cdot 10^{-6} U$	$U =$ measured value
	400 mV	10 Hz; 20 Hz 30 Hz 40 Hz, 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz 1 kHz, 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz 50 kHz; 70 kHz; 100 kHz; 200 kHz 300 kHz 500 kHz 700 kHz; 800 kHz; 1 MHz	$10 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $17 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / method	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- measuring instruments	500 mV	10 Hz	$15 \cdot 10^{-6} U$	<i>U</i> = measured value
		20 Hz	$9 \cdot 10^{-6} U$	
		30 Hz	$6 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$4 \cdot 10^{-6} U$	
		300 Hz; 400 Hz	$4 \cdot 10^{-6} U$	
		500 Hz; 1 kHz; 2 kHz; 5 kHz;	$6 \cdot 10^{-6} U$	
		10 kHz; 20 kHz; 30 kHz;	$6 \cdot 10^{-6} U$	
		50 kHz; 70 kHz; 100 kHz;	$6 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$6 \cdot 10^{-6} U$	
		500 kHz; 700 kHz; 800 kHz;	$15 \cdot 10^{-6} U$	
	1 MHz	$15 \cdot 10^{-6} U$		
	600 mV	10 Hz	$14 \cdot 10^{-6} U$	
		20 Hz	$9 \cdot 10^{-6} U$	
		30 Hz	$7 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$4 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$4 \cdot 10^{-6} U$	
		1 kHz; 2kHz	$4 \cdot 10^{-6} U$	
		5 kHz	$6 \cdot 10^{-6} U$	
		10 kHz; 20 kHz; 30 kHz;	$4 \cdot 10^{-6} U$	
		50 kHz	$4 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$5 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$6 \cdot 10^{-6} U$	
		500 kHz; 700 kHz; 800 kHz;	$15 \cdot 10^{-6} U$	
		1 MHz	$15 \cdot 10^{-6} U$	
	700 mV	10 Hz	$15 \cdot 10^{-6} U$	
		20 Hz	$9 \cdot 10^{-6} U$	
		30 Hz	$6 \cdot 10^{-6} U$	
40 Hz; 55 Hz; 60 Hz; 120 Hz;		$4 \cdot 10^{-6} U$		
300 Hz; 400 Hz; 500 Hz;		$4 \cdot 10^{-6} U$		
1 kHz		$4 \cdot 10^{-6} U$		
2 kHz; 5 kHz; 10 kHz; 20 kHz;		$5 \cdot 10^{-6} U$		
30 kHz; 50 kHz; 70 kHz;		$5 \cdot 10^{-6} U$		
100 kHz		$5 \cdot 10^{-6} U$		
200 kHz; 300 kHz		$6 \cdot 10^{-6} U$		
500 kHz; 700 kHz; 800 kHz;		$15 \cdot 10^{-6} U$		
1 MHz		$15 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / method	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- measuring instruments	1 V	10 Hz	$9 \cdot 10^{-6} U$	<i>U</i> = measured value
		20 Hz	$8 \cdot 10^{-6} U$	
		30 Hz	$7 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$4 \cdot 10^{-6} U$	
		300 Hz	$4 \cdot 10^{-6} U$	
		400 Hz; 500 Hz; 1 kHz	$2 \cdot 10^{-6} U$	
		2 kHz; 5 kHz	$3 \cdot 10^{-6} U$	
		10 kHz; 20 kHz; 30 kHz;	$4 \cdot 10^{-6} U$	
		50 kHz	$4 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$5 \cdot 10^{-6} U$	
	200 kHz; 300 kHz	$7 \cdot 10^{-6} U$		
	500 kHz	$11 \cdot 10^{-6} U$		
	700 kHz; 800 kHz; 1 MHz	$14 \cdot 10^{-6} U$		
	2 V	10 Hz; 20 Hz	$8 \cdot 10^{-6} U$	
		30 Hz	$5 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$2 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$2 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$2 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$2 \cdot 10^{-6} U$	
		70 kHz; 100 kHz	$5 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$6 \cdot 10^{-6} U$	
		500 kHz	$11 \cdot 10^{-6} U$	
		700 kHz; 800 kHz	$14 \cdot 10^{-6} U$	
	1 MHz	$16 \cdot 10^{-6} U$		
	3 V; 4 V; 5 V; 6 V; 7 V; 8 V	10 Hz	$10 \cdot 10^{-6} U$	
		20 Hz	$8 \cdot 10^{-6} U$	
		30 Hz	$5 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$3 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$3 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$3 \cdot 10^{-6} U$	
20 kHz; 30 kHz; 50 kHz		$3 \cdot 10^{-6} U$		
70 kHz		$4 \cdot 10^{-6} U$		
100 kHz		$5 \cdot 10^{-6} U$		
200 kHz; 300 kHz		$8 \cdot 10^{-6} U$		
500 kHz	$9 \cdot 10^{-6} U$			
700 kHz; 800 kHz	$12 \cdot 10^{-6} U$			
1 MHz	$15 \cdot 10^{-6} U$			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- measuring instruments	10 V	10 Hz 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz, 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz, 1 MHz	$10 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $4 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$	<i>U</i> = measured value
	20 V	10 Hz, 20 Hz 30 Hz 40 Hz, 55 Hz; 60 Hz, 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz; 1 MHz	$8 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $3 \cdot 10^{-6} U$ $5 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	
	30 V; 40 V; 50 V; 60 V; 70 V	10 Hz, 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz, 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz	$9 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$	
	100 V	10 Hz 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz	$10 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $6 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC/DC Transfer- AC voltage- measuring instruments	200 V	10 Hz; 20 Hz 30 Hz; 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz	$10 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	<i>U</i> = measured value
	300 V; 400 V; 500 V; 600 V; 700 V; 800 V; 1000 V	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz, 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz, 2 kHz; 5 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz 70 kHz, 100 kHz	$9 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $7 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$	
AC voltage Measuring devices and Sources	1 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz; 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz 1 MHz	$0.42 \cdot 10^{-3} U$ $0.36 \cdot 10^{-3} U$ $0.36 \cdot 10^{-3} U$ $0.36 \cdot 10^{-3} U$ $0.36 \cdot 10^{-3} U$ $0.36 \cdot 10^{-3} U$ $0.43 \cdot 10^{-3} U$ $0.48 \cdot 10^{-3} U$ $0.53 \cdot 10^{-3} U$	
	2 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz; 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz 1 MHz	$0.18 \cdot 10^{-3} U$ $0.14 \cdot 10^{-3} U$ $0.14 \cdot 10^{-3} U$ $0.14 \cdot 10^{-3} U$ $0.14 \cdot 10^{-3} U$ $0.18 \cdot 10^{-3} U$ $0.18 \cdot 10^{-3} U$ $0.21 \cdot 10^{-3} U$ $0.24 \cdot 10^{-3} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage  Measuring devices and sources	6 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz;  300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$70 \cdot 10^{-6} U$ $60 \cdot 10^{-6} U$ $60 \cdot 10^{-6} U$ $60 \cdot 10^{-6} U$ $60 \cdot 10^{-6} U$ $70 \cdot 10^{-6} U$ $85 \cdot 10^{-6} U$ $0.17 \cdot 10^{-3} U$ $0.17 \cdot 10^{-3} U$	$U$ = measured value
	10 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$53 \cdot 10^{-6} U$ $47 \cdot 10^{-6} U$ $47 \cdot 10^{-6} U$ $47 \cdot 10^{-6} U$ $47 \cdot 10^{-6} U$ $57 \cdot 10^{-6} U$ $70 \cdot 10^{-6} U$ $0.14 \cdot 10^{-3} U$ $0.14 \cdot 10^{-3} U$	
	20 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$37 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $56 \cdot 10^{-6} U$ $0.11 \cdot 10^{-3} U$ $0.11 \cdot 10^{-3} U$	
	40 mV	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz; 1 MHz	$33 \cdot 10^{-6} U$ $31 \cdot 10^{-6} U$ $31 \cdot 10^{-6} U$ $31 \cdot 10^{-6} U$ $31 \cdot 10^{-6} U$ $40 \cdot 10^{-6} U$ $56 \cdot 10^{-6} U$ $95 \cdot 10^{-6} U$ $95 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / method	Extended uncertainty of measurement	Remarks
AC voltage  Measuring devices and sources	60 mV	10 Hz; 20 Hz; 30 Hz  40 Hz; 55 Hz; 60 Hz; 120 Hz;  300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz 1 MHz	$31 \cdot 10^{-6} U$  $25 \cdot 10^{-6} U$  $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $29 \cdot 10^{-6} U$ $43 \cdot 10^{-6} U$ $87 \cdot 10^{-6} U$ $98 \cdot 10^{-6} U$	$U =$ measured value
	100 mV	10 Hz; 20 Hz 30 Hz 40Hz, 55Hz; 60Hz;120Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz 1 MHz	$26 \cdot 10^{-6} U$  $20 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $33 \cdot 10^{-6} U$ $53 \cdot 10^{-6} U$	
	200 mV	10 Hz; 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz; 200 kHz; 300 kHz 500 kHz; 700 kHz 800 kHz; 1 MHz	$21 \cdot 10^{-6} U$  $15 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage  Measuring devices and sources	300 mV	10 Hz; 20 Hz; 30 Hz  40 Hz; 55 Hz; 60 Hz; 120 Hz;  300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz; 70 kHz 100 kHz; 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz 1 MHz	$17 \cdot 10^{-6} U$  $12 \cdot 10^{-6} U$  $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $17 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $28 \cdot 10^{-6} U$	<i>U</i> = measured value
	500 mV; 600 mV; 700 mV	10 Hz 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz; 200 kHz; 300 kHz 500 kHz; 700 kHz; 800 kHz 1 MHz	$18 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $21 \cdot 10^{-6} U$ $40 \cdot 10^{-6} U$	
	1 V	10 Hz; 20 Hz 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz 200 kHz; 300 kHz 500 kHz 700 kHz; 800 kHz 1 MHz	$13 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $60 \cdot 10^{-6} U$	



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage Measuring devices and Sources	2 V	10 Hz; 20 Hz	$12 \cdot 10^{-6} U$	<i>U</i> = measured value
		30 Hz	$10 \cdot 10^{-6} U$	
		40 Hz; 55 Hz; 60 Hz; 120 Hz;	$9 \cdot 10^{-6} U$	
		300 Hz; 400 Hz; 500 Hz;	$9 \cdot 10^{-6} U$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz;	$9 \cdot 10^{-6} U$	
		20 kHz; 30 kHz; 50 kHz	$9 \cdot 10^{-6} U$	
		70 kHz; 100 kHz;	$10 \cdot 10^{-6} U$	
		200 kHz; 300 kHz	$10 \cdot 10^{-6} U$	
		500 kHz	$15 \cdot 10^{-6} U$	
700 kHz; 800 kHz	$25 \cdot 10^{-6} U$			
1 MHz	$67 \cdot 10^{-6} U$			
3 V; 4 V; 5 V	3 V; 4 V; 5 V	10 Hz; 20 Hz	$15 \cdot 10^{-6} U$	
		30 Hz; 40 Hz; 55 Hz; 60 Hz;	$11 \cdot 10^{-6} U$	
		120 Hz; 300 Hz; 400 Hz;	$11 \cdot 10^{-6} U$	
		500 Hz; 1 kHz; 2 kHz; 5 kHz	$11 \cdot 10^{-6} U$	
		10 kHz; 20 kHz; 30 kHz;	$11 \cdot 10^{-6} U$	
		50 kHz; 70 kHz	$11 \cdot 10^{-6} U$	
		100 kHz; 200 kHz; 300 kHz	$11 \cdot 10^{-6} U$	
		500 kHz	$15 \cdot 10^{-6} U$	
		700 kHz; 800 kHz	$25 \cdot 10^{-6} U$	
1 MHz	$67 \cdot 10^{-6} U$			
6 V; 7 V; 8 V	6 V; 7 V; 8 V	10 Hz; 20 Hz	$15 \cdot 10^{-6} U$	
		30 Hz; 40 Hz; 55 Hz; 60 Hz;	$11 \cdot 10^{-6} U$	
		120 Hz; 300 Hz; 400 Hz;	$11 \cdot 10^{-6} U$	
		500 Hz; 1 kHz; 2 kHz; 5 kHz;	$11 \cdot 10^{-6} U$	
		10 kHz; 20 kHz; 30 kHz;	$11 \cdot 10^{-6} U$	
		50 kHz; 70 kHz	$11 \cdot 10^{-6} U$	
		100 kHz; 200 kHz; 300 kHz	$13 \cdot 10^{-6} U$	
		500 kHz	$30 \cdot 10^{-6} U$	
		700 kHz; 800 kHz	$60 \cdot 10^{-6} U$	
1 MHz	$95 \cdot 10^{-6} U$			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage Measuring devices and Sources	10 V; 20 V	10 Hz; 20 Hz 30 Hz; 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz; 70 kHz 100 kHz; 200 kHz; 300 kHz; 500 kHz 700 kHz; 800 kHz; 1 MHz	$13 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $0.11 \cdot 10^{-3} U$	$U =$ measured value
	30 V; 40 V; 50 V; 60 V; 70 V	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz; 100 kHz	$15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $20 \cdot 10^{-6} U$	
	100 V; 200 V	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz; 70 kHz 100 kHz	$17 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $17 \cdot 10^{-6} U$ $17 \cdot 10^{-6} U$ $17 \cdot 10^{-6} U$ $32 \cdot 10^{-6} U$	
	300 V	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz; 1 kHz 2 kHz; 5 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz 70 kHz 100 kHz	$17 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $20 \cdot 10^{-6} U$ $32 \cdot 10^{-6} U$ $44 \cdot 10^{-6} U$ $66 \cdot 10^{-6} U$	
	500 V; 1000 V	10 Hz; 20 Hz 30 Hz; 40 Hz; 55 Hz; 60 Hz; 120 Hz; 300 Hz; 400 Hz; 500 Hz 1 kHz; 2 kHz; 5 kHz; 10 kHz 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz	$24 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $47 \cdot 10^{-6} U$ $55 \cdot 10^{-6} U$ $66 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage  Measuring devices and Sources	1 mV to 2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV to 7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
		> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$	
	> 7 mV to 22 mV	10 Hz to 20 Hz	$80 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$80 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$65 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$75 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$75 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$95 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$0.19 \cdot 10^{-3} U$		
> 500 kHz to 1 MHz		$0.21 \cdot 10^{-3} U$		
> 22 mV to 70 mV	10 Hz to 20 Hz	$70 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$58 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$35 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$35 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$45 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$55 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$0.11 \cdot 10^{-3} U$		
	> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage  Measuring devices and  Sources	> 70 mV to 220 mV	10 Hz to 20 Hz	$39 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$35 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$25 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$25 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$28 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$42 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$85 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.1 \cdot 10^{-3} U$	
	> 220 mV to 700 mV	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$22 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$12 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$12 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$14 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$27 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$40 \cdot 10^{-6} U$	
	> 700 mV to 2.2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$11 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$11 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$22 \cdot 10^{-6} U$		
> 500 kHz to 1 MHz		$68 \cdot 10^{-6} U$		
> 2.2 V to 7 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$13 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$		
	> 500 kHz to 1 MHz	$95 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage  Measuring devices and  Sources	> 7 V to 22 V	10 Hz to 20 Hz	$17 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
		> 500 kHz to 1MHz	$0.11 \cdot 10^{-3} U$	
	> 22 V to 70 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
> 70 V to 220 V	10 Hz to 20 Hz	$19 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$18 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$17 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$17 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$32 \cdot 10^{-6} U$		
> 220 V to 1000 V	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$27 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$45 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$45 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$65 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage Measuring devices and Sources	0,01 V to 0,1 V	10 Hz to 40 Hz	$0.69 \cdot 10^{-3} U$	<i>U</i> = measured value
		> 40 Hz to 20 kHz	$0.53 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.64 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$1.1 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$2.1 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$3.6 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$5.0 \cdot 10^{-3} U$	
	> 0,1 V to 0,22 V	10 Hz to 40 Hz	$0.38 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.28 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.65 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$1.1 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$1.6 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$3.3 \cdot 10^{-3} U$	
	> 0,22 V to 2,2 V	10 Hz to 40 Hz	$0.49 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.09 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.14 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.29 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.85 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$2.1 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$3.3 \cdot 10^{-3} U$	
	> 2,2 V to 22 V	10 Hz to 40 Hz	$0.45 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.07 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.13 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.21 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.6 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$2.0 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$3.1 \cdot 10^{-3} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage Measuring devices and Sources	> 22 V to 220 V	10 Hz to 40 Hz	$0.47 \cdot 10^{-3} U$	$U =$ measured value with Fluke 5720A  (areas)
		> 40 Hz to 20 kHz	$0.09 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.14 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.29 \cdot 10^{-3} U$	
	> 220 V to 1000 V	50 Hz to 1 kHz	$85 \cdot 10^{-6} U$	
AC voltage with 50 $\Omega$ input impedance	3,2 V	10 Hz; 40 Hz; 100 Hz; 500 Hz 1 kHz; 10 kHz; 50 kHz 100 kHz; 200 kHz; 500 kHz 1 MHz; 2 MHz; 4 MHz	$0.5 \cdot 10^{-3}$	at discrete points
		5 MHz; 8 MHz	$1.1 \cdot 10^{-3}$	
		10 MHz; 15 MHz; 20 MHz	$2.0 \cdot 10^{-3}$	
		26 MHz; 30 MHz; 50 MHz	$3.2 \cdot 10^{-3}$	
	1 V; 320 mV	10 Hz; 40 Hz; 100 Hz; 500 Hz 1 kHz; 10 kHz; 50 kHz 100 kHz; 200 kHz; 500 kHz 1 MHz; 2 MHz; 4 MHz	$0.7 \cdot 10^{-3}$	
		5 MHz; 8 MHz	$1.8 \cdot 10^{-3}$	
		10 MHz; 15 MHz; 20 MHz	$3.5 \cdot 10^{-3}$	
		26 MHz; 30 MHz; 50 MHz	$5.4 \cdot 10^{-3}$	
	100 mV; 32 mV 10 mV; 3.2 mV 1 mV	10 Hz; 40 Hz; 100 Hz; 500 Hz 1 kHz; 10 kHz; 50 kHz 100 kHz; 200 kHz; 500 kHz 1 MHz; 2 MHz; 4 MHz	$1.3 \cdot 10^{-3}$	
		5 MHz; 8 MHz	$2.7 \cdot 10^{-3}$	
		10 MHz; 15 MHz; 20 MHz	$5.2 \cdot 10^{-3}$	
		26 MHz; 30 MHz; 50 MHz	$7.9 \cdot 10^{-3}$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
Alternating current strength- DC current strength- Transfer	100 µA	10 Hz	$82 \cdot 10^{-6} /$	/ = measured value	
		20 Hz	$39 \cdot 10^{-6} /$		
		30 Hz	$31 \cdot 10^{-6} /$		
		40 Hz	$34 \cdot 10^{-6} /$		
		55 Hz	$0.11 \cdot 10^{-3} /$		
		400 Hz	$63 \cdot 10^{-6} /$		
		500 Hz; 1 kHz	$41 \cdot 10^{-6} /$		
		2 kHz	$39 \cdot 10^{-6} /$		
		5 kHz; 10 kHz	$31 \cdot 10^{-6} /$		
	300 µA	10 Hz	$37 \cdot 10^{-6} /$		
		20 Hz	$34 \cdot 10^{-6} /$		
		30 Hz; 40 Hz	$31 \cdot 10^{-6} /$		
		55 Hz	$41 \cdot 10^{-6} /$		
		400 Hz	$35 \cdot 10^{-6} /$		
		500 Hz; 1 kHz	$31 \cdot 10^{-6} /$		
		2 kHz; 5 kHz; 10 kHz	$32 \cdot 10^{-6} /$		
	1 mA	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$31 \cdot 10^{-6} /$ $31 \cdot 10^{-6} /$ $31 \cdot 10^{-6} /$		
		3 mA	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz 10 kHz		$7 \cdot 10^{-6} /$ $7 \cdot 10^{-6} /$ $7 \cdot 10^{-6} /$ $8 \cdot 10^{-6} /$
			5 mA		10 Hz; 20 Hz 30 Hz 40 Hz; 55 Hz 400 Hz; 500 Hz; 1 kHz; 2 kHz 5 kHz 10 kHz
	10 mA				10 Hz 20 Hz; 30 Hz 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Alternating current strength- DC current strength- Transfer	20 mA; 30 mA; 50 mA	10 Hz 20 Hz; 30 Hz; 40 Hz; 55 Hz 400 Hz; 500 Hz; 1 kHz; 2 kHz 5 kHz; 10 kHz	$8 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$ $4 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$	/ = measured value
	100 mA	10 Hz 20 Hz; 30 Hz; 40 Hz; 55 Hz 400 Hz 500 Hz; 1 kHz 2 kHz 5 kHz; 10 kHz	$8 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$ $8 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$ $8 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$	
	200 mA	10 Hz 20 Hz; 30 Hz; 40 Hz 55 Hz; 400 Hz; 500 Hz 1 kHz 2 kHz; 5 kHz; 10 kHz	$7 \cdot 10^{-6} /$ $6 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$ $8 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$	
	300 mA; 500 mA	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz 400 Hz; 500 Hz; 1 kHz; 2 kHz 5 kHz; 10 kHz	$6 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$ $4 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$	
	1 A	10 Hz 20 Hz; 30 Hz 40 Hz; 55 Hz 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$7 \cdot 10^{-6} /$ $6 \cdot 10^{-6} /$ $7 \cdot 10^{-6} /$ $5 \cdot 10^{-6} /$ $9 \cdot 10^{-6} /$	
	2 A	10 Hz 20 Hz; 30 Hz 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$7 \cdot 10^{-6} /$ $8 \cdot 10^{-6} /$ $7 \cdot 10^{-6} /$ $7 \cdot 10^{-6} /$ $7 \cdot 10^{-6} /$	
	3 A; 5 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$14 \cdot 10^{-6} /$ $14 \cdot 10^{-6} /$ $14 \cdot 10^{-6} /$	
	10 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$21 \cdot 10^{-6} /$ $21 \cdot 10^{-6} /$ $21 \cdot 10^{-6} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Alternating current strength- DC current strength- Transfer	20 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$26 \cdot 10^{-6} /$ $26 \cdot 10^{-6} /$ $26 \cdot 10^{-6} /$ $30 \cdot 10^{-6} /$	/ = measured value
	50 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$32 \cdot 10^{-6} /$ $32 \cdot 10^{-6} /$ $32 \cdot 10^{-6} /$ $40 \cdot 10^{-6} /$	
	100 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$47 \cdot 10^{-6} /$ $47 \cdot 10^{-6} /$ $47 \cdot 10^{-6} /$ $92 \cdot 10^{-6} /$	
AC power Sources	100 $\mu$ A	10 Hz 20 Hz 30 Hz; 40 Hz 55 Hz 400 Hz 500 Hz; 1 kHz; 2 kHz 5 kHz; 10 kHz; 20 kHz; 30 kHz 50 kHz 70 kHz; 100 kHz	$83 \cdot 10^{-6} /$ $40 \cdot 10^{-6} /$ $34 \cdot 10^{-6} /$ $0.11 \cdot 10^{-3} /$ $64 \cdot 10^{-6} /$ $42 \cdot 10^{-6} /$ $33 \cdot 10^{-6} /$ $47 \cdot 10^{-6} /$ $77 \cdot 10^{-6} /$	
	300 $\mu$ A	10 Hz 20 Hz 30 Hz; 40 Hz 55 Hz 400 Hz 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz 70 kHz 100 kHz	$38 \cdot 10^{-6} /$ $34 \cdot 10^{-6} /$ $32 \cdot 10^{-6} /$ $42 \cdot 10^{-6} /$ $36 \cdot 10^{-6} /$ $33 \cdot 10^{-6} /$ $33 \cdot 10^{-6} /$ $33 \cdot 10^{-6} /$ $52 \cdot 10^{-6} /$ $0.11 \cdot 10^{-3} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC power Sources	1 mA	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz; 50 kHz; 70 kHz 100 kHz	$32 \cdot 10^{-6} l$ $32 \cdot 10^{-6} l$ $32 \cdot 10^{-6} l$ $32 \cdot 10^{-6} l$ $34 \cdot 10^{-6} l$	l = measured value
	3 mA	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz 10 kHz 20 kHz; 30 kHz; 50 kHz; 70 kHz 100 kHz	$9 \cdot 10^{-6} l$ $9 \cdot 10^{-6} l$ $9 \cdot 10^{-6} l$ $10 \cdot 10^{-6} l$ $12 \cdot 10^{-6} l$ $15 \cdot 10^{-6} l$	
	5 mA	10 Hz; 20 Hz; 30 Hz 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz 10 kHz 20 kHz; 30 kHz; 50 kHz; 70 kHz 100 kHz	$9 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $10 \cdot 10^{-6} l$ $12 \cdot 10^{-6} l$ $15 \cdot 10^{-6} l$	
	10 mA	10 Hz 20 Hz; 30 Hz; 40 Hz; 55 Hz 400 Hz; 500 Hz; 1 kHz; 2 kHz 5 kHz; 10 kHz; 20 kHz; 30 kHz 50 kHz; 70 kHz 100 kHz	$9 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $7 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $10 \cdot 10^{-6} l$ $12 \cdot 10^{-6} l$	
	20 mA; 30 mA; 50 mA; 100 mA	10 Hz 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz 1 kHz; 2 kHz; 5 kHz; 10 kHz; 20 kHz; 30 kHz 50 kHz; 70 kHz 100 kHz	$10 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $10 \cdot 10^{-6} l$ $13 \cdot 10^{-6} l$	
	200 mA	10 Hz 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz 1 kHz 2 kHz; 5 kHz; 10 kHz; 20 kHz 30 kHz; 50 kHz; 70 kHz; 100 kHz	$9 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $10 \cdot 10^{-6} l$ $9 \cdot 10^{-6} l$ $13 \cdot 10^{-6} l$ $13 \cdot 10^{-6} l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC power Sources	300 mA	10 Hz; 20 Hz	$9 \cdot 10^{-6} /$	/ = measured value
		30 Hz; 40 Hz; 55 Hz; 400 Hz;	$8 \cdot 10^{-6} /$	
		500 Hz; 1 kHz; 2 kHz; 5 kHz;	$8 \cdot 10^{-6} /$	
		10 kHz; 20 kHz	$8 \cdot 10^{-6} /$	
		30 kHz; 50 kHz	$9 \cdot 10^{-6} /$	
		70 kHz; 100 kHz	$14 \cdot 10^{-6} /$	
	500 mA	10 Hz; 20 Hz; 30 Hz	$9 \cdot 10^{-6} /$	
		40 Hz; 55 Hz; 400 Hz; 500 Hz;	$8 \cdot 10^{-6} /$	
		1 kHz; 2 kHz; 5 kHz;	$8 \cdot 10^{-6} /$	
		10 kHz; 20 kHz	$8 \cdot 10^{-6} /$	
		30 kHz; 50 kHz	$9 \cdot 10^{-6} /$	
		70 kHz	$11 \cdot 10^{-6} /$	
	1 A	10 Hz; 20 Hz; 30 Hz;	$9 \cdot 10^{-6} /$	
		40 Hz; 55 Hz	$9 \cdot 10^{-6} /$	
		400 Hz; 500 Hz	$8 \cdot 10^{-6} /$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz	$11 \cdot 10^{-6} /$	
		20 kHz; 30 kHz; 50 kHz	$11 \cdot 10^{-6} /$	
		70 kHz	$13 \cdot 10^{-6} /$	
	2 A	10 Hz; 20 Hz; 30 Hz	$10 \cdot 10^{-6} /$	
		40 Hz; 55 Hz; 400 Hz; 500 Hz	$8 \cdot 10^{-6} /$	
1 kHz; 2 kHz; 5 kHz; 10 kHz		$9 \cdot 10^{-6} /$		
20 kHz; 30 kHz; 50 kHz		$13 \cdot 10^{-6} /$		
70 kHz		$18 \cdot 10^{-6} /$		
100 kHz		$23 \cdot 10^{-6} /$		
3 A	10 Hz; 20 Hz; 30 Hz	$18 \cdot 10^{-6} /$		
	40 Hz; 55 Hz; 400 Hz;	$17 \cdot 10^{-6} /$		
	500 Hz; 1k Hz; 2 kHz; 5 kHz;	$17 \cdot 10^{-6} /$		
	10 kHz; 20 kHz; 30 kHz	$17 \cdot 10^{-6} /$		
	50 kHz	$18 \cdot 10^{-6} /$		
	70 kHz	$27 \cdot 10^{-6} /$		
		100 kHz	$29 \cdot 10^{-6} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC power Sources	5 A	10 Hz; 20 Hz; 30 Hz; 40 Hz;	$16 \cdot 10^{-6} l$	l = measured value
		55 Hz; 400 Hz; 500 Hz;	$16 \cdot 10^{-6} l$	
		1 kHz; 2 kHz; 5 kHz;	$16 \cdot 10^{-6} l$	
		10 kHz; 20 kHz; 30 kHz	$16 \cdot 10^{-6} l$	
		50 kHz	$18 \cdot 10^{-6} l$	
		70 kHz	$27 \cdot 10^{-6} l$	
		100 kHz	$29 \cdot 10^{-6} l$	
	10 A	10 Hz; 20 Hz; 30 Hz; 40 Hz;	$22 \cdot 10^{-6} l$	
		55 Hz; 400 Hz; 500 Hz;	$22 \cdot 10^{-6} l$	
		1 kHz; 2 kHz; 5 kHz;	$22 \cdot 10^{-6} l$	
		10 kHz; 20 kHz	$22 \cdot 10^{-6} l$	
		30 kHz	$31 \cdot 10^{-6} l$	
		50 kHz	$41 \cdot 10^{-6} l$	
		70 kHz	$51 \cdot 10^{-6} l$	
		100 kHz	$76 \cdot 10^{-6} l$	
	20 A	10 Hz; 20 Hz; 30 Hz; 40 Hz;	$27 \cdot 10^{-6} l$	
		55 Hz; 400 Hz; 500 Hz;	$27 \cdot 10^{-6} l$	
		1 kHz; 2 kHz; 5 kHz; 10 kHz	$27 \cdot 10^{-6} l$	
		20 kHz; 30 kHz	$31 \cdot 10^{-6} l$	
		50 kHz	$46 \cdot 10^{-6} l$	
		70 kHz	$0.13 \cdot 10^{-3} l$	
		100 kHz	$0.17 \cdot 10^{-3} l$	
	50 A	10 Hz; 20 Hz; 30 Hz; 40 Hz;	$33 \cdot 10^{-6} l$	
		55 Hz; 400 Hz; 500 Hz;	$33 \cdot 10^{-6} l$	
		1 kHz; 2 kHz;	$33 \cdot 10^{-6} l$	
		5 kHz; 10 kHz	$40 \cdot 10^{-6} l$	
	100 A	10 Hz; 20 Hz; 30 Hz; 40 Hz;	$48 \cdot 10^{-6} l$	
		55 Hz; 400 Hz; 500 Hz;	$48 \cdot 10^{-6} l$	
		1 kHz; 2 kHz; 5 kHz;	$48 \cdot 10^{-6} l$	
		10 kHz	$93 \cdot 10^{-6} l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC power Measuring devices	100 µA	10 Hz	$83 \cdot 10^{-6} l$	l = measured value
		20 Hz	$40 \cdot 10^{-6} l$	
		30 Hz; 40 Hz	$34 \cdot 10^{-6} l$	
		55 Hz	$0.11 \cdot 10^{-3} l$	
		400 Hz	$64 \cdot 10^{-6} l$	
		500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$42 \cdot 10^{-6} l$ $42 \cdot 10^{-6} l$	
	300 µA	10 Hz	$38 \cdot 10^{-6} l$	
		20 Hz	$34 \cdot 10^{-6} l$	
		30 Hz; 40 Hz	$32 \cdot 10^{-6} l$	
		55 Hz	$42 \cdot 10^{-6} l$	
		400 Hz	$36 \cdot 10^{-6} l$	
		500 Hz; 1 kHz	$32 \cdot 10^{-6} l$	
		2 kHz; 5 kHz; 10 kHz	$33 \cdot 10^{-6} l$	
	1 mA	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$32 \cdot 10^{-6} l$ $32 \cdot 10^{-6} l$ $32 \cdot 10^{-6} l$	
	3 mA; 5 mA	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz 10 kHz	$10 \cdot 10^{-6} l$ $10 \cdot 10^{-6} l$ $10 \cdot 10^{-6} l$ $11 \cdot 10^{-6} l$	
	10 mA	10 Hz 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$9 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC power Measuring devices	20 mA; 30 mA; 50 mA; 100 mA	10 Hz 20 Hz; 30 Hz; 400 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz 5 kHz; 10 kHz	$10 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $8 \cdot 10^{-6} l$ $9 \cdot 10^{-6} l$	l = measured value
	200 mA; 300 mA; 500 mA	10 Hz 20 Hz; 30 Hz; 40 Hz, 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$10 \cdot 10^{-6} l$ $9 \cdot 10^{-6} l$ $9 \cdot 10^{-6} l$ $9 \cdot 10^{-6} l$	
	1 A; 2 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz, 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5kHz; 10 kHz	$11 \cdot 10^{-6} l$ $11 \cdot 10^{-6} l$ $11 \cdot 10^{-6} l$	
	3 A; 5 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5kHz; 10 kHz	$18 \cdot 10^{-6} l$ $18 \cdot 10^{-6} l$ $18 \cdot 10^{-6} l$	
	10 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5kHz; 10 kHz	$22 \cdot 10^{-6} l$ $22 \cdot 10^{-6} l$ $22 \cdot 10^{-6} l$	
	20 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5kHz; 10 kHz	$27 \cdot 10^{-6} l$ $27 \cdot 10^{-6} l$ $27 \cdot 10^{-6} l$ $31 \cdot 10^{-6} l$	
	50 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$33 \cdot 10^{-6} l$ $33 \cdot 10^{-6} l$ $33 \cdot 10^{-6} l$ $40 \cdot 10^{-6} l$	
	100 A	10 Hz; 20 Hz; 30 Hz; 40 Hz; 55 Hz; 400 Hz; 500 Hz; 1 kHz; 2 kHz; 5 kHz; 10 kHz	$48 \cdot 10^{-6} l$ $48 \cdot 10^{-6} l$ $48 \cdot 10^{-6} l$ $93 \cdot 10^{-6} l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC current strength sources and measuring devices (ranges)	100 µA to 1 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$120 \cdot 10^{-6} /$ $160 \cdot 10^{-6} /$ $60 \cdot 10^{-6} /$	/= measured value
	> 1 mA to 10 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$46 \cdot 10^{-6} /$	
	> 10 mA to 1 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$17 \cdot 10^{-6} /$	
	> 1 A to 10 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$32 \cdot 10^{-6} /$	
	> 10 A to 20 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$39 \cdot 10^{-6} /$	
	> 20 A to 100 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$69 \cdot 10^{-6} /$ $69 \cdot 10^{-6} /$ $0.17 \cdot 10^{-3} /$	
AC current strength (ranges) Measuring devices	0.1 mA to 0.2 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.40 \cdot 10^{-3} /$ $0.21 \cdot 10^{-3} /$ $0.40 \cdot 10^{-3} /$ $1.7 \cdot 10^{-3} /$	/= measured value with Fluke 5720A
	> 0.2 mA to 2.2 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.44 \cdot 10^{-3} /$ $0.30 \cdot 10^{-3} /$ $0.72 \cdot 10^{-3} /$ $4.2 \cdot 10^{-3} /$	
	> 2.2 mA to 22 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.44 \cdot 10^{-3} /$ $0.30 \cdot 10^{-3} /$ $0.46 \cdot 10^{-3} /$ $3.5 \cdot 10^{-3} /$	
	> 22 mA to 220 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.44 \cdot 10^{-3} /$ $0.25 \cdot 10^{-3} /$ $0.37 \cdot 10^{-3} /$ $1.6 \cdot 10^{-3} /$	
	> 220 mA to 2.2 A	20 Hz to 1 kHz > 1 kHz to 5 kHz > 5 kHz to 10 kHz	$0.43 \cdot 10^{-3} /$ $0.84 \cdot 10^{-3} /$ $7.6 \cdot 10^{-3} /$	
	> 2.2 A to 20 A	40 Hz to 5 kHz	$0.81 \cdot 10^{-3} / 1.2 \text{ mA}$	



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC power Sources	0.1 mA to 1 A	40 Hz to 5 kHz	$2 \cdot 10^{-3} /$	$I$ = measured value with HP3458A
Alternating current amperage current clamps	1 mA to 2.2 A	40 Hz to 5 kHz	$2 \cdot 10^{-3} /$	$I$ = measured value
	> 2.2 A to 20 A	40 Hz to 5 kHz	$3 \cdot 10^{-3} /$	
	> 20 A to 800 A	40 Hz to 65 Hz	$4 \cdot 10^{-3} /$	
AC current strength current transformer	1 A to 120 A	40 Hz to 850 Hz	$0.16 \cdot 10^{-3} /$	
	1 A to 120 A	> 850 Hz to 2 kHz	$0.47 \cdot 10^{-3} /$	
	> 120 A to 600 A	40 Hz to 400 Hz	$0.52 \cdot 10^{-3} /$	
	> 120 A to 1000 A	40 Hz to 65 Hz	$0.6 \cdot 10^{-3} /$	
Resistance ratio AC/DC measuring bridges	0.16 to 6.3	Equal and Alternating current up to 400 Hz	$0.2 \cdot 10^{-6}$	Uncertainty of measurement denotes absolute value
Phase angle between voltages Measuring devices and sources	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$ 50 mV / 50 mV measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz; 50 kHz 100 kHz	0,005° 0,008° 0,020°	$U_{SIG}$ : Signal voltage $U_{REF}$ : Reference voltage  For discrete measured values and frequencies
		$U_{REF} / U_{SIG}$ 0.5 V / 0.5 V 1 V / 1 V 0.8 V / 1 V 1 V / 0.5 V 10 V / 10 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz 1 kHz; 5 kHz 10 kHz 50 kHz 100 kHz	0,005° 0,005° 0,007° 0,008° 0,009°	
		$U_{REF} / U_{SIG}$ 100 V / 100 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz 1 kHz; 5 kHz 10 kHz 50 kHz 100 kHz	0,005° 0,005° 0,007° 0,008° 0,030°	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
Phase angle between tensions  Measuring devices and sources	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$		$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage  For discrete measured values and frequencies	
		1 V/0.05 V			
		10 V/1 V			
		1 V /10 V			
		100 V/1 V			
		1 V/100 V			
		Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz	0,009°		
		1 kHz; 5 kHz	0,009°		
		10 kHz	0,020°		
		50 kHz	0,030°		
		100 kHz	0,070°		
	-180° to 180	$U_{REF} / U_{SIG}$		$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage Measuring ranges	
		0.05 V to 0.5 V			
		Measuring frequency: 10 Hz to 1 kHz	0,006°		
		> 1 kHz to 50 kHz	0,010°		
		> 50 kHz to 100 kHz	0,025°		
		$U_{REF} / U_{SIG}$			
		> 0.5 V to 10 V			
		Measuring frequency: 10 Hz to 5 kHz	0,006°		
		> 5 kHz to 100 kHz	0,015°		
		$U_{REF} / U_{SIG}$			
		> 10 V to 100 V			
		Measuring frequency: 10 Hz to 5 kHz	0,006°		
> 5 kHz to 50 kHz	0,010°				
> 50 kHz to 100 kHz	0,035°				
$U_{REF} / U_{SIG}$					
> 100 V to 630 V					
Measuring frequency: 10 Hz to 2.5 kHz	0,008°				
> 2.5 kHz to 5 kHz	0,03°				
> 5 kHz to 10 kHz	0,04°				
> 10 kHz to 20 kHz	0,05°				
> 20 kHz to 50 kHz	0,1°				
> 50 kHz to 100 kHz	0,2°				

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage Sources	-180° to 180	$U_{REF} / U_{SIG}$ 1 mA to 2 A / 0.05 V to 100 V  Measuring frequency: 10 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 100 kHz	0,009° 0,045° 0,50°	$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage Measuring ranges
		$U_{REF} / U_{SIG}$ > 2 A to 20 A / 1 V to 100 V Measuring frequency: 10 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 100 kHz	0,02° 0,1° 1,0°	
		$U_{REF} / U_{SIG}$ > 20 A to 100 A / 1 V to 100 V Measuring frequency: 10 Hz to 1 kHz > 1 kHz to 10 kHz > 10 kHz to 100 kHz	0,025° 0,20° 2,0°	
		$U_{REF} / U_{SIG}$ 1 mA to 2 A / 0.05 V to 100 V Measuring frequency: 10 Hz to 1 kHz > 1 kHz to 10 kHz	0,009° 0,045°	
		$U_{REF} / U_{SIG}$ > 2 A to 20 A / 1 V to 100 V Measuring frequency: 10 Hz to 1 kHz > 1 kHz to 10 kHz	0,02° 0,1°	
		$U_{REF} / U_{SIG}$ > 20A to 100A / 1V to 100 V Measuring frequency: 10 Hz to 1 kHz > 1 kHz to 10 kHz	0,025° 0,20°	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage Sources	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$ 1 mA/0.0 5V 1 mA/0.5 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	0,007° 0,02° 0,08°	$U_{SIG}$ : Signal voltage $U_{REF}$ : Reference voltage For discrete measured values and frequencies
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$ 1 mA/1 V 10 mA/1 V 20 mA/1 V 50 mA/1 V 100 mA/1 V 200 mA/1 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	0,005° 0,010° 0,070°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$ 500 mA/1 V 1 A/1 V 2 A/1 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	0,006° 0,040° 0,40°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$ 5 A/1 V 10 A/1 V 20 A/1 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	0,010° 0,090° 0,90°	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage  Sources	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$	0,020°  0,15°  1,5°	$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage For discrete measured values and frequencies
		50 A/1 V		
		100 A/1 V		
		Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz		
0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$	0,006°  0,020°  0,080°	
		1 mA/10 V		
		10 mA/10 V 20 mA/10 V 50 mA/10 V 100 mA/10 V 200 mA/10 V		
		Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz		
0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$	0,007°  0,040°  0,40°	
		500 mA/10 V		
		1 A/10 V 2 A/10 V		
		Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz		
0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$	0,015°  0,09°  0,90°	
		5 A/10 V		
		10 A/10 V 20 A/10 V		
		Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz 5 kHz; 10 kHz 50 kHz; 100 kHz		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage  Sources	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  50 A/10 V  100 A/10 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	       0,020° 0,15° 1,5°	$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage For discrete measured values and frequencies
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  1 mA/100 V 10 mA/100 V 20 mA/100 V 50 mA/100 V 100 mA/100 V 200 mA/100 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	       0,008° 0,025° 0,09°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  500 mA/100 V 1 A/100 V 2 A/100 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	       0,007° 0,04° 0,40°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  5 A/100 V 10 A/100 V 20 A/100 V Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	       0,015° 0,09° 0,90°	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage  Sources	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  50 A/100 V  100 A/100 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz 50 kHz; 100 kHz	     0,020° 0,15° 1,5°	$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage For discrete measured values and frequencies
Phase angle between current and Voltage  Measuring devices	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  1 mA/0.05 V 1 mA/0.5 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz	     0,007° 0,02°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  1 mA/1 V 10 mA/1 V 20 mA/1 V 50 mA/1 V 100 mA/1 V 200 mA/1 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz	       0,005° 0,010°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  500 mA/1 V 1 A/1 V 2 A/1 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz 5 kHz; 10 kHz	       0,006° 0,040°	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage  Measuring devices	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  5 A/1 V  10 A/1 V  20 A/1 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz  5 kHz; 10 kHz	          0,010°  0,090°	$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage For discrete measured values and frequencies
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  50 A/1 V  100 A/1 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz  5 kHz; 10 kHz	          0,020°  0,15°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  1 mA/10 V  10 mA/10 V  20 mA/10 V  50 mA/10 V  100 mA/10 V  200 mA/10 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz  5 kHz; 10 kHz	          0,006°  0,020°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  500 mA/10 V  1 A/10 V  2 A/10 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz  5 kHz; 10 kHz	          0,007°  0,040°	



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage  Measuring devices	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  5 A/10 V  10 A/10 V  20 A/10 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz  5 kHz; 10 kHz	          0,015°  0,09°	$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage  For discrete measured values and frequencies
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  50 A/10 V  100 A/10 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz  5 kHz; 10 kHz	          0,020°  0,15°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  1 mA/100 V  10 mA/100 V  20 mA/100 V  50 mA/100 V  100 mA/100 V  200 mA/100 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1kHz  5 kHz; 10 kHz	          0,008°  0,025°	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase angle between current and Voltage  Measuring devices	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  500 mA/100 V  1 A/100 V  2 A/100 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz  5 kHz; 10 kHz	        0,007°  0,04°	$U_{SIG}$ : Signal voltage  $U_{REF}$ : Reference voltage  For discrete measured values and frequencies
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  5 A/100 V  10 A/100 V  20 A/100 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz  5 kHz; 10 kHz	        0,015°  0,09°	
	0°; 30°; 60°; 75°; 90°; 150°; 180°; 270°; 300°	$U_{REF} / U_{SIG}$  50 A/100 V  100 A/100 V  Measuring frequency: 10 Hz; 40 Hz; 55 Hz; 400 Hz; 1 kHz  5 kHz; 10 kHz	        0,020°  0,15°	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Inductance  Measuring devices, sources, Decadal normals	100 µH	1 kHz; 10 kHz	$0.3 \cdot 10^{-3}$	
	1 mH	1 kHz 10 kHz	$65 \cdot 10^{-6}$ $0.15 \cdot 10^{-3}$	
	10 mH	1 kHz 10 kHz	$55 \cdot 10^{-6}$ $0.16 \cdot 10^{-3}$	
	100 mH	1 kHz 10 kHz	$65 \cdot 10^{-6}$ $0.21 \cdot 10^{-3}$	
	1 H	100 Hz 1 kHz	$0.1 \cdot 10^{-3}$ $70 \cdot 10^{-6}$	
	10 H	100 Hz 1 kHz	$0.2 \cdot 10^{-3}$ $0.2 \cdot 10^{-3}$	
Inductance Measuring devices, Sources	0.1 mH to 1 mH	1 kHz to 10 kHz	$5.5 \cdot 10^{-3}$	
	> 1 mH to 10 mH	1 kHz to 10 kHz	$1.5 \cdot 10^{-3}$	
	> 10 mH to 100 mH	1 kHz to 10 kHz	$2.0 \cdot 10^{-3}$	
	> 0.1 H to 1 H	1 kHz	$0.4 \cdot 10^{-3}$	
	> 1 H to 10 H	100 Hz to 1 kHz	$1.5 \cdot 10^{-3}$	
Capacity gauges, Decadal normals	1 pF	50 Hz	$6 \cdot 10^{-3}$	
		100 Hz	$2.5 \cdot 10^{-3}$	
		1 kHz	$0.3 \cdot 10^{-3}$	
		10 kHz	$0.2 \cdot 10^{-3}$	
		400 kHz	$6.1 \cdot 10^{-3}$	
		100 kHz; 1 MHz	$3.5 \cdot 10^{-3}$	
	10 pF	50 Hz	$0.6 \cdot 10^{-3}$	
		100 Hz	$0.4 \cdot 10^{-3}$	
		1 kHz	$35 \cdot 10^{-6}$	
		10 kHz	$40 \cdot 10^{-6}$	
		100 kHz; 400 kHz	$75 \cdot 10^{-6}$	
		1 MHz	$110 \cdot 10^{-6}$	
	100 pF	50 Hz	$80 \cdot 10^{-6}$	
		100 Hz	$40 \cdot 10^{-6}$	
1 kHz; 10 kHz		$25 \cdot 10^{-6}$		
100 kHz		$35 \cdot 10^{-6}$		
400 kHz		$65 \cdot 10^{-6}$		
1 MHz		$0.35 \cdot 10^{-3}$		
1 nF	50 Hz	$35 \cdot 10^{-6}$		
	100 Hz	$25 \cdot 10^{-6}$		
	1 kHz	$15 \cdot 10^{-6}$		
	10 kHz	$22 \cdot 10^{-6}$		
	100 kHz	$75 \cdot 10^{-6}$		
	400 kHz 1 MHz	$0.45 \cdot 10^{-3}$ $3 \cdot 10^{-3}$		
10 nF	50 Hz; 100 Hz; 1 kHz	$35 \cdot 10^{-6}$		
	10 kHz	$55 \cdot 10^{-6}$		
100 nF	50 Hz; 100 Hz; 1 kHz	$55 \cdot 10^{-6}$		
	10 kHz	$75 \cdot 10^{-6}$		
1 µF	50 Hz	$55 \cdot 10^{-6}$		
	100 Hz	$70 \cdot 10^{-6}$		
	1 kHz	$55 \cdot 10^{-6}$		
	10 kHz	$110 \cdot 10^{-6}$		
10 µF	50 Hz; 100 Hz; 1 kHz	$0.2 \cdot 10^{-3}$		
	10 kHz	$0.4 \cdot 10^{-3}$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Capacity Sources, gauges, Normal	10 pF to 100 pF	50 Hz until 1 kHz > 1 kHz until 10 kHz > 10 kHz until 100 kHz > 100 kHz until 400 kHz > 400 kHz to 1 MHz	$4.9 \cdot 10^{-3} C$ $0.4 \cdot 10^{-3} C$ $3.0 \cdot 10^{-3} C$ $1.6 \cdot 10^{-3} C$ $2.2 \cdot 10^{-3} C$	C = measured value
	> 100 pF to 1 nF	50 Hz until 1 kHz > 1 kHz until 10 kHz > 10 kHz until 100 kHz > 100 kHz until 400 kHz > 400 kHz to 1 MHz	$0.2 \cdot 10^{-3} C$ $2.3 \cdot 10^{-3} C$ $3.5 \cdot 10^{-3} C$ $1.5 \cdot 10^{-3} C$ $3.6 \cdot 10^{-3} C$	
	> 1 nF to 10 nF	50 Hz to 1 kHz > 1 kHz to 10 kHz	$2.8 \cdot 10^{-3} C$ $2.4 \cdot 10^{-3} C$	
	> 10 nF to 100 nF	50 Hz to 1 kHz > 1 kHz to 10 kHz	$4.3 \cdot 10^{-3} C$ $2.0 \cdot 10^{-3} C$	
	> 100 nF to 1 µF	50 Hz to 100 Hz > 100 Hz until 1 kHz > 1 kHz until 10 kHz	$0.11 \cdot 10^{-3} C$ $1.0 \cdot 10^{-3} C$ $0.5 \cdot 10^{-3} C$	
	> 1 µF to 10 µF	50 Hz until 100 Hz > 100 Hz until 1 kHz > 1 kHz to 10 kHz	$0.58 \cdot 10^{-3} C$ $0.38 \cdot 10^{-3} C$ $0.43 \cdot 10^{-3} C$	
	Capacity Measuring devices	190 pF to < 400 pF	10 Hz to 10 kHz	
400 pF to < 1.1 nF		10 Hz to 10 kHz	$4.5 \cdot 10^{-3} C$ 8 pF	
1.1 nF to < 3.3 nF		10 Hz to 3 kHz	$4.0 \cdot 10^{-3} C$ 8 pF	
3.3 nF to < 11 nF		10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 8 pF	
11 nF to < 33 nF		10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
33 nF to < 110 nF		10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
110 nF to < 330 nF		10 Hz to 1 kHz	$4.5 \cdot 10^{-3} C$	
330 nF to < 1.1 µF		10 Hz to 600 Hz	$4.5 \cdot 10^{-3} C$	
1.1 µF to < 3.3 µF		10 Hz to 300 Hz	$4.5 \cdot 10^{-3} C$	
3.3 µF to < 11 µF		10 Hz to 150 Hz	$4.5 \cdot 10^{-3} C$	
11 µF to < 33 µF	10 Hz to 120 Hz	$6.0 \cdot 10^{-3} C$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Capacity Measuring devices	33 µF to < 110 µF	10 Hz to 80 Hz DC to 50	$6.5 \cdot 10^{-3} C$	C = measured value
	110 µF to < 330 µF	Hz DC to Hz DC to 6	$6.0 \cdot 10^{-3} C$	
	330 µF to < 1.1 mF	Hz DC to 2 Hz DC to	$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	200.6 Hz	$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 0,2 Hz	$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF		$8.0 \cdot 10^{-3} C$	
	33 mF to 110 mF		$11 \cdot 10^{-3} C$	
DC capacity Sources, measuring devices	1 µF to 70 µF	DC method	$2.6 \cdot 10^{-3}$	
	> 70 µF to 200 µF		$0.55 \cdot 10^{-3}$	
	> 200 µF to 110 mF		$0.30 \cdot 10^{-3}$	
Stress ratio	± 2 mV/V	Bridge voltage: 5 V	0.04 µV/V	Calibration of 350 Ω bridge standards and the associated indicators  at discrete points in 10% steps
		Measuring frequency 225 Hz	0.05 µV/V	
		Measuring frequency 600 Hz	1.0 µV/V	
	± 2 mV/V	Bridge voltage: 2,5 V	0.05 µV/V	
		Measuring frequency 225 Hz	0.05 µV/V	
	± 5 mV/V	Measuring frequency 600 Hz	1.0 µV/V	
		Measuring frequency 4.8 kHz		
± 5 mV/V	Bridge voltage: 5 V	0.15 µV/V		
	Measuring frequency 225 Hz	1.0 µV/V		
± 10 mV/V	Measuring frequency 4.8 kHz			
	Bridge voltage: 5 V	0.10 µV/V		
± 10 mV/V	Measuring frequency 225 Hz	0.30 µV/V		
	Measuring frequency 600 Hz			
	Measuring frequency 4.8 kHz			
± 5 mV/V	Bridge voltage: 2,5 V	0.1 µV/V		
	Measuring frequency 225 Hz	0.1 µV/V		
	Measuring frequency 600 Hz	1.0 µV/V		
± 5 mV/V	Measuring frequency 4.8 kHz			
	Bridge voltage: 2,5 V	0.4 µV/V		
	Measuring frequency 225 Hz	0.4 µV/V		
± 10 mV/V	Measuring frequency 600 Hz	0.4 µV/V		
	Measuring frequency 4.8 kHz			
	Bridge voltage: 1 V	0.40 µV/V		
± 10 mV/V	Measuring frequency 600 Hz			
	Measuring frequency 600 Hz			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Stress ratio	± 20 mV/V	Bridge voltage: 1 V  Measuring frequency 4.8 kHz	0.60 µV/V	Calibration of 350 Ω bridge standards and the associated indicators  at discrete points in 10% steps
	± 100 mV/V	Bridge voltage: 1 V  Measuring frequency 4.8 kHz	5.0 µV/V	
	± 100 mV/V	Bridge voltage: 2,5 V  Measuring frequency 4.8 kHz	5.0 µV/V	
DC voltage Bridge standards	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage Bridge voltage:  0,5 V	0.4 µV/V 0.35 µV/V 0.35 µV/V 0.35 µV/V 0.35 µV/V 0.35 µV/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage Bridge voltage:  1,0 V	0.2 µV/V 0.15 µV/V 0.15 µV/V 0.15 µV/V 0.15 µV/V 0.25 µV/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage Bridge voltage:  2,5 V	0.1 µV/V 0.07 µV/V 0.07 µV/V 0.07 µV/V 0.07 µV/V 0.20 µV/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage Bridge voltage:  5,0 V	0.04 µV/V 0.035 µV/V 0.035 µV/V 0.035 µV/V 0.045 µV/V 0.15 µV/V	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Voltage ratio DC voltage bridge standards	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage bridge voltage:  7,5 V	0.025 µV/V 0.025 µV/V 0.025 µV/V 0.025 µV/V 0.04 µV/V 0.15 µV/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage Bridge voltage:  10,0 V	0.02 µV/V 0.015 µV/V 0.020 µV/V 0.025 µV/V 0.035 µV/V 0.075 µV/V	
DC voltage bridges, measuring instruments, measuring amplifiers	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage bridge voltage:  0,5 V	0.35 µV/V 0.35 µV/V 0.40 µV/V 0.55 µV/V 2.5 µV/V	With K148
	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage Bridge voltage:  1 V	0.20 µV/V 0.20 µV/V 0.30 µV/V 0.50 µV/V 2.5 µV/V	
	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	DC voltage Bridge voltage:  2.5 V; 5 V; 7.5 V; 10 V	0.10 µV/V 0.15 µV/V 0.25 µV/V 0.45 µV/V 2.5 µV/V	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power		1 mA / 0.05 V // 10 Hz to 10 kHz		<i>P</i> = AC active power
Sources and measuring devices (discrete points)	50.0 µW	Phase angle: 0 °	0,2 - 10 <sup>-3</sup> <i>P</i>	Discrete points for  Current, voltage and Phase angle
	43.3 µW	Phase angle: ±30 °	0,3 - 10 <sup>-3</sup> <i>P</i>	
	25.0 µW	Phase angle: ±60 °	0,8 - 10 <sup>-3</sup> <i>P</i>	
	12.9 µW	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	500 µW	1 mA / 0.5 V / 10 Hz to 10 kHz Phase angle: 0 °	0,2 - 10 <sup>-3</sup> <i>P</i>	
	433 µW	Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>	
	250 µW	Phase angle: ±60 °	0,4 - 10 <sup>-3</sup> <i>P</i>	
	129 µW	Phase angle: ±75 °	0,7 - 10 <sup>-3</sup> <i>P</i>	
	1.0 mW	1 mA / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	0,1 - 10 <sup>-3</sup> <i>P</i>	
	0.9 mW	Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>	
	0.5 mW	Phase angle: ±60 °	0,4 - 10 <sup>-3</sup> <i>P</i>	
	0.3 mW	Phase angle: ±75 °	0,8 - 10 <sup>-3</sup> <i>P</i>	
	10.0 mW	10 mA / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	0,1 - 10 <sup>-3</sup> <i>P</i>	
	8.7 mW	Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>	
	5.0 mW	Phase angle: ±60 °	0,4 - 10 <sup>-3</sup> <i>P</i>	
	2.6 mW	Phase angle: ±75 °	0,8 - 10 <sup>-3</sup> <i>P</i>	
20.0 mW	20 mA / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
17.3 mW	Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
10.0 mW	Phase angle: ±60 °	0,4 - 10 <sup>-3</sup> <i>P</i>		
5.2 mW	Phase angle: ±75 °	0,8 - 10 <sup>-3</sup> <i>P</i>		
50.0 mW	50 mA / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
47.3 mW	Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
25.0 mW	Phase angle: ±60 °	0,4 - 10 <sup>-3</sup> <i>P</i>		
12.9 mW	Phase angle: ±75 °	0,8 - 10 <sup>-3</sup> <i>P</i>		
100.0 mW	100 mA / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
86.6 mW	Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
50.0 mW	Phase angle: ±60 °	0,4 - 10 <sup>-3</sup> <i>P</i>		
25.9 mW	Phase angle: ±75 °	0,8 - 10 <sup>-3</sup> <i>P</i>		



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power  Sources and measuring devices (discrete points)	200.0 mW	200 mA / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> P	P = AC active power  Discrete points for  Current, voltage and Phase angle
	173.2 mW	Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> P	
	100.0 mW	Phase angle: ±60 °	0,4 - 10 <sup>-3</sup> P	
	51.8 mW	Phase angle: ±75 °	0,8 - 10 <sup>-3</sup> P	
	500.0 mW	500 mA / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> P	
	433.0 mW	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> P	
	250.0 mW	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> P	
	129.4 mW	Phase angle: ±75 °	3,0 - 10 <sup>-3</sup> P	
	1,0 W	1 A / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> P	
	0,9 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> P	
	0,5 W	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> P	
	0,3 W	Phase angle: ±75 °	3,0 - 10 <sup>-3</sup> P	
	2,0 W	2 A / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> P	
	1,7 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> P	
	1,0 W	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> P	
	0,5 W	Phase angle: ±75 °	3,0 - 10 <sup>-3</sup> P	
	5,0 W	5 A / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> P	
	4,3 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> P	
	2,5 W	Phase angle: ±60 °	4,0 - 10 <sup>-3</sup> P	
	1,3 W	Phase angle: ±75 °	8,0 - 10 <sup>-3</sup> P	
	10,0 W	10 A / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> P	
	8,7 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> P	
	5,0 W	Phase angle: ±60 °	4,0 - 10 <sup>-3</sup> P	
	2,6 W	Phase angle: ±75 °	8,0 - 10 <sup>-3</sup> P	
20,0 W	20 A / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> P		
17,3 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> P		
10,0 W	Phase angle: ±60 °	4,0 - 10 <sup>-3</sup> P		
5,2 W	Phase angle: ±75 °	8,0 - 10 <sup>-3</sup> P		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power		50 A / 1 V / 10 Hz to 10 kHz		<i>P</i> = AC active power
Sources and measuring devices (discrete points)	50,0 W	Phase angle: 0 °	0,3 - 10 <sup>-3</sup> <i>P</i>	Discrete points for  Current, voltage and Phase angle
	43,3 W	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	25,0 W	Phase angle: ±60 °	6,0 - 10 <sup>-3</sup> <i>P</i>	
	12,9 W	Phase angle: ±75 °	10 - 10 <sup>-3</sup> <i>P</i>	
	100,0 W	100 A / 1 V / 10 Hz to 10 kHz Phase angle: 0 °	0,3 - 10 <sup>-3</sup> <i>P</i>	
	86,6 W	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	50,0 W	Phase angle: ±60 °	6,0 - 10 <sup>-3</sup> <i>P</i>	
	25,9 W	Phase angle: ±75 °	10 - 10 <sup>-3</sup> <i>P</i>	
	10.0 mW	1 mA / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	0,1 - 10 <sup>-3</sup> <i>P</i>	
	8.7 mW	Phase angle: ±30 °	0,3 - 10 <sup>-3</sup> <i>P</i>	
	5.0 mW	Phase angle: ±60 °	0,7 - 10 <sup>-3</sup> <i>P</i>	
	2.6 mW	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	100.0 mW	10 mA / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	86.6 mW	Phase angle: ±30 °	0,3 - 10 <sup>-3</sup> <i>P</i>	
	50.0 mW	Phase angle: ±60 °	0,7 - 10 <sup>-3</sup> <i>P</i>	
	25.9 mW	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
200.0 mW	20 mA / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
173.2 mW	Phase angle: ±30 °	0,3 - 10 <sup>-3</sup> <i>P</i>		
100.0 mW	Phase angle: ±60 °	0,7 - 10 <sup>-3</sup> <i>P</i>		
51.8 mW	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>		
500.0 mW	50 mA / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
433.0 mW	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>		
250.0 mW	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
129.4 mW	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>		
1,0 W	100 mA / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
0,9 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>		
0,5 W	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
0,3 W	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power		200 mA / 10 V / 10 Hz to 10 kHz		<i>P</i> = AC active power
Sources and measuring devices (discrete points)	2,0 W	Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	Discrete points for  Current, voltage and Phase angle
	1,7 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>	
	1,0 W	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>	
	0,5 W	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	10,0 W	1 A / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	8,7 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>	
	5,0 W	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	2,6 W	Phase angle: ±75 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	20,0 W	2 A / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	17,3 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>	
	10,0 W	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	5,2 W	Phase angle: ±75 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	50,0 W	5 A / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	43,3 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> <i>P</i>	
	25,0 W	Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	12,9 W	Phase angle: ±75 °	7,0 - 10 <sup>-3</sup> <i>P</i>	
100,0 W	10 A / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
86,6 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
50,0 W	Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> <i>P</i>		
25,9 W	Phase angle: ±75 °	7,0 - 10 <sup>-3</sup> <i>P</i>		
200,0 W	20 A / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
173,2 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
100,0 W	Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> <i>P</i>		
51,8 W	Phase angle: ±75 °	7,0 - 10 <sup>-3</sup> <i>P</i>		
500,0 W	50 A / 10 V / 10 Hz to 10 kHz Phase angle: 0 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
433,0 W	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> <i>P</i>		
250,0 W	Phase angle: ±60 °	5,0 - 10 <sup>-3</sup> <i>P</i>		
129,4 W	Phase angle: ±75 °	1,0 - 10 <sup>-2</sup> <i>P</i>		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power		100 A / 10 V / 10 Hz to 10 kHz		<i>P</i> = AC active power
Sources and measuring devices (discrete points)	1000,0 W	Phase angle: 0 °	0,2 - 10 <sup>-3</sup> <i>P</i>	Discrete points for  Current, voltage and Phase angle
	866,0 W	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	500,0 W	Phase angle: ±60 °	5,0 - 10 <sup>-3</sup> <i>P</i>	
	258,8 W	Phase angle: ±75 °	1,0 - 10 <sup>-2</sup> <i>P</i>	
	100.0 mW	1 mA / 100 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	86.6 mW	Phase angle: ±30 °	0,3 - 10 <sup>-3</sup> <i>P</i>	
	50.0 mW	Phase angle: ±60 °	0,8 - 10 <sup>-3</sup> <i>P</i>	
	25.9 mW	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	1,0 W	10 mA / 100 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	0,9 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>	
	0,5 W	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>	
	0,3 W	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	2,0 W	20 mA / 100 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	1,7 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>	
	1,0 W	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>	
	0,5 W	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
5,0 W	50 mA / 100 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
4,3 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>		
2,5 W	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
1,3 W	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>		
10,0 W	100 mA / 100 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
8,7 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>		
5,0 W	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
2,6 W	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>		
20,0 W	200 mA / 100 V / 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
17,3 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>		
10,0 W	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
5,2 W	Phase angle: ±75 °	2,0 - 10 <sup>-3</sup> <i>P</i>		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power		1 A / 100 V // 10 Hz to 10 kHz		<i>P</i> = AC active power
Sources and measuring devices (discrete points)	100,0 W	Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	Discrete points for  Current, voltage and Phase angle
	86,6 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>	
	50,0 W	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	25,9 W	Phase angle: ±75 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	200,0 W	Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	173,2 W	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> <i>P</i>	
	100,0 W	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> <i>P</i>	
	51,8 W	Phase angle: ±75 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	500,0 W	Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	433,0 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> <i>P</i>	
	250,0 W	Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	129,4 W	Phase angle: ±75 °	7,0 - 10 <sup>-3</sup> <i>P</i>	
	1000 W	10 A / 100 V // 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>	
	866 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> <i>P</i>	
	500 W	Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> <i>P</i>	
	258,8 W	Phase angle: ±75 °	7,0 - 10 <sup>-3</sup> <i>P</i>	
2000 W	20 A / 100 V // 10 Hz to 10 kHz Phase angle: 0 °	50 - 10 <sup>-6</sup> <i>P</i>		
1732 W	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
1000 W	Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> <i>P</i>		
517,6 W	Phase angle: ±75 °	7,0 - 10 <sup>-3</sup> <i>P</i>		
5000 W	50 A / 100 V // 10 Hz to 10 kHz Phase angle: 0 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
4330 W	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> <i>P</i>		
2500 W	Phase angle: ±60 °	5,0 - 10 <sup>-3</sup> <i>P</i>		
1294 W	Phase angle: ±75 °	1,0 - 10 <sup>-2</sup> <i>P</i>		
10000 W	100 A / 100 V // 10 Hz to 10 kHz Phase angle: 0 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
8660 W	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> <i>P</i>		
5000 W	Phase angle: ±60 °	5,0 - 10 <sup>-3</sup> <i>P</i>		
2588 W	Phase angle: ±75 °	10 - 10 <sup>-3</sup> <i>P</i>		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
AC active power  Sources and measuring devices (discrete points)	25 kW  21.7 kW 12,5 kW 6,5 kW	50 A / 500 V / 40 Hz to 850 Hz		<i>P</i> = AC active power  Discrete points for  Current, voltage and Phase angle	
		Phase angle: 0 °	85 - 10 <sup>-6</sup> <i>P</i>		
		Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
		Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> <i>P</i>		
		Phase angle: ±75 °	1,0 - 10 <sup>-3</sup> <i>P</i>		
	40 kW 34.6 kW 20 kW 10,4 kW	80 A / 500 V / 40 Hz up to 850 Hz			
		Phase angle: 0 °	85 - 10 <sup>-6</sup> <i>P</i>		
		Phase angle: ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
Phase angle: ±60 °		0,5 - 10 <sup>-3</sup> <i>P</i>			
AC active power sources and meters (ranges)	50 µW to 500 mW	50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: 0 °	0,17 - 10 <sup>-3</sup> <i>P</i>		
		50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: 0° to ±30 °	0,2 - 10 <sup>-3</sup> <i>P</i>		
	50 µW to 500 mW	50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: 30 ° to 60 ° and -30 ° to -60 °	0,3 - 10 <sup>-3</sup> <i>P</i>		
		50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: 60 ° to 75 ° and -60 ° to -75 °	0,6 - 10 <sup>-3</sup> <i>P</i>		
	> 500 mW to 500 W	5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: 0 °	0,1 - 10 <sup>-3</sup> <i>P</i>		
		5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: 0° to ±30 °	0,1 - 10 <sup>-3</sup> <i>P</i>		
		5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: ±30 ° to ±60 °	0,3 - 10 <sup>-3</sup> <i>P</i>		
		5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: ±60 ° to ±75 °	0,6 - 10 <sup>-3</sup> <i>P</i>		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
AC active power Sources and measuring devices (ranges)	> 500 W to 10 kW	25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: 0 °	0,1 - 10 <sup>-3</sup> P	P = AC active power	
		25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: 0° to ±30 °	0,15 - 10 <sup>-3</sup> P		
		25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: ±30 ° to ± 60 °	0,35 - 10 <sup>-3</sup> P		
		25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: ±60 ° to ± 75 °	0,75 - 10 <sup>-3</sup> P		
	50 µW to 500 µW	0.05 V to 0.5 V 1 mA >1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75°	0,3 - 10 <sup>-3</sup> P 0,8 - 10 <sup>-3</sup> P 7,0 - 10 <sup>-3</sup> P		
	> 0.5 mW to 200 mW	0.5 V to 1 V 1 mA to 200 mA > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75°	0,2 - 10 <sup>-3</sup> P 0,4 - 10 <sup>-3</sup> P 0,8 - 10 <sup>-3</sup> P		
	> 200 mW to 20 W	0.1 V to 1 V > 200 mA to 20 A > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75°	1,0 - 10 <sup>-3</sup> P 4,0 - 10 <sup>-3</sup> P 8,0 - 10 <sup>-3</sup> P		
	> 20 W to 100 W	0.1 V to 1 V > 20 A to 100 A >1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75°	3,0 - 10 <sup>-3</sup> P 6,0 - 10 <sup>-3</sup> P 10 - 10 <sup>-3</sup> P		
	10 mW to 200 mW	> 1 V to 10 V 1 mA to < 200 mA > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30 ° to ±60° Phase angle: >±60° to ±75°	0,3 - 10 <sup>-3</sup> P 0,7 - 10 <sup>-3</sup> P 2,0 - 10 <sup>-3</sup> P		
	> 200 mW to 20 W	> 1 V to 10 V > 200 mA to 20 A > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30 ° to ±60° Phase angle: >±60° to ±75°	0,5 - 10 <sup>-3</sup> P 2,0 - 10 <sup>-3</sup> P 3,0 - 10 <sup>-3</sup> P		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power Sources and measuring devices (ranges)	> 20 W to 1000 W	> 1 V to 10 V > 20 A to 100 A > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75	2,0 - 10 <sup>-3</sup> P 5,0 - 10 <sup>-3</sup> P 10 - 10 <sup>-3</sup> P	P = AC active power
	100 mW to 20 W	> 10 V to 100 V 1 mA to 200 mA > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75	0,5 - 10 <sup>-3</sup> P 1,0 - 10 <sup>-3</sup> P 2,0 - 10 <sup>-3</sup> P	
	> 20 W to 1000 W	> 10 V to 100 V > 200 mA to 10 A > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75	1,5 - 10 <sup>-3</sup> P 3,0 - 10 <sup>-3</sup> P 7,0 - 10 <sup>-3</sup> P	
	> 1 kW to 10 kW	> 10 V to 100 V > 10 A to 100 A > 1 kHz to 10 kHz Phase angle: 0° to ±30° Phase angle: >±30° to ±60° Phase angle: >±60° to ±75	2,0 - 10 <sup>-3</sup> P 5,0 - 10 <sup>-3</sup> P 10 - 10 <sup>-3</sup> P	
	> 10 kW to 80 kW	500 V ≤ U ≤ 1000 V 20 A ≤ I ≤ 80 A 40 Hz to 850 Hz Phase angle: 0°	85 - 10 <sup>-6</sup> P	
		500 V ≤ U ≤ 1000 V 20 A ≤ I ≤ 80 A 40 Hz to 850 Hz Phase angle: 0° to ±30°	0,25 - 10 <sup>-3</sup> P	
		500 V ≤ U ≤ 1000 V 20 A ≤ I ≤ 80 A 40 Hz to 850 Hz Phase angle: 30° to 60° Phase angle: -30° to -60°	0,65 - 10 <sup>-3</sup> P	
		500 V ≤ U ≤ 1000 V 20 A ≤ I ≤ 80 A 40 Hz to 850 Hz Phase angle: 60° to 75° Phase angle: -60° to -75°	1,5 - 10 <sup>-3</sup> P	



Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC reactive power Sources and measuring devices (discrete points)	50.0 $\mu$ VAr	1 mA / 0.05 V / 10 Hz to 10 kHz Phase angle: 90 °	0,2 - 10 <sup>-3</sup> Q	Q=AC reactive power  Discrete points for current, voltage and phase angle
	43.3 $\mu$ VAr	Phase angle: $\pm 60$ °	0,3 - 10 <sup>-3</sup> Q	
	25.0 $\mu$ VAr	Phase angle: $\pm 30$ °	0,8 - 10 <sup>-3</sup> Q	
	12.9 $\mu$ VAr	Phase angle: $\pm 15$ °	2,0 - 10 <sup>-3</sup> Q	
	500 $\mu$ VAr	1 mA / 0.5 V / 10 Hz to 10 kHz Phase angle: 90 °	0,2 - 10 <sup>-3</sup> Q	
	433 $\mu$ VAr	Phase angle: $\pm 60$ °	0,2 - 10 <sup>-3</sup> Q	
	250 $\mu$ VAr	Phase angle: $\pm 30$ °	0,4 - 10 <sup>-3</sup> Q	
	129 $\mu$ VAr	Phase angle: $\pm 15$ °	0,7 - 10 <sup>-3</sup> Q	
	1.0 mVAr	1 mA / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	0,1 - 10 <sup>-3</sup> Q	
	0.9 mVAr	Phase angle: $\pm 60$ °	0,2 - 10 <sup>-3</sup> Q	
	0.5 mVAr	Phase angle: $\pm 30$ °	0,4 - 10 <sup>-3</sup> Q	
	0.3 mVAr	Phase angle: $\pm 15$ °	0,8 - 10 <sup>-3</sup> Q	
	10.0 mVAr	10 mA / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	0,1 - 10 <sup>-3</sup> Q	
	8.7 mVAr	Phase angle: $\pm 60$ °	0,2 - 10 <sup>-3</sup> Q	
	5.0 mVAr	Phase angle: $\pm 30$ °	0,4 - 10 <sup>-3</sup> Q	
	2.6 mVAr	Phase angle: $\pm 15$ °	0,8 - 10 <sup>-3</sup> Q	
	20.0 mVAr	20 mA / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	17.3 mVAr	Phase angle: $\pm 60$ °	0,2 - 10 <sup>-3</sup> Q	
	10.0 mVAr	Phase angle: $\pm 30$ °	0,4 - 10 <sup>-3</sup> Q	
	5.2 mVAr	Phase angle: $\pm 15$ °	0,8 - 10 <sup>-3</sup> Q	
	50.0 mVAr	50 mA / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	47.3 mVAr	Phase angle: $\pm 60$ °	0,2 - 10 <sup>-3</sup> Q	
	25.0 mVAr	Phase angle: $\pm 30$ °	0,4 - 10 <sup>-3</sup> Q	
	12.9 mVAr	Phase angle: $\pm 15$ °	0,8 - 10 <sup>-3</sup> Q	
	100.0 mVAr	100 mA / 1 V 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	86.6 mVAr	Phase angle: $\pm 60$ °	0,2 - 10 <sup>-3</sup> Q	
	50.0 mVAr	Phase angle: $\pm 30$ °	0,4 - 10 <sup>-3</sup> Q	
	25.9 mVAr	Phase angle: $\pm 15$ °	0,8 - 10 <sup>-3</sup> Q	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC reactive power Sources and measuring devices (discrete points)	200.0 mVAr	200 mA / 1 V / 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	Q=AC reactive power Discrete points for current, voltage and phase angle
	173.2 mVAr	Phase angle: ±60 °	0,2 - 10 <sup>-3</sup> Q	
	100.0 mVAr	Phase angle: ±30 °	0,4 - 10 <sup>-3</sup> Q	
	51.8 mVAr	Phase angle: ±15 °	0,8 - 10 <sup>-3</sup> Q	
	500.0 mVAr	500 mA / 1 V / 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	433.0 mVAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	250.0 mVAr	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> Q	
	129.4 mVAr	Phase angle: ±15 °	3,0 - 10 <sup>-3</sup> Q	
	1.0 VAr	1 A / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	0.9 VAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	0.5 VAr	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> Q	
	0.3 VAr	Phase angle: ±15 °	3,0 - 10 <sup>-3</sup> Q	
	2.0 VAr	2 A / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	1.7 VAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	1.0 VAr	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> Q	
	0.5 VAr	Phase angle: ±15 °	3,0 - 10 <sup>-3</sup> Q	
	5.0 VAr	5 A / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	4.3 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q	
	2.5 VAr	Phase angle: ±30 °	4,0 - 10 <sup>-3</sup> Q	
	1.3 VAr	Phase angle: ±15 °	8,0 - 10 <sup>-3</sup> Q	
	10.0 VAr	10 A / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	8.7 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q	
	5.0 VAr	Phase angle: ±30 °	4,0 - 10 <sup>-3</sup> Q	
	2.6 VAr	Phase angle: ±15 °	8,0 - 10 <sup>-3</sup> Q	
20.0 VAr	20 A / 1 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q		
17.3 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q		
10.0 VAr	Phase angle: ±30 °	4,0 - 10 <sup>-3</sup> Q		
5.2 VAr	Phase angle: ±15 °	8,0 - 10 <sup>-3</sup> Q		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks	
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement		
AC reactive power Sources and measuring devices (discrete points)	50.0 VAr 43.3 VAr 25.0 VAr 12.9 VAr	50 A / 1 V // 10 Hz to 10 kHz		Q=AC reactive power  Discrete points for current, voltage and phase angle	
		Phase angle: 90 °	0,3 - 10 <sup>-3</sup> Q		
		Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> Q		
		Phase angle: ±30 °	6,0 - 10 <sup>-3</sup> Q		
	Phase angle: ±15 °	10 - 10 <sup>-3</sup> Q			
	100.0 VAr 86.6 VAr 50.0 VAr 25.9 VAr	100 A / 1 V // 10 Hz to 10 kHz			
		Phase angle: 90 °	0,3 - 10 <sup>-3</sup> Q		
		Phase angle: ±60 °	3,0 - 10 <sup>-3</sup> Q		
		Phase angle: ±30 °	6,0 - 10 <sup>-3</sup> Q		
	Phase angle: ±15 °	10 - 10 <sup>-3</sup> Q			
	10.0 mVAr 8.7 mVAr 5.0 mVAr 2.6 mVAr	1 mA / 10 V // 10 Hz to 10 kHz			
		Phase angle: 90 °	0,1 - 10 <sup>-3</sup> Q		
		Phase angle: ±60 °	0,3 - 10 <sup>-3</sup> Q		
		Phase angle: ±30 °	0,7 - 10 <sup>-3</sup> Q		
	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q			
	100.0 mVAr 86.6 mVAr 50.0 mVAr 25.9 mVAr	10 mA / 10 V / 10 Hz to 10 kHz			
		Phase angle: 90 °	50 - 10 <sup>-6</sup> Q		
		Phase angle: ±60 °	0,3 - 10 <sup>-3</sup> Q		
		Phase angle: ±30 °	0,7 - 10 <sup>-3</sup> Q		
	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q			
	200.0 mVAr 173.2 mVAr 100.0 mVAr 51.8 mVAr	20 mA / 10 V / 10 Hz to 10 kHz			
		Phase angle: 90 °	50 - 10 <sup>-6</sup> Q		
		Phase angle: ±60 °	0,3 - 10 <sup>-3</sup> Q		
		Phase angle: ±30 °	0,7 - 10 <sup>-3</sup> Q		
	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q			
	500.0 mVAr 433.0 mVAr 250.0 mVAr 129.4 mVAr	50 mA / 10 V / 10 Hz to 10 kHz			
		Phase angle: 90 °	50 - 10 <sup>-6</sup> Q		
		Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q		
Phase angle: ±30 °		1,0 - 10 <sup>-3</sup> Q			
Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q				
1.0 VAr 0.9 VAr 0.5 VAr 0.3 VAr	100 mA / 10 V / 10 Hz to 10 kHz				
	Phase angle: 90 °	50 - 10 <sup>-6</sup> Q			
	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q			
	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> Q			
Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q				

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC reactive power Sources and measuring devices (discrete points)	2.0 VAr	200 mA / 10 V / 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	Q=AC reactive power  Discrete points for current, voltage and phase angle
	1.7 VAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	1.0 VAr	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> Q	
	0.5 VAr	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q	
	10.0 VAr	1 A / 10 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	8.7 VAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	5.0 VAr	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> Q	
	2.6 VAr	Phase angle: ±15 °	3,0 - 10 <sup>-3</sup> Q	
	20.0 VAr	2 A / 10 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	17.3 VAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	10.0 VAr	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> Q	
	5.2 VAr	Phase angle: ±15 °	3,0 - 10 <sup>-3</sup> Q	
	50.0 VAr	5 A / 10 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	43.3 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q	
	25.0 VAr	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> Q	
	12.9 VAr	Phase angle: ±15 °	7,0 - 10 <sup>-3</sup> Q	
	100.0 VAr	10 A / 10 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	86.6 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q	
	50.0 VAr	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> Q	
	25.9 VAr	Phase angle: ±15 °	7,0 - 10 <sup>-3</sup> Q	
	200.0 VAr	20 A / 10 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	173.2 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q	
	100.0 VAr	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> Q	
	51.8 VAr	Phase angle: ±15 °	7,0 - 10 <sup>-3</sup> Q	
	500.0 VAr	50 A / 10 V // 10 Hz to 10 kHz Phase angle: 90 °	0,2 - 10 <sup>-3</sup> Q	
	433.0 VAr	Phase angle: ±60 °	2,0 - 10 <sup>-2</sup> Q	
	250.0 VAr	Phase angle: ±30 °	2,0 - 10 <sup>-2</sup> Q	
	129.4 VAr	Phase angle: ±15 °	1,0 - 10 <sup>-2</sup> Q	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC reactive power Sources and measuring devices (discrete points)	1000.0 VAR	100 A / 10 V 10 Hz to 10 kHz Phase angle: 90 °	0,2 - 10 <sup>-3</sup> Q	Q=AC reactive power  Discrete points for current, voltage and phase angle
	866.0 VAR	Phase angle: ±60 °	2,0 - 10 <sup>-2</sup> Q	
	500.0 VAR	Phase angle: ±30 °	2,0 - 10 <sup>-2</sup> Q	
	258.8 VAR	Phase angle: ±15 °	1,0 - 10 <sup>-2</sup> Q	
	100.0 m VAR	1 mA / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	86.6 m VAR	Phase angle: ±60 °	0,3 - 10 <sup>-3</sup> Q	
	50.0 m VAR	Phase angle: ±30 °	0,8 - 10 <sup>-3</sup> Q	
	25.9 m VAR	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q	
	1.0 VAR	10 mA / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	0.9 VAR	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	0.5 VAR	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> Q	
	0.3 VAR	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q	
	2.0 VAR	20 mA / 100 V // 1 0 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	1.7 VAR	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	1.0 VAR	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> Q	
	0.5 VAR	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q	
	5.0 VAR	50 mA / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	4.3 VAR	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	2.5 VAR	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> Q	
	1.3 VAR	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q	
	10.0 VAR	100 mA / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	8.7 VAR	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	5.0 VAR	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> Q	
	2.6 VAR	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q	
	20.0 VAR	200 mA / 100 V // 0 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	17.3 VAR	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	10.0 VAR	Phase angle: ±30 °	1,0 - 10 <sup>-3</sup> Q	
	5.2 VAR	Phase angle: ±15 °	2,0 - 10 <sup>-3</sup> Q	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC reactive power Sources and measuring devices (discrete points)		1 A / 100 V // 10 Hz to 10 kHz		Q=AC reactive power
	100.0 VAr	Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	Discrete points for current, voltage and phase angle
	86.6 VAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	50.0 VAr	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> Q	
	25.9 VAr	Phase angle: ±15 °	3,0 - 10 <sup>-3</sup> Q	
	200.0 VAr	2 A / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	173.2 VAr	Phase angle: ±60 °	0,5 - 10 <sup>-3</sup> Q	
	100.0 VAr	Phase angle: ±30 °	2,0 - 10 <sup>-3</sup> Q	
	51.8 VAr	Phase angle: ±15 °	3,0 - 10 <sup>-3</sup> Q	
	500.0 VAr	5 A / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	433.0 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q	
	250.0 VAr	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> Q	
	129.4 VAr	Phase angle: ±15 °	7,0 - 10 <sup>-3</sup> Q	
	1000 VAr	10 A / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q	
	866 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q	
	500 VAr	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> Q	
258.8 VAr	Phase angle: ±15 °	7,0 - 10 <sup>-3</sup> Q		
2000 VAr	20 A / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	50 - 10 <sup>-6</sup> Q		
1732 VAr	Phase angle: ±60 °	1,0 - 10 <sup>-3</sup> Q		
1000 VAr	Phase angle: ±30 °	3,0 - 10 <sup>-3</sup> Q		
517.6 VAr	Phase angle: ±15 °	7,0 - 10 <sup>-3</sup> Q		
5000 VAr	50 A / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	0,2 - 10 <sup>-3</sup> Q		
4330 VAr	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> Q		
2500 VAr	Phase angle: ±30 °	5,0 - 10 <sup>-3</sup> Q		
1294 VAr	Phase angle: ±15 °	1,0 - 10 <sup>-2</sup> Q		
10000 VAr	100 A / 100 V // 10 Hz to 10 kHz Phase angle: 90 °	0,2 - 10 <sup>-3</sup> Q		
8660 VAr	Phase angle: ±60 °	2,0 - 10 <sup>-3</sup> Q		
5000 VAr	Phase angle: ±30 °	5,0 - 10 <sup>-3</sup> Q		
2588 VAr	Phase angle: ±15 °	1,0 - 10 <sup>-3</sup> Q		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC reactive power Sources and measuring devices (discrete points)	25 kVAr	50 A / 500 V // 40 Hz to 850 Hz Phase angle: 90 °	85 - 10 <sup>-6</sup> Q	Q= AC reactive power  Discrete points for current, voltage and phase angle
	21.7 kVAr	Phase angle: ±60 °	0,2 - 10 <sup>-3</sup> Q	
	12.5 kVAr	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> Q	
	6.5 kVAr	Phase angle: ±15 °	1,0 - 10 <sup>-3</sup> Q	
	40 kVAr	80 A / 500 V // 40 Hz to 850 Hz Phase angle: 90 °	85 - 10 <sup>-6</sup> Q	
	34.6 kVAr	Phase angle: ±60 °	0,2 - 10 <sup>-3</sup> Q	
	20 kVAr	Phase angle: ±30 °	0,5 - 10 <sup>-3</sup> Q	
	10.4 kVAr	Phase angle: ±15 °	1,0 - 10 <sup>-3</sup> Q	
AC reactive power Sources and measuring devices (areas)	50 µVAr to 500 mVAr	50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: ±90 °	0,17 - 10 <sup>-3</sup> Q	Q=AC reactive power
		50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: 90 ° to 60 Phase angle: -90 ° to -60 °	0,2 - 10 <sup>-3</sup> Q	
		50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: 60 ° to 30 ° Phase angle: -60 ° to -30 °	0,3 - 10 <sup>-3</sup> Q	
		50 mV ≤ U ≤ 5 V 1 mA ≤ I ≤ 100 mA 10 Hz to 1 kHz Phase angle: 30 ° to 15 ° Phase angle: -30 ° to -15 °	0,6 - 10 <sup>-3</sup> Q	
	> 500 mVAr to 500 VAr	5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: 90 °	0,1 - 10 <sup>-3</sup> Q	
		5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: 90 ° to 60 ° Phase angle: -90 ° to -60 °	0,1 - 10 <sup>-3</sup> Q	
		5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: 60 ° to 30 ° Phase angle: -60 ° to -30 °	0,3 - 10 <sup>-3</sup> Q	
		5 V ≤ U ≤ 500 V 100 mA ≤ I ≤ 1 A 16 Hz to 1 kHz Phase angle: 30 ° to 15 ° Phase angle: -30 ° to -15 °	0,6 - 10 <sup>-3</sup> Q	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC reactive power Sources and measuring devices (ranges)	> 500 VAR to 10 kVAR	25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: 90 °	0,1 - 10 <sup>-3</sup> Q	Q=AC reactive power
		25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: 90 ° to 60 ° Phase angle: -90 ° to -60 °	0,15 - 10 <sup>-3</sup> Q	
		25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: 60 ° to 30 ° Phase angle: -60 ° to -30 °	0,35 - 10 <sup>-3</sup> Q	
		25 V ≤ U ≤ 500 V 1 A ≤ I ≤ 20 A 16 Hz to 1 kHz Phase angle: 30 ° to 15 ° Phase angle: -30 ° to -15 °	0,75 - 10 <sup>-3</sup> Q	
	50 μVAR to 500 μVAR	0.05 V to 0.5 V 1 mA > 1 kHz to 10 kHz Phase angle: ±90° to ±60° Phase angle: <±60° to ±30° Phase angle: <±30° to ±15°	0,3 - 10 <sup>-3</sup> Q	
			0,8 - 10 <sup>-3</sup> Q 7,0 - 10 <sup>-3</sup> Q	
	> 0.5 mVAR to 200 mVAR	0.5 V to 1 V 1 mA to < 200 mA > 1 kHz to 10 kHz Phase angle: ±90° to ±60° Phase angle: <±60° to ±30° Phase angle: <±30° to ±15°	0,2 - 10 <sup>-3</sup> Q 0,4 - 10 <sup>-3</sup> Q 0,8 - 10 <sup>-3</sup> Q	
	> 200 mVAR to 20 VAR	> 0.1 V to 1 V > 200 mA to 20 A > 1 kHz to 10 kHz Phase angle: ±90° to ±60° Phase angle: <±60° to ±30° Phase angle: <±30° to ±15°	1,0 - 10 <sup>-3</sup> Q 4,0 - 10 <sup>-3</sup> Q 8,0 - 10 <sup>-3</sup> Q	
> 20 VAR to 100 VAR	> 0.1 V to 1 V > 20 A to 100 A > 1 kHz to 10 kHz Phase angle: ±90° to ±60° Phase angle: <±60° to ±30° Phase angle: <±30° to ±15°	3,0 - 10 <sup>-3</sup> Q 6,0 - 10 <sup>-3</sup> Q 10,0 - 10 <sup>-3</sup> Q		
10 mVAR to 200 mVAR	> 1 V to 10 V 1 mA to < 200 mA > 1 kHz to 10 kHz Phase angle: ±90° to ±60° Phase angle: < ±60 ° to ±30° Phase angle: < ±30° to ±15°	0,3 - 10 <sup>-3</sup> Q		
		0,7 - 10 <sup>-3</sup> Q		
		2,0 - 10 <sup>-3</sup> Q		



Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC reactive power Sources and measuring devices (ranges)	> 200 mVAr to 20 VAr	> 1 V to 10 V > 200 mA to 20 A > 1 kHz to 10 kHz Phase angle: $\pm 90^\circ$ to $\pm 60^\circ$ Phase angle: $< \pm 60^\circ$ to $\pm 30^\circ$ Phase angle: $< \pm 30^\circ$ to $\pm 15^\circ$	0,3 - $10^{-3} Q$ 0,7 - $10^{-3} Q$ 2,0 - $10^{-3} Q$	Q=AC reactive power
	> 20 VAr to 1000 VAr	> 1 V to 10 V > 20 A to 100 A > 1 kHz to 10 kHz Phase angle: $\pm 90^\circ$ to $\pm 60^\circ$ Phase angle: $< \pm 60^\circ$ to $\pm 30^\circ$ Phase angle: $< \pm 30^\circ$ to $\pm 15^\circ$	2,0 - $10^{-3} Q$ 5,0 - $10^{-3} Q$ 10 - $10^{-3} Q$	
	100 mVAr to 20 VAr	> 10 V to 100 V 1 mA to 200 mA > 1 kHz to 10 kHz Phase angle: $\pm 90^\circ$ to $\pm 60^\circ$ Phase angle: $< \pm 60^\circ$ to $\pm 30^\circ$ Phase angle: $< \pm 30^\circ$ to $\pm 15^\circ$	0,5 - $10^{-3} Q$ 1,0 - $10^{-3} Q$ 2,0 - $10^{-3} Q$	
	> 20 VAr to 1000 VAr	> 10 V to 100 V > 200 mA to 10 A > 1 kHz to 10 kHz Phase angle: $\pm 90^\circ$ to $\pm 60^\circ$ Phase angle: $< \pm 60^\circ$ to $\pm 30^\circ$ Phase angle: $< \pm 30^\circ$ to $\pm 15^\circ$	1,5 - $10^{-3} Q$ 3,0 - $10^{-3} Q$ 7,0 - $10^{-3} Q$	
	> 1 kVAr to 10 kVAr	> 10 V to 100 V > 10 A to 100 A > 1 kHz to 10 kHz Phase angle: $\pm 90^\circ$ to $\pm 60^\circ$ Phase angle: $< \pm 60^\circ$ to $\pm 30^\circ$ Phase angle: $< \pm 30^\circ$ to $\pm 15^\circ$	2,0 - $10^{-3} Q$ 5,0 - $10^{-3} Q$ 10,0 - $10^{-3} Q$	
	> 10 kVAr to 80 kVAr	500 V $\leq$ U $\leq$ 1000 V 20 A $\leq$ I $\leq$ 80 A 40 Hz to 850 Hz Phase angle: $\pm 90^\circ$	85 - $10^{-6} Q$	
		500 V $\leq$ U $\leq$ 1000 V 20 A $\leq$ I $\leq$ 80 A 40 Hz to 850 Hz Phase angle: $90^\circ$ to $60^\circ$ Phase angle: $-90^\circ$ to $-60^\circ$	0,25 - $10^{-3} Q$	
		500 V $\leq$ U $\leq$ 1000 V 20 A $\leq$ I $\leq$ 80 A 40 Hz to 850 Hz Phase angle: $60^\circ$ to $30^\circ$ Phase angle: $-60^\circ$ to $-30^\circ$	0,65 - $10^{-3} Q$	
		500 V $\leq$ U $\leq$ 1000 V 20 A $\leq$ I $\leq$ 80 A 40 Hz to 850 Hz Phase angle: $30^\circ$ to $15^\circ$ Phase angle: $-30^\circ$ to $-15^\circ$	1,5 - $10^{-3} Q$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC apparent power Sources and measuring devices (ranges)	50 $\mu$ VA to 500 $\mu$ VA	$50 \text{ mV} \leq U \leq 500 \text{ mV}$ $I = 1 \text{ mA}$ 10 Hz to 10 kHz	0,17 - $10^{-3} S$	S = AC apparent power
	500 $\mu$ VA up to 5 mVA	$U = 500 \text{ mV}$ $10 \text{ mA} \leq I \leq 100 \text{ mA}$ 10 Hz to 10 kHz	0,16 - $10^{-3} S$	
	5 mVA to 50 mVA	$U = 500 \text{ mV}$ $10 \text{ mA} \leq I \leq 100 \text{ mA}$ 10 Hz to 10 kHz	55 - $10^{-6} S$	
	50 mVA to 500 mVA	$500 \text{ mV} \leq U \leq 5 \text{ V}$ $I = 100 \text{ mA}$ 10 Hz to 10 kHz	30 - $10^{-6} S$	
	500 mVA up to 5 VA	$5 \text{ V} \leq U \leq 50 \text{ V}$ $I = 100 \text{ mA}$ 16 Hz to 10 kHz	25 - $10^{-6} S$	
	5 VA to 500 VA	$50 \text{ V} \leq U \leq 500 \text{ V}$ $100 \text{ mA} \leq I \leq 1 \text{ A}$ 16 Hz to 10 kHz	50 - $10^{-6} S$	
	500 VA to 5 kVA	$U = 500 \text{ V}$ $1 \text{ A} \leq I \leq 10 \text{ A}$ 16 Hz to 5 kHz	60 - $10^{-6} S$	
	5 kVA to 10 kVA	$U = 500 \text{ V}$ $1 \text{ A} \leq I \leq 20 \text{ A}$ 16 Hz to 5 kHz	60 - $10^{-6} S$	
	10 kVA to 80 kVA	$500 \text{ V} \leq U \leq 1000 \text{ V}$ $20 \text{ A} \leq I \leq 80 \text{ A}$ 40 Hz to 850 Hz	85 - $10^{-6} S$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Harmonic voltage Measuring devices	1 V to 1000 V	40 Hz to 850 Hz	$0.5 \cdot 10^{-3}$	Maximum up to the 100th harmonic
	1 V to 180 V	> 850 Hz to 5 kHz	$0.8 \cdot 10^{-3}$	
Sources	1 V to 1000 V	40 Hz to 850 Hz	$0.4 \cdot 10^{-3}$	
	1 V to 180 V	> 850 Hz to 5 kHz	$0.4 \cdot 10^{-3}$	
Power Measuring devices	0.01 A to 80 A	40 Hz to 850 Hz	$4.0 \cdot 10^{-3}$	
	0.01 A to 20 A	> 850 Hz to 5 kHz	3,0 %	
Sources	0.01 A to 80 A	40 Hz to 850 Hz	$4.0 \cdot 10^{-3}$	
	0.01 A to 20 A	> 850 Hz to 5 kHz	3,0 %	
Frequency	10 MHz	Measuring time > 30 min	$1 \cdot 10^{-11} f$	$f$ = current measured value
Frequency measurement	1 mHz to 46 GHz	Measuring time > 5 min	$\sqrt{(1 \cdot 10^{-10} \cdot f)^2 + U^2}$	At low frequencies are possible trigger uncertainties $U_{Tr}$ or $U_{Tr}$ must be taken into account.
	1 mHz to 50 GHz		$1 \cdot 10^{-10} f$	
Frequency synthesis				
Time interval	1 ns to 1000 s		$\sqrt{(1 \cdot 10^{-10} \cdot t)^2 + U^2 + I_{NS}^2 / T_r^2}$	
Speed optical	1 min <sup>-1</sup> up to 2 · 10 <sup>5</sup> min <sup>-1</sup>	with light pulse generator	$6 \cdot 10^{-6}$ but not smaller than 0.001 min <sup>-1</sup>	
	1 min <sup>-1</sup> up to 10000 min <sup>-1</sup>		$4 \cdot 10^{-4}$ but not smaller than 0.01 min <sup>-1</sup>	
Oscilloscope calibrators		Square wave voltage 10 Hz to 10 kHz $R_i = 50 \Omega$	$20 \cdot 10^{-6} 1 \mu V$	
Deflection vertical	1 mV to 5 V 1 mV to 200 V	$R_i = 1 M\Omega$		
Deflection horizontal	1 ns to 1 s	Time stamps	$\sqrt{(1 \cdot 10^{-10} \cdot t)^2 + U^2}$	$t$ = current measured value Trigger uncertainty $U_{Tr}$ to consider
	> 1 s to 5 s	Measuring time > 5 min $R_i = 1 M\Omega, 50 \Omega$	$\sqrt{(5 \cdot 10^{-10} \cdot t)^2 + U^2}$	
Rise time $t_r$	18 ps to 100 ps > 100 ps to 10 ms	20 mV to 1 V	8 ps $4.5 \cdot 10^{-2} \cdot t_r$ 3 ps	External trigger signal required

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - Electrical measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Oscilloscope measurement		Square wave voltage 10 Hz to 10 kHz		
Deflection vertical	1 mV to 5V 1 mV to 120 V	$R_i = 50 \Omega$ $R_i = 1 M\Omega$	0,35 % 0,35 %	$R_i$ Internal resistance
Deflection horizontal	50 ps to < 1 $\mu$ s 1 $\mu$ s to 5 s	Time stamps or sine  < 1 V	6 ps  $1.5 \cdot 10^{-3} \cdot t$	$t$ = current measured value
Rise time $t_r$	180 ps to 450 ps > 450 ps to 10 ms	250 mV  250 mV to 1 V	40 ps  $4.5 \cdot 10^{-2} \cdot t_r$	$t_r$ = current rise time
Bandwidth $B$	$f_c$  50 MHz to 26.5 GHz	0.2 V to 2 V  $R_i = 50 \Omega$  $  \Gamma_{oszi}   \leq 0,05$ $  \Gamma_{oszi}   \leq 0,1$ $  \Gamma_{oszi}   \leq 0,15$ $  \Gamma_{oszi}   \leq 0,2$	12 MHz 13 MHz 14 MHz 15 MHz	$f_c$ = frequency at which -3dB point $f_{Ref} = 5\% f_c$ $  \Gamma_{oszi}  $ : Reflection factor osci
Total Harmonic Distortion	0 to 0.3	100 Hz to 50 kHz	0,0001 0,0165 $\cdot$ THD	
THD / distortion	0 to 0.3	100 kHz to 2 GHz	0,0001 0,0675 $\cdot$ THD	
factor $\Gamma_{THDAudio}$	0 to 0.3	100 Hz to 50 kHz	0.001 0.007 $\cdot$ $\Gamma_{THDAudio}$	
Flicker*) $\Delta U / U$ Frequency	0.4 to 5 0.0083 Hz to 40 Hz	DIN EN 61000-4-15:2011	$7 \cdot 10^{-3} \cdot \Delta U / U$ $3 \cdot 10^{-3} \cdot \Delta U / U$	
$p_{St}$ (Short Term) $p_{Lt}$ (Long Term)	10 minutes 2 hours	(115 V, 60 Hz);(230 V 50 Hz) (115 V, 60 Hz);(230 V 50 Hz)	0,5 % 1,7%	
Current transformer transformation ratio	40 A to 400 A (primary) 10 mA to 5 A (secondary) > 400 A to 4000 A (primary) 100 mA to 5 A (secondary)	DC	20 - $10^{-6}$ 24 - $10^{-6}$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - High-frequency measurands

Calibration and measurement capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
HF power  Power meters	> 1 pW to 0.1 mW	2.5 MHz to 2 GHz > 2 GHz to 18 GHz > 18 GHz to 26.5 GHz	(0.025 0.14 -  Γ ) - P (0.049 0.21 -  Γ ) - P (0.071 0.32 -  Γ ) - P	Connector system: N, PC-3.5 ; 50 Ω;  Γ  KG ≤ 0.2  Connector system: PC-3.5 ; 50 Ω;  Γ  KG ≤ 0.2
HF power  Signal generators	> 1 pW to 0.1 mW	2.5 MHz to 2 GHz  > 2 GHz to 18 GHz > 18 GHz to 26.5 GHz	(0.035 0.13 -  Γ ) - P  (0.053 0.2 -  Γ ) - P (0.074 0.31 -  Γ ) - P	Connector system: N, PC-3.5 ; 50 Ω;  Γ  KG ≤ 0.2  Connector: PC-3.5; 50 Ω,  Γ  KG ≤ 0.2
	0.1 mW to 10 mW	9 kHz to <0.1 MHz 0.1 MHz to 50 MHz >50 MHz to 6 GHz >6 GHz to 18 GHz  0.1 MHz to 50 MHz > 50 MHz to 6 GHz > 6 GHz to 18 GHz	17 · 10 <sup>-3</sup> - P 10 · 10 <sup>-3</sup> - P 15 · 10 <sup>-3</sup> - P 20 · 10 <sup>-3</sup> - P  20 · 10 <sup>-3</sup> - P 30 · 10 <sup>-3</sup> - P 40 · 10 <sup>-3</sup> - P	N connector; 50 Ω  Γ  ≤ 0,3      Γ  ≤ 0,5
	10 mW to 50 W	0.1 MHz to 2 GHz 0.1 MHz to 2 GHz 0.1 MHz to 2 GHz	48 · 10 <sup>-3</sup> - P 63 · 10 <sup>-3</sup> - P 123 · 10 <sup>-3</sup> - P	Γ  of the KG ≤ 0.1  Γ  of the KG ≤ 0.3  Γ  of the KG ≤ 0.5 N connector; PC-3.5 3)
	0.1 mW to 10 mW	10 MHz to 1 GHz > 1 GHz to 10 GHz > 10 GHz to 18 GHz > 18 GHz to 26.5 GHz 10 MHz to 1 GHz > 1 GHz to 10 GHz > 10 GHz to 18 GHz > 18 GHz to 26.5 GHz	20 · 10 <sup>-3</sup> - P 30 · 10 <sup>-3</sup> - P 40 · 10 <sup>-3</sup> - P 45 · 10 <sup>-3</sup> - P 40 · 10 <sup>-3</sup> - P 80 · 10 <sup>-3</sup> - P 100 · 10 <sup>-3</sup> - P 110 · 10 <sup>-3</sup> - P	Connector PC-3.5; 50 Ω <sup>3)</sup>  Γ  ≤ 0,3     Γ  ≤ 0,5

3) When using other connector systems, the measurement uncertainty increases.

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - High-frequency measurands

Calibration and measurement capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
HF power Power meters	0.1 mW to 10 mW	9 kHz to < 0.1 MHz 0.1 MHz to 50 MHz > 50 MHz to 6 GHz > 6 GHz to 18 GHz	$17 \cdot 10^{-3} - P$ $6.0 \cdot 10^{-3} - P$ $12 \cdot 10^{-3} - P$ $20 \cdot 10^{-3} - P$	N connector; 50 Ω <sub>3</sub> $ \Gamma  \leq 0,3$
	> 10 mW to 50 W	32 MHz to 1 GHz	$20 \cdot 10^{-3} - P$	
	0.1 mW to 10 mW	10 MHz to 1 GHz > 1 GHz to 10 GHz > 10 GHz to 18 GHz > 18 GHz to 26.5 GHz	$10 \cdot 10^{-3} - P$ $15 \cdot 10^{-3} - P$ $20 \cdot 10^{-3} - P$ $25 \cdot 10^{-3} - P$	Connector PC-3.5; 50 Ω <sub>3</sub> $ \Gamma  \leq 0,3$
Phase noise Signal generators	Phase noise related to carrier amplitude in dBc/Hz	Offset frequency referred to carrier frequency		Carrier frequency: 100 MHz - 1 GHz
	> -87 dBc/Hz	100 Hz	2.5 dB	
	> -99 dBc/Hz	1 kHz	2.5 dB	
	> -104 dBc/Hz	10 kHz	2.5 dB	
	> -111 dBc/Hz	100 kHz	2.5 dB	
	> -131 dBc/Hz	1 MHz	2.5 dB	
	> -137 dBc/Hz	10 MHz	2.5 dB	
	> -80 dBc/Hz	100 Hz	2.5 dB	> 1 MHz - 3 GHz
	> -96 dBc/Hz	1 kHz	2.5 dB	
	> -101 dBc/Hz	10 kHz	2.5 dB	
	> -109 dBc/Hz	100 kHz	2.5 dB	
	> -126 dBc/Hz	1 MHz	2.5 dB	
	> -136 dBc/Hz	10 MHz	2.5 dB	
	> -72 dBc/Hz	100 Hz	2.5 dB	> 3 GHz - 6 GHz
	> -93 dBc/Hz	1 kHz	2.5 dB	
	> -98 dBc/Hz	10 kHz	2.5 dB	
	> -106 dBc/Hz	100 kHz	2.5 dB	
	> -120 dBc/Hz	1 MHz	2.5 dB	
	> -135 dBc/Hz	10 MHz	2.5 dB	
HF noise display Receiver / Measuring devices	10 Hz to 50 GHz	-165 to 0 dBm/Hz dbm/Hz	1 dB	
Signal level difference Measuring devices / sources	0 dBc to 90 dBc	9 kHz to 7 GHz > 7 GHz to 13.6 GHz > 13.6 GHz to 26.5 GHz	1.5 dB 2.3 dB 3 dB	SNR > 20 dB
	> 90 dBc to 100 dBc	9 kHz to 7 GHz > 7 GHz to 13.6 GHz > 13.6 GHz to 26.5 GHz	4.5 dB 4.8 dB 5.3 dB	SNR > 20 dB

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent laboratory Kirchzarten, Gewerbestraße 3 - High-frequency measurands**

**Calibration and measurement capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Filter bandwidth Measuring devices	1 Hz to 40 MHz		1 %	SNR > 70 dB
Form factor Measuring devices	1:1 bis 4:1 > 4:1 to 10:1 > 10:1 bis 18:1		5,5 % 7 % 8,5 %	SNR > 20 dB
Amplitude modulation: Modulation depth $m$	0.0 to $\leq 1.0$	$f_{MOD} < 1 \text{ MHz}$	0,004 0,025 · $m$	$f_{HF}$ = carrier frequency $f_{HF} < 4 \text{ GHz}$ $f_{MOD}$ = modulation freq. Absolute measuremen t uncertainty
Frequency modulation Frequency deviation $\Delta f$	0 Hz to 5 MHz	$f_{MOD} < 1 \text{ MHz}$	0.041 · $\Delta f$ 25 Hz	$f_{HF}$ = Carrier frequency $f_{HF} < 4 \text{ GHz}$ $f_{MOD}$ = Modulation frequency $\Delta f$ = frequency deviation Absolute measuremen t uncertainty
Phase modulation Phase deviation $\Delta \Phi$	0 to (4 MHz / $f_{MOD}$ ) rad	$f_{MOD} < 1 \text{ MHz}$	0.025 rad 0.041 · $\Delta \Phi$	$f_{HF}$ = Carrier frequency $f_{HF} < 4 \text{ GHz}$ $f_{MOD}$ = Modulation frequency $\Delta \Phi$ = phase deviation Absolute measuremen t uncertainty
Distortion factor $k$	> 0.0001 to 0.01 > 0.01 to 0.1 > 0.1 to 0.2	AM- Demodulation method $f_{HF}$ : 150 kHz to 2 GHz $f_{MOD} = 1 \text{ kHz}$ $P_{HF} = 0 \text{ dBm}$	0,030 0,029 0,025	$f_{HF}$ = Carrier frequency $f_{MOD}$ = Modulation frequency $P_{HF}$ = Carrier level
	> 0.0001 to 0.01 > 0.01 to 0.1 > 0.1 to 0.2	FM & PM- Demodulation method $f_{HF}$ : 150 kHz to 2 GHz $f_{MOD} = 1 \text{ kHz}$ $P_{HF} = 0 \text{ dBm}$ $\Delta f \leq 50 \text{ kHz}$	0,09	Absolute measurement uncertainty

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - High-frequency measurands

Calibration and measurement capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Pulsed measurands *) Spectral voltage amplitude density (Measure/ Display)	$S_{\sigma} = 13.5 \mu\text{Vs}$	DIN EN 55016-1-1:2020 CISPR 16-1-1:2019  CISPR Band A 9 kHz to 0.15 MHz	0.30 dB	Pulse frequency 1 Hz to 100 Hz  $r_{\sigma}, r_{\tau} \leq 0.05$ (represent)
	$S_{\sigma} = 0.316 \mu\text{Vs}$	CISPR Band B > 0.15 MHz to 30 MHz	0.30 dB	Pulse frequency 1 Hz to 1000 Hz  $r_{\sigma}, r_{\tau} \leq 0.07$ (represent)
	$S_{\sigma} = 0.0044 \mu\text{Vs}$	CISPR Band C > 30 MHz to 300 MHz	0.36 dB	Pulse frequency 1 Hz to 1000 Hz  $r_{\sigma}, r_{\tau} \leq 0.12$ (represent)
	$S_{\sigma} = 0.0044 \mu\text{Vs}$	CISPR Band D > 300 MHz to 1 GHz	0.40 dB	
HF current transformer clamp *) Transfer light resistance dB(Q)	9 kHz to 100 MHz	DIN EN 55016-1-2:2019	0.3 dB	
	> 100 MHz to 400 MHz	4.4 mA	0.5 dB	
	> 400 MHz to 1 GHz		0.8 dB	
HF Bulk Current Injection Transducer Clamp*) Insertion loss dB	9 kHz to 100 MHz	DIN EN 61000-4-6:2014	0.3 dB	
	> 100 MHz to 400 MHz	4.4 mA	0.5 dB	
	> 400 MHz to 1 GHz		1.5 dB	
Burst generators *) Voltage pulse	100 V to 4400 V	DIN EN 61000-4-4:2013 under load (RL)	2,2 %	$R_L = \text{load resistance}$
		at $R_L = 50 \Omega$		
		at $R_L = 1 \text{ k}\Omega$		
Rise time and Pulse width	3 ns to 1 $\mu\text{s}$		2,5 %	
Burst duration and Burst period	100 ns to 1 s		0,25 %	
Surge generators *) Measure and display voltage amplitude	250 V to 7000 V	DIN EN 61000-4-5:2019 with or without  Coupling and  Decoupling network	3,5 %	
			3,5 %	
			3,5 %	
Current amplitude	5 A to 5 kA			
Rise time and Pulse width	400 ns to 1 ms			



Annex to the partial accreditation certificate D-K-15070-01-01

Permanent laboratory Kirchzarten, Gewerbestraße 3 - High-frequency measurands

Calibration and measurement capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
RF attenuation	0 dB to 60 dB	9 kHz to 26.5 GHz > 26.5 GHz to 40 GHz > 40 GHz to 50 GHz	0.06 dB 0.09 dB 0.11 dB	Connector 50 Ω: N50; PC-3.5; PC- 2.4 <sub>3</sub> )
	> 60 dB to 90 dB	9 kHz to 26.5 GHz	0.12 dB	
HF reflection coefficient	0 to 1	9 kHz to 10 GHz > 10 GHz to 18 GHz > 18 GHz to 26.5 GHz > 26.5 GHz to 50 GHz	0,003 0,0035/ Γ  0,004 0,0040/ Γ  0,004 0,0045/ Γ  0,007 0,0060/ Γ	Connector 50 Ω: N50; PC-3.5; PC-2.4 <sub>3</sub> )  Γ : Amount of the complex Reflection factor
RF attenuation Phase angle φ	-180° to 180°	9 kHz to 50 GHz	$u_s - 180^\circ/\pi K - f$	$u_s = \arcsin(U( \Gamma )/ \Gamma )$ K = 0.025°/GHz Connector 50 Ω: N50; PC- 3.5; PC-2.4 <sub>3</sub> )
RF transmission phase Phase angle φ	-180° to 180°	9 kHz to 50 GHz	$U_T - 180^\circ/\pi K - f 0.3^\circ$	$U_T = \arcsin(\frac{10^{U/20} - 1}{K})$ K: 0.05°/GHz U: Uncertainty of the Attenuation in dB
Electrostatic Discharge (ESD) Current pulse $I_p$ Support values current pulse $I_{30}$ Current pulse $I_{60}$ Rise time $t_r$ DC voltage $U_L$	1 A to 120 A  1 A until 120 A 1 A until 120 A 0.6 ns until 1 μs 1 kV to 30 kV	DIN EN61000-4-2	3,0 %  3,5 % 3,5 % 5,0 % 0,5 %	$I_p$ = first discharge current peak $I_{30}$ = current at 30 ns $I_{60}$ = current at 60 ns

3) When using other connector systems, the measurement uncertainty increases.

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - Dimensional measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Length of Cylindrical setting standards, ring gauges: Diameter	1 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006 Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ = is the measured Diameter
Plug gauges: Diameter	1 mm to 200 mm		$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Test pins: Diameter	0,1 mm to 30 mm	VDI/VDE/DGQ 2618 Sheet 4.2:2007 Point 3.2.2 (Opt. 1)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Thread gauges (one and multi-start cylindrical male and female threads with straight flanks, symmetrical profile) Threaded mandrels: simple Flank diameter	1.4 mm to 200 mm nominal pitch: 0.3 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.8:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Three-wire method $d$ = is the measured diameter
Threaded rings: simpler Flank diameter	3 mm to 200 mm nominal pitch: 0.5 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.9:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Two-ball method $d$ = is the measured Diameter
Length of plane-parallel, spherical or cylindrical measuring surfaces	0,01 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 19.1:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Diameter	0,01 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006 Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ is the measured diameter
Feeler gauges	0,03 mm to 2,00 mm	DIN 2275:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Adjustment dimensions for Outside micrometers	25 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 4.4:2009	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Dimensional measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Throat gauges	3 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.7:2005 Point 3.3.2 (Opt. 2)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ is the measured Diameter
Caliper for Exterior, interior and Depth measurements	0 mm until 500 mm	VDI/VDE/DGQ 2618 Sheet 9.1:2006	$30 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	$l$ is the measured Length
Depth caliper,	> 500 mm until 1000 mm	VDI/VDE/DGQ 2618 Sheet 9.2:2006	$50 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	
Height caliper		VDI/VDE/DGQ 2618 Sheet 9.3:2006		
Outside micrometers	0 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.1:2001	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Precision micrometers	0 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 10.3:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Micrometer head screws	0 mm to 50 mm	VDI/VDE/DGQ 2618 Sheet 10.4:2008	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Depth gauges	0 mm to 300 mm	VDI/VDE/DGQ 2618 Sheet 10.5:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 2-point contact on the calibration object	13 mm until 300 mm > 300 mm until 500 mm	VDI/VDE/DGQ 2618 Sheet 10.7:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$ $5 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 3-line contact on the calibration object	3 mm to 150 mm	VDI/VDE/DGQ 2618 Sheet 10.8:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	$d$ is the measured diameter
Lever gauges (quick probe) for outdoor measurements	to 200 mm	VDI/VDE/DGQ 2618 Sheet 12.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Lever gauges (quick probes) for Internal measurements	2 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 13.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Dial gauges	0 mm to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.1:2021	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	mechanical dial gauges
		VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	electronic digital dial gauges
Fine pointer	0 mm to 3 mm	VDI/VDE/DGQ 2618 Sheet 11.2:2002	$0.6 \mu\text{m}$	
Feeler lever gauges	0 mm to 1,6 mm	VDI/VDE/DGQ 2618 Sheet 11.3:2002	$1.0 \mu\text{m}$	
Electr. inductive Linear Encoders	to 100 mm	VDI/VDE/DGQ 2618 Sheet 14.1:2010	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	
Electr. incremental Linear Encoders	to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - Dimensional measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Coordinate measuring technology Coordinate measuring machines with optical probing  Measuring projectors, Measuring microscopes	Devices with one measuring level with a Area diagonals  ≤ 450 mm	Calibration of the metrological  Properties according to DKD-R 4-3 Sheet 18.1:2018, and the below mentioned Standards and guidelines DIN EN ISO 10360 VDI/VDE 2617		Measuring systems with visual probing  or electronic Edge detection
		Determination of the Probing deviation $P_{Sx}$ , $P_{Sy}$ and $P_{S2D}$ by means of a Circle standard according to VDI/VDE 2617 Sheet 6.1:2021	0,5 $\mu$ m	
		Determination of the Probing deviation of the Image processing system $P_{SVx}$ , $P_{SVy}$ and $P_{SV2D}$ by means of a Circle standard according to VDI/VDE 2617 Sheet 6.1:2021	0,5 $\mu$ m	
		Determination of the Length measurement deviation $E_{UXy}$ , $E_{UX}$ and $E_{UY}$ by means of a Line scale or Circle matrix according to DIN EN ISO 10360-7:2011	0.5 $\mu$ m $0.7 \cdot 10^{-6} \cdot l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Dimensional measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Coordinate measuring technology  Coordinate measuring machines with optical probing  Measuring projectors, measuring microscopes		Determination of the length measurement deviation of the image processing system $E_{uv}$ by means of a line scale or circle matrix according to DIN EN ISO 10360-7:2011	0,5 $\mu\text{m}$	
	to 100 mm  to 100 $\mu\text{m}$	Determination of the length measurement deviation $E_{uz}$ by means of gauge blocks or depth setting standard according to DIN EN ISO 10360-7:2011	0.5 $\mu\text{m}$ $0.7 \cdot 10^{-6} \cdot l$  0.25 $\mu\text{m}$	$l$ is the measured length

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC voltage Measuring devices	0 V 1 mV to 2,2 V > 2,2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		0.1 $\mu$ V $7 \cdot 10^{-6} U$ 1 $\mu$ V $9 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U = \text{set value}$
DC voltage Sources	0 V 1 mV to 100 mV > 100 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		0.1 $\mu$ V $8 \cdot 10^{-6} U$ 1 $\mu$ V $11 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$	$U = \text{measured value}$
High Voltage	> 1 kV to 10 kV		$2.5 \cdot 10^{-3} U$ 2.5 V	$U = \text{measured value}$
DC current strength meters and sources	0 A 0.1 $\mu$ A until < 1 $\mu$ A 1 $\mu$ A until < 10 $\mu$ A 10 $\mu$ A until < 100 $\mu$ A 100 $\mu$ A until 320 mA > 320 mA until 1 A > 1 A to 10 A > 10 A > 150 A 150 A to 2000 A	Precision Open Normal resistance and voltmeter Current transformer	0.2 nA $35 \cdot 10^{-6} / 21$ pA $50 \cdot 10^{-6} / 6$ pA $15 \cdot 10^{-6} / 0.4$ nA $18 \cdot 10^{-6} /$ $15 \cdot 10^{-6} / 6$ $\mu$ A $0,2 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$	$I = \text{set value}$
Direct current strength Current clamps	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 1000 A		$1 \cdot 10^{-3} /$ $2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks		
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement			
DC resistance	0 Ω		50 μΩ	R = set value Fluke 5700A		
	1 Ω; 1.9 Ω		95 · 10 <sup>-6</sup> R			
	10 Ω		28 · 10 <sup>-6</sup> R			
	19 Ω		27 · 10 <sup>-6</sup> R			
	100 Ω; 190 Ω		17 · 10 <sup>-6</sup> R			
	1 kΩ		13 · 10 <sup>-6</sup> R			
	1.9 kΩ		13 · 10 <sup>-6</sup> R			
	10 kΩ		12 · 10 <sup>-6</sup> R			
	19 kΩ		12 · 10 <sup>-6</sup> R			
	100 kΩ		14 · 10 <sup>-6</sup> R			
	190 kΩ		14 · 10 <sup>-6</sup> R			
	1 MΩ		20 · 10 <sup>-6</sup> R			
	1.9 MΩ		21 · 10 <sup>-6</sup> R			
	10 MΩ		40 · 10 <sup>-6</sup> R			
	19 MΩ		48 · 10 <sup>-6</sup> R			
	100 MΩ		110 · 10 <sup>-6</sup> R			
	0 Ω	1 Ω to 10 Ω			100 μΩ	R = measured value HP 3458A
		> 10 Ω to 100 Ω			16 · 10 <sup>-6</sup> R 50 μΩ	
		> 100 Ω to 1 kΩ			12 · 10 <sup>-6</sup> R 500 μΩ	
> 1 kΩ to 10 kΩ			15 · 10 <sup>-6</sup> R			
> 10 kΩ to 100 kΩ			15 · 10 <sup>-6</sup> R			
> 100 kΩ to 1 MΩ			35 · 10 <sup>-6</sup> R			
> 1 MΩ to 10 MΩ			150 · 10 <sup>-6</sup> R			
> 10 MΩ to 100 MΩ			600 · 10 <sup>-6</sup> R			
> 100 MΩ to 1 GΩ			5 · 10 <sup>-3</sup> R			
0,001 Ω to 0,1 Ω		Substitution procedure	50 · 10 <sup>-6</sup> R			
> 0,1 Ω to 1 MΩ	with normal resistance	20 · 10 <sup>-6</sup> R				
> 1 MΩ to 100 MΩ		30 · 10 <sup>-6</sup> R				

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC resistance (areas) Measuring devices	1 Ω to < 11 Ω		$0.12 \cdot 10^{-3} R$	R = set value Fluke 5520A /5522A
	11 Ω to < 33 Ω		$33 \cdot 10^{-6} R$	
	33 Ω to < 110 Ω		$29 \cdot 10^{-6} R$	
	110 Ω to < 330 Ω		$28 \cdot 10^{-6} R$	
	330 Ω to < 1.1 kΩ		$28 \cdot 10^{-6} R$	
	1.1 kΩ to < 3.3 kΩ		$28 \cdot 10^{-6} R$	
	3.3 kΩ to < 11 kΩ		$28 \cdot 10^{-6} R$	
	11 kΩ to < 33 kΩ		$28 \cdot 10^{-6} R$	
	33 kΩ to < 110 kΩ		$28 \cdot 10^{-6} R$	
	110 kΩ to < 330 kΩ		$32 \cdot 10^{-6} R$	
	330 kΩ to < 1.1 MΩ		$33 \cdot 10^{-6} R$	
	1.1 MΩ to < 3.3 MΩ		$62 \cdot 10^{-6} R$	
	3.3 MΩ to < 11 MΩ		$0.13 \cdot 10^{-3} R$	
	11 MΩ to < 33 MΩ		$0.25 \cdot 10^{-3} R$	
	33 MΩ to < 110 MΩ		$0.5 \cdot 10^{-3} R$	
	110 MΩ to < 330 MΩ		$3 \cdot 10^{-3} R$	
330 MΩ to < 1.1 GΩ		$15 \cdot 10^{-3} R$		
AC voltage Measuring devices and Sources	1 mV to 2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	U = measured value
		> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV to 7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
		> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$	
	> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$		
	> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$		



**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and Sources	> 7 mV to 22 mV	10 Hz to 20 Hz	$80 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$80 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$65 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$75 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$75 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$95 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$0.19 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.21 \cdot 10^{-3} U$	
	> 22 mV to 70 mV	10 Hz to 20 Hz	$70 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$58 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$35 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$35 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$45 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$55 \cdot 10^{-6} U$	
	> 300 kHz to 500 kHz	$0.11 \cdot 10^{-3} U$		
	> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and Sources	> 70 mV to 220 mV	10 Hz to 20 Hz	$39 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$35 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$25 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$25 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$28 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$42 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$85 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.1 \cdot 10^{-3} U$	
	> 220 mV to 700 mV	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$22 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$12 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$12 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$14 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$27 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$40 \cdot 10^{-6} U$	
	> 700 mV to 2.2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$11 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$11 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$22 \cdot 10^{-6} U$		
> 500 kHz to 1 MHz		$68 \cdot 10^{-6} U$		
> 2.2 V to 7 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$13 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$		
	> 500 kHz to 1 MHz	$95 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks		
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement			
AC voltage  Measuring devices and Sources	> 7 V to 22 V	10 Hz to 20 Hz	$17 \cdot 10^{-6} U$	<i>U</i> = measured value		
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$			
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$			
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$			
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$			
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$			
		> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$			
		> 500 kHz to 1MHz	$0.11 \cdot 10^{-3} U$			
	> 22 V to 70 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$			
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$			
		> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$			
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$			
		> 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$			
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$			
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$			
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$			
	> 70 V to 220 V	10 Hz to 20 Hz	$19 \cdot 10^{-6} U$			
		> 20 Hz to 40 Hz	$18 \cdot 10^{-6} U$			
		> 40 Hz to 20 kHz	$17 \cdot 10^{-6} U$			
		> 20 kHz to 50 kHz	$17 \cdot 10^{-6} U$			
		> 50 kHz to 100 kHz	$32 \cdot 10^{-6} U$			
		10 Hz to 20 Hz	$25 \cdot 10^{-6} U$			
		> 20 Hz to 40 Hz	$27 \cdot 10^{-6} U$			
		> 40 Hz to 20 kHz	$45 \cdot 10^{-6} U$			
	> 220 V to 1000 V	> 20 kHz to 50 kHz	$45 \cdot 10^{-6} U$			
		> 50 kHz to 100 kHz	$65 \cdot 10^{-6} U$			
		> 0.7 kV to 1 kV	50 Hz		$2.5 \cdot 10^{-3} U$ 0.25 V	
		> 1 kV to 7 kV			$3.5 \cdot 10^{-3} U$ 2.0 V	
Alternating current strength sources and measuring devices	100 µA to 1 mA	10 Hz to 40 Hz	$120 \cdot 10^{-6} I$	<i>I</i> = measured value		
		> 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$160 \cdot 10^{-6} I$ $60 \cdot 10^{-6} I$			
	> 1 mA to 10 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$46 \cdot 10^{-6} I$			

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Alternating current strength sources and measuring devices	> 10 mA to 1 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$17 \cdot 10^{-6} /$	I = measured value
	> 1 A to 10 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$32 \cdot 10^{-6} /$	
	> 10 A to 20 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$39 \cdot 10^{-6} /$	
	> 20 A to 100 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$69 \cdot 10^{-6} /$ $69 \cdot 10^{-6} /$ $0.17 \cdot 10^{-3} /$	
Sources	100 A to 2000 A	50 Hz	$3.0 \cdot 10^{-3} /$	Current transformer
Current clamps	1 mA to 2.2 A	40 Hz to 5 kHz	$2 \cdot 10^{-3} /$	I = measured value
	> 2.2 A to 20 A	40 Hz to 5 kHz	$3 \cdot 10^{-3} /$	
	> 20 A to 800 A	40 Hz to 65 Hz	$4 \cdot 10^{-3} /$	
Capacity gauges	190 pF to < 400 pF	10 Hz to 10 kHz	$4 \cdot 10^{-3} C$ 8 pF	With 5520A / 5522A
	400 pF to < 1.1 nF	10 Hz to 10 kHz	$4.5 \cdot 10^{-3} C$ 8 pF	
	1.1 nF to < 3.3 nF	10 Hz to kHz	$4.0 \cdot 10^{-3} C$ 8 pF	
	3.3 nF to < 11 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 8 pF	
	11 nF to < 33 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	33 nF to < 110 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	110 nF to < 330 nF	10 Hz to 1 kHz	$4.5 \cdot 10^{-3} C$	
	330 nF to < 1.1 μF	10 Hz to 600 Hz	$4.5 \cdot 10^{-3} C$	
	1.1 μF to < 3.3 μF	10 Hz to 300 Hz	$4.5 \cdot 10^{-3} C$	
	3.3 μF to < 11 μF	10 Hz to 150 Hz	$4.5 \cdot 10^{-3} C$	
	11 μF to < 33 μF	10 Hz to 120 Hz	$6.0 \cdot 10^{-3} C$	
	33 μF to < 110 μF	10 Hz to 80 Hz DC to 50	$6.5 \cdot 10^{-3} C$	
	110 μF to < 330 μF	Hz DC to Hz DC to 6	$6.0 \cdot 10^{-3} C$	
	330 μF to < 1.1 mF	Hz DC to 2 Hz DC to	$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	200.6 Hz	$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 0,2 Hz	$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF		$8.0 \cdot 10^{-3} C$	
33 mF to 110 mF		$11 \cdot 10^{-3} C$		
1 nF to 100 nF	50 Hz to 10 kHz	$1,0 \cdot 10^{-3} C$	C: measured value with normal capacitances	
> 100 nF to 1000 nF	50 Hz to 1 kHz	$1,0 \cdot 10^{-3} C$		
	> 1 kHz to 10 kHz	$2,5 \cdot 10^{-3} C$		

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Frequency	1 mHz to 46 GHz		$2 \cdot 10^{-9} \cdot f_{UTF}$	$f$ = current measured value $_{UTF}$ = Trigger uncertainty
Time interval	1 $\mu$ s to 1000 s		$2 \cdot 10^{-9} \cdot t$ 2 ns	$t$ = current measured value
Speed optical	1 min <sup>-1</sup> up to 100.000 min <sup>-1</sup>	with light pulse generator	$8 \cdot 10^{-6}$ but not less than 0.006 min <sup>-1</sup>	
AC active power measuring devices	109 $\mu$ W to < 11kW	33 mV to 1000 V 45 Hz to 65 kHz $PF = 1$ 33 mA to < 11 A	$1.4 \cdot 10^{-3} P$	$P$ = set value with Fluke 5520A/5522A $PF$ : Power factor
	363 mW up to 20 kW	11 A to 20 A	$2.0 \cdot 10^{-3} P$	
DC power Measuring devices	1 mW to 300 W		$0.5 \cdot 10^{-3} P$	
	> 300 W to 20 kW		$1.0 \cdot 10^{-3} P$	
Sources	1 mW to 300 W > 300 W to 1 kW > 1 kW to 1 MW	Product of U and I 1 mV $\leq$ U $\leq$ 1000 V 100 $\mu$ A $\leq$ I $\leq$ 2000 A	$30 - 10^{-6} P$ $200 - 10^{-6} P$ $300 - 10^{-6} P$	$P$ Calculated power
AC active power	5 mW up to 50 kW > 2.5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 - 10^{-6}$ not smaller than $0.14 - 10^{-3}$	$w(U_F)$ Uncertainty of the amplitude of the voltage fundamental $w(I_F)$ Uncertainty of amplitude of Current fundamental $w(\Phi_F)$ Uncertainty of the phase shift angle $w(U_{rmc})$ Uncertainty of the voltage rms value $w(I_{rmc})$ Uncertainty of the rms current value
	5 mvar to 50 kvar > 2.5 var up to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 - 10^{-6}$ not smaller than $0.14 - 10^{-3}$	
Apparent power	0.1 VA up to 50 kVA > 50 VA to 120 kVA	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 - 10^{-6}$ not smaller than $0.14 - 10^{-3}$	

**Annex to the partial accreditation certificate D-K-15070-01-01**
**On-site calibration - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks	
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement		
Stress ratio	± 2 mV/V	Bridge voltage: 5 V		Calibration of 350 Ω bridge standards and the associated indicators  at discrete points in 10% steps	
		Measuring frequency 225 Hz	0.04 μV/V		
		Measuring frequency 600 Hz	0.05 μV/V		
	± 2 mV/V	Measuring frequency 4.8 kHz	1.0 μV/V		
		Bridge voltage: 2,5 V			
		Measuring frequency 225 Hz	0.05 μV/V		
	± 2 mV/V	Measuring frequency 600 Hz	0.05 μV/V		
		Measuring frequency 4.8 kHz	1.0 μV/V		
		Bridge voltage: 5 V			
	± 5 mV/V	Measuring frequency 225 Hz	0.15 μV/V		
		Measuring frequency 4.8 kHz	1.0 μV/V		
	± 10 mV/V	Bridge voltage: 5 V			
		Measuring frequency 225 Hz	0.10 μV/V		
± 10 mV/V	Measuring frequency 4.8 kHz	0.30 μV/V			
	Bridge voltage: 2,5 V				
± 5 mV/V	Measuring frequency 225 Hz	0.1 μV/V			
	Measuring frequency 600 Hz	0.1 μV/V			
	Measuring frequency 4.8 kHz	1.0 μV/V			
± 10 mV/V	Bridge voltage: 2,5 V				
	Measuring frequency 225 Hz	0.4 μV/V	Calibration of 350 Ω bridge standards and the associated indicators		
	Measuring frequency 600 Hz	0.4 μV/V			
Measuring frequency 4.8 kHz	0.4 μV/V				
± 10 mV/V	Bridge voltage: 1 V		at discrete points in 10% steps		
	Measuring frequency 600 Hz	0.40 μV/V			
± 20 mV/V	Bridge voltage: 1 V				
	Measuring frequency 4.8 kHz	0.60 μV/V			
± 100 mV/V	Bridge voltage: 1 V				
	Measuring frequency 4.8 kHz	5.0 μV/V			
± 100 mV/V	Bridge voltage: 2,5 V				
	Measuring frequency 4.8 kHz	5.0 μV/V			

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Voltage ratio DC voltage bridge standards	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  0,5 V	2.0 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  1,0 V	1.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  2,5 V	0.5 $\mu$ V/V 0.5 $\mu$ V/V 0.5 $\mu$ V/V 0.5 $\mu$ V/V 0.5 $\mu$ V/V 1.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  5,0 V	0.3 $\mu$ V/V 0.25 $\mu$ V/V 0.25 $\mu$ V/V 0.25 $\mu$ V/V 0.35 $\mu$ V/V 1.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  7,5 V	0.2 $\mu$ V/V 0.2 $\mu$ V/V 0.2 $\mu$ V/V 0.2 $\mu$ V/V 0.3 $\mu$ V/V 1.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  10,0 V	0.1 $\mu$ V/V 0.15 $\mu$ V/V 0.15 $\mu$ V/V 0.2 $\mu$ V/V 0.3 $\mu$ V/V 1.5 $\mu$ V/V	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Voltage ratio DC voltage bridges, measuring instruments, measuring amplifier	-2 mV/V to 2 mV/V	Bridge voltage: 0,5 V	0.35 $\mu$ V/V	With K148
	-5 mV/V to 5 mV/V		0.35 $\mu$ V/V	
	-10 mV/V to 10 mV/V		0.40 $\mu$ V/V	
	-20 mV/V to 20 mV/V		0.55 $\mu$ V/V	
	-100 mV/V to 100 mV/V		2.5 $\mu$ V/V	
	-2 mV/V to 2 mV/V	Bridge voltage:  1 V	0.20 $\mu$ V/V	
	-5 mV/V to 5 mV/V		0.20 $\mu$ V/V	
	-10 mV/V to 10 mV/V		0.30 $\mu$ V/V	
	-20 mV/V to 20 mV/V		0.50 $\mu$ V/V	
-100 mV/V to 100 mV/V	2.5 $\mu$ V/V			
-2 mV/V to 2 mV/V	Bridge voltage:  2.5 V; 5 V; 7.5 V; 10 V	0.10 $\mu$ V/V		
-5 mV/V to 5 mV/V		0.15 $\mu$ V/V		
-10 mV/V to 10 mV/V		0.25 $\mu$ V/V		
-20 mV/V to 20 mV/V		0.45 $\mu$ V/V		
-100 mV/V to 100 mV/V		2.5 $\mu$ V/V		



**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - high-frequency and radiation measurement quantities**

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
Oscilloscope measured variables	Vertical deflection 5 mV to 5 V 5 mV to 120 V	$R_i = 50 \Omega$	$3.5 \cdot 10^{-3} U \pm 35 \mu V$	Square wave voltage 10 Hz to 10 kHz	
		$R_i = 1 M\Omega$	$2.4 \cdot 10^{-3} U \pm 40 \mu V$		
	Horizontal deflection 5 ns to 520 ms > 20 ms to 5 s			$3 \cdot 10^{-6} t \pm 1 \text{ ns}$	$t$ : current time
				$30 \cdot 10^{-6} t \pm 1.2 \cdot 10^{-3} t^2$	
Rise time	150 ps to 10 ms 250 ps to 10 ms	250 mV	$35 \cdot 10^{-3} - t_r \pm 5 \text{ ps}$	$t_r$ = intrinsic rise time of the oscilloscope	
		> 250 mV to 2.5 V $R_i = 50 \Omega$	$35 \cdot 10^{-3} - t_r \pm 8 \text{ ps}$		
RF impedance (reflection factor) Single measurement $ S_{11} $ Amount $ \Gamma $	0,0 to 1,0	45 MHz to 5 GHz	0,01 0,01 $ \Gamma $	Connector; PC-7; 50 $\Omega$	
		> 5 GHz to 18 GHz	0,015 0,01 $ \Gamma $		
		9 kHz to 5 GHz	0,01 0,01 $ \Gamma $	N connector; 50 $\Omega$	
		> 5 GHz to 18 GHz	0,015 0,01 $ \Gamma $		
Phase $\varphi$	-180° to 180°	9 kHz to 18 GHz $0,1 \leq  \Gamma  \leq 1$	—	N connector; 50 $\Omega$ .	
		45 MHz to 18 GHz $0,1 \leq  \Gamma  \leq 1$	$\arcsin \frac{U( \Gamma )}{ \Gamma } \cdot \frac{180^\circ}{\pi}$		PC-7; 50 $\Omega$
		45 MHz to 26.5 GHz $0,1 \leq  \Gamma  \leq 1$		PC-3.5	

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - high-frequency and radiation measurement quantities

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
RF impedance (reflection factor) Two-port measurement <sub> S11 </sub> & <sub> S22 </sub> Amount $ \Gamma $	0.0 to 1.0	45 MHz to 5 GHz	0,015 0,01 $ \Gamma $	PC-7; 50 Ω <sub>3</sub> )
		> 5 GHz to 18 GHz	0,02 0,01 $ \Gamma $	
		9 kHz to 5 GHz	0,015 0,01 $ \Gamma $	N connector; 50 Ω. <sub>3</sub> )
		> 5 GHz to 18 GHz	0,02 0,01 $ \Gamma $	
Phase φ	-180° to 180	45 MHz to 5 GHz	0,01 0,005 $ \Gamma $	PC-3,5 <sub>3</sub> )
		> 5 GHz to 18 GHz	0,015 0,01 $ \Gamma $	
		> 18 GHz to 26.5 GHz	0,02 0,02 $ \Gamma $	
		9 kHz to 18 GHz $0,1 \leq  \Gamma  \leq 1$	$\arcsin \frac{U( \Gamma )}{ \Gamma } \cdot \frac{180^\circ}{\pi}$	N connector; 50 Ω.
		45 MHz to 18 GHz $0,1 \leq  \Gamma  \leq 1$		PC-7; 50 Ω
		45 MHz to 26.5 GHz $0,1 \leq  \Gamma  \leq 1$		PC-3.5
HF - Attenuation Switchable- Attenuators, Fixed attenuators  Absolute attenuation values	0 dB to 60 dB	9 kHz to 18 GHz	0.3 dB	Connector system: N; 50 Ω $ \Gamma  \leq 0.1$
	> 60 dB to 90 dB			
	0 dB to 60 dB	45 MHz to 20 GHz	0.3 dB	Connector system PC- 3.5; 50 Ω 45 MHz to 20 GHz $ \Gamma  \leq 0.1$
	> 60 dB to 90 dB	> 20 GHz to 26.5 GHz	0.5 dB	
	> 60 dB to 90 dB	45 MHz to 20 GHz	0.3 dB	< 20 GHz to 26.5 GHz $ \Gamma  \leq 0.15$
		> 20 GHz to 26.5 GHz	0.5 dB	
HF power  Power meters	> 1 pW to 0.1 mW	2.5 MHz to 2 GHz	$(0.025 \ 0.14 -  \Gamma ) - P$	Connector system: N, PC- 3.5 ; 50 Ω; $ \Gamma /KG \leq 0.2$
		> 2 GHz to 18 GHz	$(0.049 \ 0.21 -  \Gamma ) - P$	
		> 18 GHz to 26.5GHz	$(0.071 \ 0.32 -  \Gamma ) - P$	
HF power  Signal generators	> 1 pW to 0.1 mW	2.5 MHz to 2 GHz	$(0.035 \ 0.13 -  \Gamma ) - P$	Connector system: N, PC- 3.5 ; 50 Ω; $ \Gamma /KG \leq 0.2$
		> 2 GHz to 18 GHz	$(0.053 \ 0.2 -  \Gamma ) - P$	
		> 18 GHz to 26.5GHz	$(0.074 \ 0.31 -  \Gamma ) - P$	
	0.1 mW to 10 mW	9 kHz to 50 MHz	$17 \cdot 10^{-3} - P$	N connector; 50 Ω <sub>3</sub> ) $ \Gamma  \leq 0,3$
		> 50 MHz to 5 GHz	$22 \cdot 10^{-3} - P$	
		> 5 GHz to 18 GHz	$30 \cdot 10^{-3} - P$	
0.1 mW to 10 mW	50 MHz to 5 GHz	$22 \cdot 10^{-3} - P$	Connector PC-3.5; 50 Ω <sub>3</sub> ) $ \Gamma  \leq 0,3$	
	> 5 GHz to 18 GHz	$32 \cdot 10^{-3} - P$		
	> 18 GHz to 26.5 GHz	$40 \cdot 10^{-3} - P$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - high-frequency and radiation measurement quantities**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
HF power Power meters	0.1 mW to 10 mW	9 kHz to 50 MHz > 50 MHz until 5 GHz > 5 GHz until 18 GHz	$17 \cdot 10^{-3} - P$ $21 \cdot 10^{-3} - P$ $28 \cdot 10^{-3} - P$	N connector; 50 Ω <sub>3</sub>   $\Gamma$   ≤ 0,3
	0.1 mW to 10 mW	50 MHz to 5 GHz > 5 GHz to 18 GHz > 18 GHz to 26.5 GHz	$22 \cdot 10^{-3} - P$ $32 \cdot 10^{-3} - P$ $40 \cdot 10^{-3} - P$	Connector PC-3.5; 50 Ω <sub>3</sub>   $\Gamma$   ≤ 0,3
Signal level difference Measuring devices / sources	0 dBc to 90 dBc	9 kHz to 7 GHz > 7 GHz to 13,6 GHz > 13.6 GHz to 26,5 GHz	1.5 dB 2.3 dB 3 dB	SNR > 20 dB
	> 90 dBc to 100 dBc	9 kHz to 7 GHz > 7 GHz to 13,6 GHz > 13.6 GHz to 26,5 GHz	4.5 dB 4.8 dB 5.3 dB	
Filter bandwidth Measuring devices	1 Hz to 40 MHz		1 %	SNR > 70 dB
HF noise display Receiver / Measuring devices	10 Hz to 50 GHz	-165 dbm/Hz to 0 dBm/Hz	1 dB	
Form factor Measuring devices	1:1 to 4:1		5,5 %	SNR > 20 dB
	> 4:1 to 10:1		7 %	
	> 10:1 to 18:1		8,5 %	

3) When using other connector systems, the measurement uncertainty increases.

Annex to the partial accreditation certificate D-K-15070-01-01

On-site calibration - high-frequency and radiation measurement quantities

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Amplitude modulation: Modulation depth $m$	0.0 to $\leq 1.0$	$f_{HF} < 4$ GHz $f_{MOD} < 1$ MHz	0,004 0,025 · $m$	$f_{HF}$ = carrier frequency $f_{MOD}$ = Modulation freq.
Frequency modulation Frequency deviation $\Delta f$	0 Hz to 5 MHz	$f_{HF} < 4$ GHz $f_{MOD} < 1$ MHz	0.041 · $\Delta f$ 25 Hz	$f_{HF}$ = Carrier frequency $f_{MOD}$ = Modulation freq.
Phase modulation Phase deviation $\Delta \Phi$	0 to (4 MHz / $f_{MOD}$ ) rad	$f_{HF} < 4$ GHz $f_{MOD} < 1$ MHz	0.025 rad 0.041 · $\Delta \Phi$	$f_{HF}$ = Carrier frequency $f_{MOD}$ = Modulation freq.
Distortion factor $k$	> 0.0001 to 0.01 > 0.01 to 0.1 > 0.1 to 0.2	AM demodulation method $f_{HF}$ : 150 kHz to 2 GHz $f_{MOD} = 1$ kHz $P_{HF} = 0$ dBm	0,030 0,029 0,025	$f_{HF}$ = Carrier frequency $f_{MOD}$ = Modulation freq. $P_{HF}$ = Carrier level
	> 0.0001 to 0.01 > 0.01 to 0.1 > 0.1 to 0.2	FM & PM- Demodulation method $f_{HF}$ : 150 kHz to 2 GHz $f_{MOD} = 1$ kHz $P_{HF} = 0$ dBm $\Delta f \leq 50$ kHz	0,09	Absolute measurement uncertainty
Pulsed measurands <sup>1)</sup> Spectral voltage amplitude density (Measure/ Display)	$S_{\sigma} = 13.5$ $\mu$ Vs	DIN EN 55016-1-1:2020 CISPR 16-1-1:2019  CISPR Band A 9 kHz to 0.15 MHz	0.50 dB	Pulse frequency 1 Hz to 100 Hz $\Gamma_G, \Gamma_L \leq 0.05$ (represent)
	$S_{\sigma} = 0.316$ $\mu$ Vs	CISPR Band B > 0.15 MHz to 30 MHz	0.50 dB	Pulse frequency 1 Hz to 1000 Hz $\Gamma_G, \Gamma_L \leq 0.07$ (represent)
	$S_{\sigma} = 0.0044$ $\mu$ Vs	CISPR Band C > 30 MHz to 300 MHz	0.6 dB	Pulse frequency 1 Hz to 1000 Hz $\Gamma_G, \Gamma_L \leq 0.12$ (represent)
	$S_{\sigma} = 0.0044$ $\mu$ Vs	CISPR Band D > 300 MHz to 1 GHz	0.6 dB	Pulse frequency 1 Hz to 1000 Hz $\Gamma_G, \Gamma_L \leq 0.12$ (represent)
HF current transformer clamp <sup>1)</sup> Transfer resistance dB( $\Omega$ )	9 kHz to 100 MHz	DIN EN 55016-1-2:2019	0.3 dB	
	> 100 MHz to 400 MHz	4.4 mA	0.5 dB	
	> 400 MHz to 1 GHz		0.8 dB	
HF Bulk Current Injection Transducer Clamp Insertion Loss dB	9 kHz to 100 MHz	DIN EN 61000-4-6:2014	0.3 dB	
	> 100 MHz to 400 MHz	4.4 mA	0.5 dB	
	> 400 MHz to 1 GHz		1.5 dB	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**On-site calibration - high-frequency and radiation measurement quantities**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase noise	Phase noise related to carrier amplitude in dBc/Hz	Offset frequency referred to carrier frequency		Carrier frequency:
Signal generators	> -87 dBc/Hz	100 Hz	2.5 dB	100 MHz - 1 GHz
	> -99 dBc/Hz	1 kHz	2.5 dB	
	> -104 dBc/Hz	10 kHz	2.5 dB	
	> -111 dBc/Hz	100 kHz	2.5 dB	
	> -131 dBc/Hz	1 MHz	2.5 dB	
	> -137 dBc/Hz	10 MHz	2.5 dB	
	> -80 dBc/Hz	100 Hz	2.5 dB	> 1 MHz - 3 GHz
	> -96 dBc/Hz	1 kHz	2.5 dB	
	> -101 dBc/Hz	10 kHz	2.5 dB	
	> -109 dBc/Hz	100 kHz	2.5 dB	
	> -126 dBc/Hz	1 MHz	2.5 dB	
	> -136 dBc/Hz	10 MHz	2.5 dB	
	> -72 dBc/Hz	100 Hz	2.5 dB	> 3 GHz - 6 GHz
	> -93 dBc/Hz	1 kHz	2.5 dB	
	> -98 dBc/Hz	10 kHz	2.5 dB	
	> -106 dBc/Hz	100 kHz	2.5 dB	
	> -120 dBc/Hz	1 MHz	2.5 dB	
	> -135 dBc/Hz	10 MHz	2.5 dB	
Burst generators				
voltage pulse	100 V to 4400 V	DIN EN 61000-4-4:2013	2,2 %	R <sub>L</sub> = load resistance
Rise time and pulse width	3 ns to 1 μs	under load (R <sub>L</sub> ) at <sub>R<sub>L</sub></sub> = 50 Ω	2,5 %	
Burst duration and burst period	100 ns to 1 s	at <sub>R<sub>L</sub></sub> = 1 kΩ	0,25 %	
Surge generators voltage amplitude	250 V to 7000 V		3,5 %	
Current amplitude	5 A to 5 kA	DIN EN 61000-4-5:2019	3,5 %	
Rise time and pulse width	400 ns to 1 ms	with or without coupling and decoupling network	3,5 %	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Mobile laboratory - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC voltage Measuring devices	0 V 1 mV to 2,2 V > 2,2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		0.1 $\mu$ V $7 \cdot 10^{-6} U$ 1 $\mu$ V $9 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U = \text{set value}$
DC voltage Sources	0 V 1 mV to 100 mV > 100 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		0.1 $\mu$ V $8 \cdot 10^{-6} U$ 1 $\mu$ V $11 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$	$U = \text{measured value}$
High Voltage	> 1 kV to 10 kV		$2.5 \cdot 10^{-3} U$ 2.5 V	$U = \text{measured value}$
Direct current strength Measuring devices and sources	0 A 0.1 $\mu$ A to < 1 $\mu$ A 1 $\mu$ A to < 10 $\mu$ A 10 $\mu$ A to < 100 $\mu$ A 100 $\mu$ A to 320 mA > 320 mA to 1 A > 1 A to 10 A > 10 A to > 150 A 150 A to 2000 A	Precision Open  Normal resistance and  Voltmeter    Current transformer	0.2 nA  $35 \cdot 10^{-6} / 21$ pA  $50 \cdot 10^{-6} / 6$ pA $15 \cdot 10^{-6} / 0.4$ nA $18 \cdot 10^{-6} /$ $15 \cdot 10^{-6} / 6$ $\mu$ A $0,2 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$	$I = \text{set value}$
Direct current strength Current clamps	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 1000 A		$1 \cdot 10^{-3} /$ $2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Mobile laboratory - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks	
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement		
DC resistance	0 Ω		50 μΩ	R = set value Fluke 5700A	
	1 Ω; 1.9 Ω		95 · 10 <sup>-6</sup> R		
	10 Ω		28 · 10 <sup>-6</sup> R		
	19 Ω		27 · 10 <sup>-6</sup> R		
	100 Ω; 190 Ω		17 · 10 <sup>-6</sup> R		
	1 kΩ		13 · 10 <sup>-6</sup> R		
	1.9 kΩ		13 · 10 <sup>-6</sup> R		
	10 kΩ		12 · 10 <sup>-6</sup> R		
	19 kΩ		12 · 10 <sup>-6</sup> R		
	100 kΩ		14 · 10 <sup>-6</sup> R		
	190 kΩ		14 · 10 <sup>-6</sup> R		
	1 MΩ		20 · 10 <sup>-6</sup> R		
	1.9 MΩ		21 · 10 <sup>-6</sup> R		
	10 MΩ		40 · 10 <sup>-6</sup> R		
	19 MΩ		48 · 10 <sup>-6</sup> R		
	100 MΩ		110 · 10 <sup>-6</sup> R		
	0 Ω		100 μΩ		R = measured value HP 3458A
	1 Ω to 10 Ω		16 · 10 <sup>-6</sup> R 50 μΩ		
	> 10 Ω to 100 Ω		12 · 10 <sup>-6</sup> R 500 μΩ		
> 100 Ω to 1 kΩ		15 · 10 <sup>-6</sup> R			
> 1 kΩ to 10 kΩ		15 · 10 <sup>-6</sup> R			
> 10 kΩ to 100 kΩ		15 · 10 <sup>-6</sup> R			
> 100 kΩ to 1 MΩ		35 · 10 <sup>-6</sup> R			
> 1 MΩ to 10 MΩ		150 · 10 <sup>-6</sup> R			
> 10 MΩ to 100 MΩ		600 · 10 <sup>-6</sup> R			
> 100 MΩ to 1 GΩ		5 · 10 <sup>-3</sup> R			
0.001 Ω to 0.1 Ω	Substitution procedure	50 · 10 <sup>-6</sup> R			
> 0.1 Ω to 1 MΩ	with normal resistance	20 · 10 <sup>-6</sup> R			
> 1 MΩ to 100 MΩ		30 · 10 <sup>-6</sup> R			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Mobile laboratory - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC resistance (areas) Measuring devices	1 Ω to < 11 Ω		$0.12 \cdot 10^{-3} R$	<i>R</i> = set value Fluke 5520A /5522A
	11 Ω to < 33 Ω		$33 \cdot 10^{-6} R$	
	33 Ω to < 110 Ω		$29 \cdot 10^{-6} R$	
	110 Ω to < 330 Ω		$28 \cdot 10^{-6} R$	
	330 Ω to < 1.1 kΩ		$28 \cdot 10^{-6} R$	
	1.1 kΩ to < 3.3 kΩ		$28 \cdot 10^{-6} R$	
	3.3 kΩ to < 11 kΩ		$28 \cdot 10^{-6} R$	
	11 kΩ to < 33 kΩ		$28 \cdot 10^{-6} R$	
	33 kΩ to < 110 kΩ		$28 \cdot 10^{-6} R$	
	110 kΩ to < 330 kΩ		$32 \cdot 10^{-6} R$	
	330 kΩ to < 1.1 MΩ		$33 \cdot 10^{-6} R$	
	1.1 MΩ to < 3.3 MΩ		$62 \cdot 10^{-6} R$	
	3.3 MΩ to < 11 MΩ		$0.13 \cdot 10^{-3} R$	
	11 MΩ to < 33 MΩ		$0.25 \cdot 10^{-3} R$	
	33 MΩ to < 110 MΩ		$0.5 \cdot 10^{-3} R$	
	110 MΩ to < 330 MΩ		$3 \cdot 10^{-3} R$	
330 MΩ to < 1.1 GΩ		$15 \cdot 10^{-3} R$		
AC voltage Measuring devices and Sources	1 mV to 2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV to 7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
		> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$	
> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$			
> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$			



Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and Sources	> 7 mV to 22 mV	10 Hz to 20 Hz	$80 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$80 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$65 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$75 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$75 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$95 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$0.19 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.21 \cdot 10^{-3} U$	
	> 22 mV to 70 mV	10 Hz to 20 Hz	$70 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$58 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$35 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$35 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$45 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$55 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$0.11 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
	> 70 mV to 220 mV	10 Hz to 20 Hz	$39 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$35 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$25 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$25 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$28 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$42 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$85 \cdot 10^{-6} U$		
> 500 kHz to 1 MHz		$0.1 \cdot 10^{-3} U$		
> 220 mV to 700 mV	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$22 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$12 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$12 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$14 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$27 \cdot 10^{-6} U$		
	> 500 kHz to 1 MHz	$40 \cdot 10^{-6} U$		

Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and  Sources	> 700 mV to 2.2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$11 \cdot 10^{-6} U$	
> 2.2 V to 7 V	10 Hz to 20 Hz	> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$13 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
> 7 V to 22 V	10 Hz to 20 Hz	> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
> 22 V to 70 V	10 Hz to 20 Hz	> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$	
> 70 V to 220 V	10 Hz to 20 Hz	> 20 Hz to 40 Hz	$18 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$17 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$17 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$32 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$32 \cdot 10^{-6} U$	

Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage Measuring devices and Sources	> 220 V to 1000 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$ $27 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $65 \cdot 10^{-6} U$	$U =$ measured value
High Voltage	> 0.7 kV to 1 kV > 1 kV to 7 kV	50 Hz	$2.5 \cdot 10^{-3} U 0.25 V$ $3.5 \cdot 10^{-3} U 2.0 V$	
Alternating current strength sources and measuring devices	100 $\mu$ A to 1 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$120 \cdot 10^{-6} I$ $160 \cdot 10^{-6} I$ $60 \cdot 10^{-6} I$	$I =$ measured value
	> 1 mA to 10 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$46 \cdot 10^{-6} I$	
	> 10 mA to 1 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$17 \cdot 10^{-6} I$	$I =$ measured value
	> 1 A to 10 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$32 \cdot 10^{-6} I$	
	> 10 A to 20 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$39 \cdot 10^{-6} I$	
	> 20 A to 100 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$69 \cdot 10^{-6} I$ $69 \cdot 10^{-6} I$ $0.17 \cdot 10^{-3} I$	
Current clamps	1 mA to 2.2 A > 2.2 A to 20 A > 20 A to 800 A	40 Hz to 5 kHz 40 Hz to 5 kHz 40 Hz to 65 Hz	$2 \cdot 10^{-3} I$ $3 \cdot 10^{-3} I$ $4 \cdot 10^{-3} I$	$I =$ measured value
Capacity gauges	190 pF to < 400 pF 400 pF to < 1.1 nF 1.1 nF to < 3.3 nF 3.3 nF to < 11 nF 11 nF to < 33 nF 33 nF to < 110 nF 110 nF to < 330 nF 330 nF to < 1.1 $\mu$ F 1.1 $\mu$ F to < 3.3 $\mu$ F 3.3 $\mu$ F to < 11 $\mu$ F	10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 600 Hz 10 Hz to 300 Hz 10 Hz to 150 Hz	$4 \cdot 10^{-3} C 8 pF$ $4.5 \cdot 10^{-3} C 8 pF$ $4.0 \cdot 10^{-3} C 8 pF$ $2.5 \cdot 10^{-3} C 8 pF$ $2.5 \cdot 10^{-3} C 80 pF$ $2.5 \cdot 10^{-3} C 80 pF$ $4.5 \cdot 10^{-3} C$ $4.5 \cdot 10^{-3} C$ $4.5 \cdot 10^{-3} C$ $4.5 \cdot 10^{-3} C$	With 5520A / 5522A

Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Capacity gauges	11 $\mu$ F to < 33 $\mu$ F	10 Hz to 120 Hz	$6.0 \cdot 10^{-3} C$	With 5520A / 5522A
	33 $\mu$ F to < 110 $\mu$ F	10 Hz to 80 Hz	$6.5 \cdot 10^{-3} C$	
	110 $\mu$ F to < 330 $\mu$ F	DC to 50 Hz	$6.0 \cdot 10^{-3} C$	
	330 $\mu$ F to < 1.1 mF	DC to Hz	$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	DC to 6 Hz	$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 2 Hz	$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF	DC to 200.6 Hz	$8.0 \cdot 10^{-3} C$	
	33 mF to 110 mF	DC to 0,2 Hz	$11 \cdot 10^{-3} C$	
	1 nF to 100 nF > 100 nF to 1000 nF	50 Hz to 10 kHz 50 z to 1 kHz >1 kHz to 10 kHz	$1,0 \cdot 10^{-3} C$ $1,0 \cdot 10^{-3} C$ $2,5 \cdot 10^{-3} C$	C: measured value with normal capacitances
Frequency	1 mHz to 46 GHz		$2 \cdot 10^{-9} \cdot f_{UTF}$	$f$ = current measured value $UTF$ = Trigger uncertainty
Time interval	1 $\mu$ s to 1000 s		$2 \cdot 10^{-9} \cdot t$ 2 ns	$t$ = current measured value
Speed optical	1 $\text{min}^{-1}$ to 100.000 $\text{min}^{-1}$	with light pulse generator	$8 \cdot 10^{-6}$ but not less than 0.006 $\text{min}^{-1}$	
AC active power measuring devices		33 mV to 1000 V 45 Hz to 65 kHz $PF = 1$		$P$ = set value with Fluke 5520A/5522A $PF$ : Power factor
	109 $\mu$ W to < 11kW 363 mW up to 20 kW	33 mA to < 11 A 11 A to 20 A	$1.4 \cdot 10^{-3} P$ $2.0 \cdot 10^{-3} P$	
DC power Measuring devices	1 mW to 300 W > 300 W to 20 kW		$0.5 \cdot 10^{-3} P$ $1.0 \cdot 10^{-3} P$	
	Sources	Product of $U$ and $I$ 1 mV $\leq U \leq$ 1000 V 100 $\mu$ A $\leq I \leq$ 2000 A	$30 \cdot 10^{-6} P$ $200 \cdot 10^{-6} P$ $300 \cdot 10^{-6} P$	$P$ : calculated power
AC active power	5 mW to 50 kW > 2.5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	$w(U_F)$ ... Uncertainty of the amplitude of the voltage fundamental $w(I_F)$ ... Uncertainty of the amplitude of the current fundamentals $w(\Phi_F)$ ... Uncertainty of the phase shift angle $w(U_{rnc})$ ... uncertainty of the stress rms value $w(I_{rnc})$ ... Uncertainty of the rms current value
AC reactive power	5 mvar to 50 kvar > 2.5 var to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	

Valid 19.04.2023

Date of issue: 19.04.2023

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Mobile laboratory - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Apparent power	0.1 VA until 50 kVA > 50 VA until 120 kVA	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_p)^2 + w(I_p)^2 + w(\Phi_p)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Mobile laboratory - high frequency and radiation measurement quantities**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Oscilloscope measured variables Vertical deflection	5 mV to 5 V	$R_i = 50 \Omega$	$3.5 \cdot 10^{-3} U \pm 35 \mu V$	Square wave voltage
	5 mV to 120 V	$R_i = 1 M\Omega$	$2.4 \cdot 10^{-3} U \pm 40 \mu V$	10 Hz to 10 kHz
Horizontal deflection	5 ns to 520 ms > 20 ms to 5 s		$3 \cdot 10^{-6} t \pm 1 \text{ ns}$ $30 \cdot 10^{-6} t \pm 1.2 \cdot 10^{-3} t^2$	t: current time
Rise time	150 ps until 10 ms 250 ps until 10 ms	250 mV > 250 mV to 2.5 V $R_i = 50 \Omega$	$35 \cdot 10^{-3} - t_r \pm 5 \text{ ps}$ $35 \cdot 10^{-3} - t_r \pm 8 \text{ ps}$	$t_r$ = natural rise time of the oscilloscope
RF impedance (reflection coefficient) Single measurement <sub>S11</sub> A mount  Γ	0,0 to 1,0	45 MHz to 5 GHz	0,01 0,01  Γ	Connector; PC-7; 50 Ω <sub>3</sub>
		> 5 GHz to 18 GHz	0,015 0,01  Γ	
		9 kHz to 5 GHz	0,01 0,01  Γ	N connector; 50 Ω <sub>3</sub>
		> 5 GHz to 18 GHz	0,015 0,01  Γ	
45 MHz until 5 GHz	0,01 0,005  Γ	Connector; PC-3.5; 50 Ω <sub>3</sub>		
> 5 GHz until 18 GHz	0,015 0,01  Γ			
> 18 GHz to 26.5 GHz	0,02 0,02  Γ			
Phase φ	-180° to 180°	9 kHz to 18 GHz $0,1 \leq  Γ  \leq 1$	—	N connector; 50 Ω.
		45 MHz to 18 GHz $0,1 \leq  Γ  \leq 1$	$\arcsin \frac{U( Γ )}{ Γ } \cdot \frac{180^\circ}{\pi}$	PC-7; 50 Ω
		45 MHz to 26.5 GHz $0,1 \leq  Γ  \leq 1$		PC-3.5

Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - high frequency and radiation measurement quantities

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
RF impedance (reflection factor) Two-port measurement <sub> S<sub>11</sub> </sub> & <sub> S<sub>22</sub> </sub> Amount $ Γ $	0,0 to 1,0	45 MHz to 5 GHz	0,015 0,01 $ Γ $	PC-7; 50 Ω <sub>3</sub> )
		> 5 GHz to 18 GHz	0,02 0,01 $ Γ $	
		9 kHz to 5 GHz	0,015 0,01 $ Γ $	N connector; 50 Ω. <sub>3</sub> )
		> 5 GHz to 18 GHz	0,02 0,01 $ Γ $	
Phase φ	-180° to 180°	45 MHz to 5 GHz	0,01 0,005 $ Γ $	PC-3,5 <sub>3</sub> )
		> 5 GHz to 18 GHz	0,015 0,01 $ Γ $	
		> 18 GHz to 26.5 GHz	0,02 0,02 $ Γ $	
Phase φ	-180° to 180°	9 kHz to 18 GHz $0,1 ≤  Γ  ≤ 1$	$\arcsin \frac{U( Γ )}{ Γ } \cdot \frac{180^\circ}{\pi}$	N connector; 50 Ω.
		45 MHz to 18 GHz $0,1 ≤  Γ  ≤ 1$		PC-7; 50 Ω
		45 MHz to 26.5 GHz $0,1 ≤  Γ  ≤ 1$		PC-3.5
HF - Attenuation Switchable- Attenuators, Fixed attenuators  Absolute Damping values	0 dB to 60 dB	9 kHz to 18 GHz	0.3 dB	Connector system: N; 50 Ω $ Γ  ≤ 0.1$
	> 60 dB to 90 dB		0.3 dB	
	0 dB to 60 dB	45 MHz to 20 GHz	0.3 dB	Connector system PC-3.5; 50 Ω 45 MHz to 20 GHz $ Γ  ≤ 0.1$ <20 GHz to 26.5 GHz $ Γ  ≤ 0.15$
	> 60 dB to 90 dB	45 MHz to 20 GHz > 20 GHz to 26.5 GHz	0.3 dB 0.5 dB	
HF power Power meters	> 1 pW to 0.1 mW	2.5 MHz to 2 GHz	$(0.025 \ 0.14 -  Γ ) - P$	Connector system: N, PC-3.5 ; 50 Ω; $ Γ  / KG ≤ 0.2$
		> 2 GHz to 18 GHz	$(0.049 \ 0.21 -  Γ ) - P$	
		> 18 GHz to 26.5GHz	$(0.071 \ 0.32 -  Γ ) - P$	
HF power Signal generators	> 1 pW to 0.1 mW	2.5 MHz to 2 GHz	$(0.035 \ 0.13 -  Γ ) - P$	Connector system: N, PC-3.5 ; 50 Ω; $ Γ _{KG} ≤ 0.2$
		> 2 GHz to 18 GHz	$(0.053 \ 0.2 -  Γ ) - P$	
		> 18 GHz to 26.5GHz	$(0.074 \ 0.31 -  Γ ) - P$	
	0.1 mW to 10 mW	9 kHz to 50 MHz	$17 \cdot 10^{-3} - P$	N connector; 50 Ω <sub>3</sub> ) $ Γ  ≤ 0,3$
		> 50 MHz until 5 GHz	$22 \cdot 10^{-3} - P$	
		> 5 GHz until 18 GHz	$30 \cdot 10^{-3} - P$	
0.1 mW to 10 mW	50 MHz to 5 GHz	$22 \cdot 10^{-3} - P$	Connector PC-3.5; 50 Ω <sub>3</sub> ) $ Γ  ≤ 0,3$	
	> 5 GHz to 18 GHz	$32 \cdot 10^{-3} - P$		
	> 18 GHz to 26.5 GHz	$40 \cdot 10^{-3} - P$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Mobile laboratory - high frequency and radiation measurement quantities**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Extended uncertainty of measurement	Remarks
	Measuring range	Measuring conditions / Procedures			
HF power Power meters	0.1 mW to 10 mW	9 kHz to 50 MHz > 50 MHz until 5 GHz > 5 GHz until 18 GHz		$17 \cdot 10^{-3} - P$ $21 \cdot 10^{-3} - P$ $28 \cdot 10^{-3} - P$	N connector; 50 Ω <sub>3</sub> $ I  \leq 0,3$
	0.1 mW to 10 mW	50 MHz to 5 GHz > 5 GHz to 18 GHz > 18 GHz to 26.5 GHz		$22 \cdot 10^{-3} - P$ $32 \cdot 10^{-3} - P$ $40 \cdot 10^{-3} - P$	Connector PC-3.5; 50 Ω <sub>3</sub> $ I  \leq 0,3$
Signal level difference Measuring devices / sources	0 dBc to 90 dBc	9 kHz until 7 GHz > 7 GHz until 13.6 GHz > 13.6 GHz until 26.5 GHz		1.5 dB 2.3 dB 3 dB	SNR > 20 dB
	> 90 dBc to 100 dBc	9 kHz until 7 GHz > 7 GHz until 13.6 GHz > 13.6 GHz to 26.5 GHz		4.5 dB 4.8 dB 5.3 dB	SNR > 20 dB
Filter bandwidth Measuring devices	1 Hz to 40 MHz			1 %	SNR > 70 dB
HF noise display Receiver / Measuring devices	10 Hz to 50 GHz	-165 dbm/Hz to 0 dBm/Hz		1 dB	
Form factor Measuring devices	1:1 to 4:1			5,5 %	SNR > 20 dB
	> 4:1 to 10:1			7 %	
	> 10:1 to 18:1			8,5 %	

3) When using other connector systems, the measurement uncertainty increases.



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Mobile laboratory - high frequency and radiation measurement quantities**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Amplitude modulation: Modulation depth $m$	0.0 to $\leq 1.0$	$f_{HF} < 4$ GHz $f_{MOD} < 1$ MHz	0,004 0,025 · $m$	$f_{HF}$ = carrier frequency $f_{MOD}$ = modulation freq.
Frequency modulation Frequency deviation $\Delta f$	0 Hz to 5 MHz	$f_{HF} < 4$ GHz $f_{MOD} < 1$ MHz	0.041 · $\Delta f$ 25 Hz	$f_{HF}$ = Carrier frequency $f_{MOD}$ = Modulation freq.
Phase modulation Phase deviation $\Delta \Phi$	0 to (4 MHz / $f_{MOD}$ ) rad	$f_{HF} < 4$ GHz $f_{MOD} < 1$ MHz	0.025 rad 0.041 · $\Delta \Phi$	$f_{HF}$ = Carrier frequency $f_{MOD}$ = Modulation freq.
Distortion factor $k$	> 0.0001 to 0.01 > 0.01 to 0.1 > 0.1 to 0.2	AM demodulation method $f_{HF}$ : 150 kHz to 2 GHz $f_{MOD} = 1$ kHz $P_{HF} = 0$ dBm	0,030 0,029 0,025	$f_{HF}$ = Carrier frequency $f_{MOD}$ = Modulation freq. $P_{HF}$ = Carrier level
	> 0.0001 to 0.01 > 0.01 to 0.1 > 0.1 to 0.2	FM & PM- Demodulation method $f_{HF}$ : 150 kHz to 2 GHz $f_{MOD} = 1$ kHz $P_{HF} = 0$ dBm $\Delta f \leq 50$ kHz	0,09	Absolute measurement uncertainty
Pulsed measurands <sup>1)</sup> Spectral voltage amplitude density (Measure/ Display)	$S_{\sigma} = 13.5$ $\mu$ Vs	<i>DIN EN 55016-1-1:2020</i> CISPR 16-1-1:2019  CISPR Band A 9 kHz to 0.15 MHz	0.50 dB	Pulse frequency 1 Hz to 100 Hz $\Gamma_G, \Gamma_L \leq 0.05$ (represent)
	$S_{\sigma} = 0.316$ $\mu$ Vs	CISPR Band B > 0.15 MHz to 30 MHz	0.50 dB	Pulse frequency 1 Hz to 1000 Hz $\Gamma_G, \Gamma_L \leq 0.07$ (represent)
	$S_{\sigma} = 0.0044$ $\mu$ Vs	CISPR Band C > 30 MHz to 300 MHz	0.6 dB	Pulse frequency 1 Hz to 1000 Hz $\Gamma_G, \Gamma_L \leq 0.12$ (represent)
	$S_{\sigma} = 0.0044$ $\mu$ Vs	CISPR Band D > 300 MHz to 1 GHz	0.6 dB	Pulse frequency 1 Hz to 1000 Hz $\Gamma_G, \Gamma_L \leq 0.12$ (represent)
HF current transformer clamp <sup>2)</sup> Transfer resistance dB( $\Omega$ )	9 kHz to 100 MHz	DIN EN 55016-1-2:2019	0.3 dB	
	> 100 MHz to 400 MHz	4.4 mA	0.5 dB	
	> 400 MHz to 1 GHz		0.8 dB	
HF Bulk Current Injection <sup>2)</sup> Transducer clamp insertion loss dB	9 kHz to 100 MHz	DIN EN 61000-4-6:2014	0.3 dB	
	> 100 MHz to 400 MHz	4.4 mA	0.5 dB	
	> 400 MHz to 1 GHz		1.5 dB	

Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - high frequency and radiation measurement quantities

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Phase noise  Signal generators	Phase noise related to carrier amplitude in dBc/Hz	Offset frequency referred to carrier frequency		Carrier frequency: 100 MHz - 1 GHz
	> -87 dBc/Hz	100 Hz	2.5 dB	
	> -99 dBc/Hz	1 kHz	2.5 dB	
	> -104 dBc/Hz	10 kHz	2.5 dB	
	> -111 dBc/Hz	100 kHz	2.5 dB	
	> -131 dBc/Hz	1 MHz	2.5 dB	
	> -137 dBc/Hz	10 MHz	2.5 dB	
	> -80 dBc/Hz	100 Hz	2.5 dB	> 1 MHz - 3 GHz
	> -96 dBc/Hz	1 kHz	2.5 dB	
	> -101 dBc/Hz	10 kHz	2.5 dB	
	> -109 dBc/Hz	100 kHz	2.5 dB	
	> -126 dBc/Hz	1 MHz	2.5 dB	
	> -136 dBc/Hz	10 MHz	2.5 dB	
	> -72 dBc/Hz	100 Hz	2.5 dB	> 3 GHz - 6 GHz
	> -93 dBc/Hz	1 kHz	2.5 dB	
	> -98 dBc/Hz	10 kHz	2.5 dB	
	> -106 dBc/Hz	100 kHz	2.5 dB	
	> -120 dBc/Hz	1 MHz	2.5 dB	
> -135 dBc/Hz	10 MHz	2.5 dB		
Burst generators voltage pulse	100 V to 4400 V	DIN EN 61000-4-4 under load ( $R_L$ ) at $R_L = 50 \Omega$ at $R_L = 1 k\Omega$	2,2 %	$R_L$ = load resistance
	Rise time and pulse width 3 ns to 1 $\mu$ s		2,5 %	
	Burst duration and burst period 100 ns to 1 s		0,25 %	
Surge generators voltage amplitude	250 V to 7000 V	DIN EN 61000-4-5 with or without coupling and decoupling network	3,5 %	
	Current amplitude 5 A to 5 kA		3,5 %	
	Rise time and pulse width 400 ns to 1 ms		3,5 %	

Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - Dimensional measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Length *) Cylindrical setting standards, ring gauges: Diameter	1 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ = is the measured Diameter
Plug gauges: Diameter	1 mm to 200 mm	Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Test pins: Diameter	0,1 mm to 30 mm	VDI/VDE/DGQ 2618 Sheet 4.2:2007 Point 3.2.2 (Opt. 1)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Thread gauges (one and multi-start cylindrical male and female threads with straight flanks, symmetrical profile) Threaded mandrels: simple Flank diameter	1.4 mm to 200 mm nominal pitch: 0.3 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.8:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Three-wire method $d$ = is the measured diameter
Threaded rings: simple pitch diameter	3 mm to 200 mm nominal pitch: 0.5 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.9:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Two-ball method $d$ = is the measured diameter
Length of plane-parallel, spherical or cylindrical measuring surfaces  Diameter	0,01 mm to 500 mm  0,01 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 19.1:2014  VDI/VDE/DGQ 2618 Sheet 4.1:2006 Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$  $1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$l$ is the measured length  $d$ is the measured diameter
Feeler gauges	0,03 mm to 2,00 mm	DIN 2275:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Adjustment dimensions for Outside micrometers	25 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 4.4:2009	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	length

Annex to the partial accreditation certificate D-K-15070-01-01

Mobile laboratory - Dimensional measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Throat gauges	3 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.7:2005 Point 3.3.2 (Opt. 2)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured Diameter
Caliper for Exterior, interior and Depth measurements	0 mm until 500 mm	VDI/VDE/DGQ 2618 Sheet 9.1:2006	$30 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured Length
Depth caliper,	> 500 mm until 1000 mm	VDI/VDE/DGQ 2618 Sheet 9.2:2006	$50 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	
Height caliper		VDI/VDE/DGQ 2618 Sheet 9.3:2006		
Outside micrometers	0 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.1:2001	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Fine pointer measurement- screws	0 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 10.3:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Micrometer head screws	0 mm to 50 mm	VDI/VDE/DGQ 2618 Sheet 10.4:2008	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Depth gauges	0 mm to 300 mm	VDI/VDE/DGQ 2618 Sheet 10.5:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 2-point contact on the calibration object	13 mm to 300 mm > 300 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.7:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$ $5 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 3-line contact on the calibration object	3 mm to 150 mm	VDI/VDE/DGQ 2618 Sheet 10.8:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured diameter
Lever gauges (quick probes) for External measurements	to 200 mm	VDI/VDE/DGQ 2618 Sheet 12.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length
Lever gauges (quick probes) for Internal measurements	2 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 13.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Dial gauges	0 mm to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.1:2021	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	mechanical dial gauges
		VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	electronic digital dial gauges
Fine pointer	0 mm to 3 mm	VDI/VDE/DGQ 2618 Sheet 11.2:2002	0.6 $\mu\text{m}$	
Feeler lever gauges	0 mm to 1,6 mm	VDI/VDE/DGQ 2618 Sheet 11.3:2002	1.0 $\mu\text{m}$	
electr. inductive Linear Encoders	to 100 mm	VDI/VDE/DGQ 2618 Sheet 14.1:2010	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	
electr. incremental Linear Encoders	to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	

**Location Calibration Laboratory Munich, Nikolaus-Otto-Straße 2, 85221 Dachau**  
**Permanent Laboratory Munich - Electrical measurands**

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC voltage Measuring devices	0 V 1 mV to 2,2 V > 2,2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		0.1 $\mu$ V $7 \cdot 10^{-6} U$ 1 $\mu$ V $9 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U = \text{set value}$
DC voltage Sources	0 V 1 mV to 100 mV > 100 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		0.1 $\mu$ V $8 \cdot 10^{-6} U$ 1 $\mu$ V $11 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$	$U = \text{measured value}$
High Voltage	> 1 kV to 10 kV		$2.5 \cdot 10^{-3} U$ 2.5 V	$U = \text{measured value}$
DC current strength meters and sources	0 A 0.1 $\mu$ A until < 1 $\mu$ A 1 $\mu$ A until < 10 $\mu$ A 10 $\mu$ A until < 100 $\mu$ A 100 $\mu$ A until 320 mA > 320 mA until 1 A > 1 A to 10 A > 10 A > 150 A 150 A to 2000 A	Precision Open Normal resistance and voltmeter Current transformer	0.2 nA $35 \cdot 10^{-6} / 21$ pA $50 \cdot 10^{-6} / 6$ pA $15 \cdot 10^{-6} / 0.4$ nA $18 \cdot 10^{-6} /$ $15 \cdot 10^{-6} / 6$ $\mu$ A $0,2 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$	$I = \text{set value}$
Direct current strength Current clamps	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 1000 A		$1 \cdot 10^{-3} /$ $2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks		
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement			
DC resistance	0 Ω		50 μΩ	R = set value Fluke 5700A		
	1 Ω; 1.9 Ω		95 · 10 <sup>-6</sup> R			
	10 Ω		28 · 10 <sup>-6</sup> R			
	19 Ω		27 · 10 <sup>-6</sup> R			
	100 Ω; 190 Ω		17 · 10 <sup>-6</sup> R			
	1 kΩ		13 · 10 <sup>-6</sup> R			
	1.9 kΩ		13 · 10 <sup>-6</sup> R			
	10 kΩ		12 · 10 <sup>-6</sup> R			
	19 kΩ		12 · 10 <sup>-6</sup> R			
	100 kΩ		14 · 10 <sup>-6</sup> R			
	190 kΩ		14 · 10 <sup>-6</sup> R			
	1 MΩ		20 · 10 <sup>-6</sup> R			
	1.9 MΩ		21 · 10 <sup>-6</sup> R			
	10 MΩ		40 · 10 <sup>-6</sup> R			
	19 MΩ		48 · 10 <sup>-6</sup> R			
	100 MΩ		110 · 10 <sup>-6</sup> R			
		0 Ω			100 μΩ	R = measured value HP 3458A
		1 Ω to 10 Ω			16 · 10 <sup>-6</sup> R 50 μΩ	
> 10 Ω to 100 Ω			12 · 10 <sup>-6</sup> R 500 μΩ			
> 100 Ω to 1 kΩ			15 · 10 <sup>-6</sup> R			
> 1 kΩ to 10 kΩ			15 · 10 <sup>-6</sup> R			
> 10 kΩ to 100 kΩ			15 · 10 <sup>-6</sup> R			
> 100 kΩ to 1 MΩ			35 · 10 <sup>-6</sup> R			
> 1 MΩ to 10 MΩ			150 · 10 <sup>-6</sup> R			
> 10 MΩ to 100 MΩ			600 · 10 <sup>-6</sup> R			
> 100 MΩ to 1 GΩ			5 · 10 <sup>-3</sup> R			
	0,001 Ω to 0,1 Ω	Substitution procedure	50 · 10 <sup>-6</sup> R			
	> 0,1 Ω to 1 MΩ	with normal resistance	20 · 10 <sup>-6</sup> R			
	> 1 MΩ to 100 MΩ		30 · 10 <sup>-6</sup> R			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC resistance	1 Ω to < 11 Ω		$0.12 \cdot 10^{-3} R$	<i>R</i> = set value
(areas)	11 Ω to < 33 Ω		$33 \cdot 10^{-6} R$	Fluke 5520A /5522A
Measuring devices	33 Ω to < 110 Ω		$29 \cdot 10^{-6} R$	
	110 Ω to < 330 Ω		$28 \cdot 10^{-6} R$	
	330 Ω to < 1.1 kΩ		$28 \cdot 10^{-6} R$	
	1.1 kΩ to < 3.3 kΩ		$28 \cdot 10^{-6} R$	
	3.3 kΩ to < 11 kΩ		$28 \cdot 10^{-6} R$	
	11 kΩ to < 33 kΩ		$28 \cdot 10^{-6} R$	
	33 kΩ to < 110 kΩ		$28 \cdot 10^{-6} R$	
	110 kΩ to < 330 kΩ		$32 \cdot 10^{-6} R$	
	330 kΩ to < 1.1 MΩ		$33 \cdot 10^{-6} R$	
	1.1 MΩ to < 3.3 MΩ		$62 \cdot 10^{-6} R$	
	3.3 MΩ to < 11 MΩ		$0.13 \cdot 10^{-3} R$	
	11 MΩ to < 33 MΩ		$0.25 \cdot 10^{-3} R$	
	33 MΩ to < 110 MΩ		$0.5 \cdot 10^{-3} R$	
	110 MΩ to < 330 MΩ		$3 \cdot 10^{-3} R$	
	330 MΩ to < 1.1 GΩ		$15 \cdot 10^{-3} R$	
AC resistance	0,1 Ω to 2 Ω	50 Hz to 400 Hz	$10 \cdot 10^{-3} R$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	1 mV to 2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV to 7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
		> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$	
	> 7 mV to 22 mV	10 Hz to 20 Hz	$80 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$80 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$65 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$75 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$75 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$95 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$0.19 \cdot 10^{-3} U$		
> 500 kHz to 1 MHz		$0.21 \cdot 10^{-3} U$		
> 22 mV to 70 mV	10 Hz to 20 Hz	$70 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$58 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$35 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$35 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$45 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$55 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$0.11 \cdot 10^{-3} U$		
	> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$		



Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Munich - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 70 mV to 220 mV	10 Hz to 20 Hz	$39 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$35 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$25 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$25 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$28 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$42 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$85 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.1 \cdot 10^{-3} U$	
	> 220 mV to 700 mV	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$22 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$12 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$12 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$14 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$27 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$40 \cdot 10^{-6} U$	
	> 700 mV to 2.2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$11 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$11 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$22 \cdot 10^{-6} U$		
> 500 kHz to 1 MHz		$68 \cdot 10^{-6} U$		
> 2.2 V to 7 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$13 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$		
	> 500 kHz to 1 MHz	$95 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 7 V to 22 V	10 Hz to 20 Hz	$17 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
	> 22 V to 70 V	> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
		> 500 kHz to 1MHz	$0.11 \cdot 10^{-3} U$	
	> 70 V to 220 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$	
	> 220 V to 1000 V	> 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
High Voltage	> 0.7 kV to 1 kV > 1 kV to 7 kV	10 Hz to 20 Hz	$19 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$18 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$17 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$17 \cdot 10^{-6} U$	
	> 50 kHz to 100 kHz	> 50 kHz to 100 kHz	$32 \cdot 10^{-6} U$	
		> 10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$27 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$45 \cdot 10^{-6} U$	
	> 20 kHz to 50 kHz	> 20 kHz to 50 kHz	$45 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$65 \cdot 10^{-6} U$	
		50 Hz	$2.5 \cdot 10^{-3} U$ 0.25 V	
			$3.5 \cdot 10^{-3} U$ 2.0 V	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC power Sources and measuring devices	100 $\mu$ A to 1 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	120 $\cdot$ 10 <sup>-6</sup> / 160 $\cdot$ 10 <sup>-6</sup> / 60 $\cdot$ 10 <sup>-6</sup> /	/ = measured value
	> 1 mA to 10 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	46 $\cdot$ 10 <sup>-6</sup> /	
	> 10 mA to 1 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	17 $\cdot$ 10 <sup>-6</sup> /	
	> 1 A to 10 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	32 $\cdot$ 10 <sup>-6</sup> /	
	> 10 A to 20 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	39 $\cdot$ 10 <sup>-6</sup> /	
Current clamps	1 mA to 2,2 A	40 Hz to 5 kHz	2 $\cdot$ 10 <sup>-3</sup> /	
	> 2,2 A to 20 A	40 Hz to 5 kHz	3 $\cdot$ 10 <sup>-3</sup> /	
	> 20 A to 800 A	40 Hz to 65 Hz	4 $\cdot$ 10 <sup>-3</sup> /	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Munich - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Capacity gauges	190 pF to < 400 pF	10 Hz to 10 kHz	$4 \cdot 10^{-3} C$ 8 pF	With 5520A / 5522A
	400 pF to < 1.1 nF	10 Hz to 10 kHz	$4.5 \cdot 10^{-3} C$ 8 pF	
	1.1 nF to < 3.3 nF	10 Hz to kHz	$4.0 \cdot 10^{-3} C$ 8 pF	
	3.3 nF to < 11 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 8 pF	
	11 nF to < 33 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	33 nF to < 110 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	110 nF to < 330 nF	10 Hz to 1 kHz	$4.5 \cdot 10^{-3} C$	
	330 nF to < 1.1 μF	10 Hz to 600 Hz	$4.5 \cdot 10^{-3} C$	
	1.1 μF to < 3.3 μF	10 Hz to 300 Hz	$4.5 \cdot 10^{-3} C$	
	3.3 μF to < 11 μF	10 Hz to 150 Hz	$4.5 \cdot 10^{-3} C$	
	11 μF to < 33 μF	10 Hz to 120 Hz	$6.0 \cdot 10^{-3} C$	
	33 μF to < 110 μF	10 Hz to 80 Hz	$6.5 \cdot 10^{-3} C$	
	110 μF to < 330 μF	DC to 50 Hz	$6.0 \cdot 10^{-3} C$	
	330 μF to < 1.1 mF	DC to Hz	$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	DC to 6 Hz	$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 2 Hz	$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF	DC to 200.6 Hz	$8.0 \cdot 10^{-3} C$	
	33 mF to 110 mF	DC to 0,2 Hz	$11 \cdot 10^{-3} C$	
	1 nF to 100 nF	50 Hz to 10 kHz	$1,0 \cdot 10^{-3} C$	C: measured value with normal capacitances
	> 100 nF to 1000 nF	50 Hz to 1 kHz	$1,0 \cdot 10^{-3} C$	
		> 1kHz to 10 kHz	$2,5 \cdot 10^{-3} C$	
Frequency	1 mHz to 1 GHz		$2 \cdot 10^{-9} \cdot f$ $U_{Tf}$	$f$ = current measured value  $U_{Tf}$ = Trigger uncertainty
Time interval	1 μs to 1000 s		$2 \cdot 10^{-9} \cdot t$ 2 ns	$t$ = current measured value
Speed optical	1 min <sup>-1</sup> up to 100.000 min <sup>-1</sup>	with light pulse generator	$8 \cdot 10^{-6}$ but not less than 0.006 min <sup>-1</sup>	
AC active power measuring devices	109 μW to < 11kW	33 mV to 1000 V 45 Hz to 65 kHz  PF = 1	$1.4 \cdot 10^{-3} P$	P = set value with Fluke 5520A/5522A PF: Power factor
	363 mW up to 20 kW	33 mA to < 11 A 11 A to 20 A	$2.0 \cdot 10^{-3} P$	
DC power Measuring devices	1 mW to 300 W		$0.5 \cdot 10^{-3} P$	
	> 300 W to 20 kW		$1.0 \cdot 10^{-3} P$	
Sources	1 mW to 300 W	Product of U and I   1 mV ≤ U ≤ 1000 V 100 μA ≤ I ≤ 2000 A	$30 \cdot 10^{-6} P$	P Calculat ed power
	> 300 W to 1 kW		$200 \cdot 10^{-6} P$	
	> 1 kW to 1 MW		$300 \cdot 10^{-6} P$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power	5 mW to 50 kW > 2,5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	$w(U_F)$ . .Uncertainty of the amplitude of the voltage fundamental $w(I_F)$ . .Uncertainty of the amplitude Current fundamental $w(\Phi_F)$ ...Uncertainty of the phase shift angle $w(U_{rmc})$ ... Uncertainty of the voltage rms value $w(I_{rmc})$ ... uncertainty of the rms current value
AC reactive power	5 mvar to 50 kvar > 2,5 var to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	
Apparent power	0.1 VA to 50 kVA > 50 VA to 120 kVA	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Munich - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Stress ratio	± 2 mV/V	Bridge voltage: 5 V		Calibration of 350 Ω bridge standards and the associated indicators  at discrete points In 10 % steps
		Measuring frequency 225 Hz	0.04 μV/V	
		Measuring frequency 600 Hz	0.05 μV/V	
	± 2 mV/V	Measuring frequency 4.8 kHz	0.12 μV/V	
		Bridge voltage: 2,5 V		
		Measuring frequency 225 Hz	0.04 μV/V	
	± 5 mV/V	Measuring frequency 600 Hz	0.04 μV/V	
		Measuring frequency 4.8 kHz	0.12 μV/V	
		Bridge voltage: 5 V		
	± 10 mV/V	Measuring frequency 225 Hz	0.06 μV/V	
		Measuring frequency 4.8 kHz	0.22 μV/V	
Bridge voltage: 5 V				
± 10 mV/V	Measuring frequency 225 Hz	0.06 μV/V		
	Measuring frequency 4.8 kHz	0.45 μV/V		
	Bridge voltage: 5 V			
± 5 mV/V	Measuring frequency 225 Hz	0.06 μV/V		
	Measuring frequency 600 Hz	0.06 μV/V		
	Measuring frequency 4.8 kHz	0.22 μV/V		
± 10 mV/V	Bridge voltage: 2,5 V			
	Measuring frequency 225 Hz	0.06 μV/V		
	Measuring frequency 600 Hz	0.10 μV/V		
± 10 mV/V	Measuring frequency 4.8 kHz	0.45 μV/V		
	Bridge voltage: 1 V			
	Measuring frequency 600 Hz	0.11 μV/V		
± 20 mV/V	Bridge voltage: 1 V			
	Measuring frequency 4.8 kHz	0.6 μV/V		
± 100 mV/V	Bridge voltage: 1 V			
	Measuring frequency 4.8 kHz	3.5 μV/V		
± 100 mV/V	Bridge voltage: 2,5 V			
	Measuring frequency 4.8 kHz	4.0 μV/V		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Munich - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Voltage ratio DC voltage bridge standards	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  0,5 V	2.0 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V 2.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  1 V	1.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V 2.0 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  2,5 V	0.5 $\mu$ V/V 0.5 $\mu$ V/V 0.5 $\mu$ V/V 0.5 $\mu$ V/V 0.5 $\mu$ V/V 1.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  5 V	0.30 $\mu$ V/V 0.25 $\mu$ V/V 0.25 $\mu$ V/V 0.25 $\mu$ V/V 0.35 $\mu$ V/V 1.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  7,5 V	0.20 $\mu$ V/V 0.20 $\mu$ V/V 0.20 $\mu$ V/V 0.20 $\mu$ V/V 0.3 $\mu$ V/V 1.5 $\mu$ V/V	
	0 mV/V -2 mV/V to 2 mV/V -5 mV/V to 5 mV/V	Bridge voltage:  10 V	0.10 $\mu$ V/V 0.15 $\mu$ V/V 0.15 $\mu$ V/V	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Munich - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Voltage ratio DC voltage Bridge standards	-10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V		0.20 $\mu$ V/V 0.3 $\mu$ V/V 1.5 $\mu$ V/V	
Voltage ratio DC voltage bridges, measuring instruments, measuring amplifiers	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  0,5 V	0.35 $\mu$ V/V 0.35 $\mu$ V/V 0.40 $\mu$ V/V 0.55 $\mu$ V/V 2.5 $\mu$ V/V	
	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  1 V	0.20 $\mu$ V/V 0.20 $\mu$ V/V 0.3 $\mu$ V/V 0.5 $\mu$ V/V 2.5 $\mu$ V/V	
	-2 mV/V to 2 mV/V -5 mV/V to 5 mV/V -10 mV/V to 10 mV/V -20 mV/V to 20 mV/V -100 mV/V to 100 mV/V	Bridge voltage:  2.5 V; 5 V; 7.5 V; 10 V	0.10 $\mu$ V/V 0.15 $\mu$ V/V 0.25 $\mu$ V/V 0.45 $\mu$ V/V 2.5 $\mu$ V/V	
Charge charge amplifiers, charge meters	1 pC to 10 <sup>4</sup> pC	0.2 Hz to < 1 Hz	0,5 %	
		1 Hz to 10 kHz	0,4 %	
		> 10 kHz to 20 kHz	0,6 %	
		> 20 kHz to 50 kHz	1,0 %	
Oscilloscope measured variables				<i>U</i> - measured value
Vertical deflection	5 mV to 5 V 5 mV to 120 V	$R_i = 50 \Omega$ $R_i = 1 M\Omega$	$3.5 \cdot 10^{-3} U$ 35 $\mu$ V $2.4 \cdot 10^{-3} U$ 40 $\mu$ V	Square wave voltage 10 Hz to 10 kHz
Horizontal deflection	5 ns to 520 ms > 20 ms to 5 s		$3 \cdot 10^{-6} T$ 1 ns $30 \cdot 10^{-6} T$ $1.2 \cdot 10^{-4} T^2$	
Rise time	150 ps to 10 ms 250 ps to 10 ms	250 mV > 250 mV to 2.5 V	$35 \cdot 10^{-3} - t_r$ 5 ps $35 \cdot 10^{-3} - t_r$ 8 ps	$t_r$ = intrinsic rise time of the oscilloscope



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Dimensional measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Length *) Cylindrical setting standards, ring gauges: Diameter	1 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ = is the measured Diameter
Plug gauges: Diameter	1 mm to 200 mm	Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	Diameter
Test pins: Diameter	0.1 mm to 30 mm	VDI/VDE/DGQ 2618 Sheet 4.2:2007 Point 3.2.2 (Opt. 1)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Thread gauges (one and multi-start cylindrical male and female threads with straight flanks, symmetrical profile) Threaded mandrels: simple Flank diameter	1.4 mm to 200 mm Nominal slope: 0.3 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.8:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Three-wire method $d$ = is the measured diameter
Threaded rings: simpler Flank diameter	3 mm to 200 mm nominal pitch: 0.5 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.9:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Two-ball method $d$ = is the measured Diameter
Length of plane-parallel, spherical or cylindrical Measuring surfaces	0.01 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 19.1:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	$l$ is the measured Length
Diameter	0.01 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006 Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ is the measured Diameter
Feeler gauges	0.03 mm to 2.00 mm	DIN 2275:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Adjustment dimensions for Outside micrometers	25 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 4.4:2009	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Munich - Dimensional measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Throat gauges	3 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.7:2005 Point 3.3.2 (Opt. 2)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ is the measured Diameter
Caliper for Exterior, interior and Depth measurements	0 mm until 500 mm	VDI/VDE/DGQ 2618 Sheet 9.1:2006	$30 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	$l$ is the measured Length
Depth caliper,	> 500 mm until 1000 mm	VDI/VDE/DGQ 2618 Sheet 9.2:2006	$50 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	
Height caliper		VDI/VDE/DGQ 2618 Sheet 9.3:2006		
Outside micrometers	0 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.1:2001	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Fine pointer measurement- screws	0 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 10.3:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Micrometer head screws	0 mm to 50 mm	VDI/VDE/DGQ 2618 Sheet 10.4:2008	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Depth gauges	0 mm to 300 mm	VDI/VDE/DGQ 2618 Sheet 10.5:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 2-point contact on the calibration object	13 mm to 300 mm > 300 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.7:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$ $5 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 3-line contact on the calibration object	3 mm to 100 mm	VDI/VDE/DGQ 2618 Sheet 10.8:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	$d$ is the measured diameter
Lever gauges (quick probes) for External measurements	to 200 mm	VDI/VDE/DGQ 2618 Sheet 12.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Lever gauges (quick probes) for Internal measurements	2 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 13.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Dial gauges	0 mm to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.1:2021	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	mechanical dial gauges
		VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	electronic digital dial gauges
Fine pointer	0 mm to 3 mm	VDI/VDE/DGQ 2618 Sheet 11.2:2002	$0.6 \mu\text{m}$	
Feeler lever gauges	0 mm to 1,6 mm	VDI/VDE/DGQ 2618 Sheet 11.3:2002	$1.0 \mu\text{m}$	
Rotation angle Direct rotary encoders	0° to 360°	VDI/VDE 2648 Sheet 1	$0,06^\circ$	
Indirect rotary encoders	0° to 360°	VDI/VDE 2648 Sheet 2 Rotation speed > 0.21/min Rotation speed < 0.21/min	$0,5^\circ$ $1,0^\circ$	

**Location Calibration Laboratory Essen, Alte Landstraße 3c, 45329 Essen, Germany**

**Permanent Laboratory Essen - Electrical Measurands**

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC voltage Measuring devices	0 V 1 mV to 2,2 V > 2,2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		0.1 $\mu$ V $7 \cdot 10^{-6} U$ 1 $\mu$ V $9 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U = \text{set value}$
DC voltage Sources	0 V 1 mV to 100 mV > 100 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		0.1 $\mu$ V $8 \cdot 10^{-6} U$ 1 $\mu$ V $11 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$	$U = \text{measured value}$
High Voltage	> 1 kV to 10 kV		$2.5 \cdot 10^{-3} U$ 2.5 V	$U = \text{measured value}$
DC current strength meters and sources	0 A 0.1 $\mu$ A until < 1 $\mu$ A 1 $\mu$ A until < 10 $\mu$ A 10 $\mu$ A until < 100 $\mu$ A 100 $\mu$ A until 320 mA > 320 mA until 1 A > 1 A to 10 A > 10 A > 150 A 150 A to 2000 A	Precision Open Normal resistance and voltmeter Current transformer	0.2 nA $35 \cdot 10^{-6} / 21$ pA $50 \cdot 10^{-6} / 6$ pA $15 \cdot 10^{-6} / 0.4$ nA 18 - $10^{-6} /$ $15 \cdot 10^{-6} / 6$ $\mu$ A 0,2 - $10^{-3} /$ 0,3 - $10^{-3} /$ 0,3 - $10^{-3} /$	$I = \text{set value}$
Direct current strength Current clamps	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 1000 A		$1 \cdot 10^{-3} /$ $2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Essen - Electrical Measurands**

Calibration and Measurement Capabilities (CMC)						
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks		
DC resistance	0 Ω		50 μΩ	R = set value Fluke 5700A		
	1 Ω; 1.9 Ω		95 · 10 <sup>-6</sup> R			
	10 Ω		28 · 10 <sup>-6</sup> R			
	19 Ω		27 · 10 <sup>-6</sup> R			
	100 Ω; 190 Ω		17 · 10 <sup>-6</sup> R			
	1 kΩ		13 · 10 <sup>-6</sup> R			
	1.9 kΩ		13 · 10 <sup>-6</sup> R			
	10 kΩ		12 · 10 <sup>-6</sup> R			
	19 kΩ		12 · 10 <sup>-6</sup> R			
	100 kΩ		14 · 10 <sup>-6</sup> R			
	190 kΩ		14 · 10 <sup>-6</sup> R			
	1 MΩ		20 · 10 <sup>-6</sup> R			
	1.9 MΩ		21 · 10 <sup>-6</sup> R			
	10 MΩ		40 · 10 <sup>-6</sup> R			
	19 MΩ		48 · 10 <sup>-6</sup> R			
	100 MΩ		110 · 10 <sup>-6</sup> R			
		0 Ω			100 μΩ	R = measured value HP 3458A
		1 Ω to 10 Ω			16 · 10 <sup>-6</sup> R 50 μΩ	
		> 10 Ω to 100 Ω			12 · 10 <sup>-6</sup> R 500 μΩ	
> 100 Ω to 1 kΩ			15 · 10 <sup>-6</sup> R			
> 1 kΩ to 10 kΩ			15 · 10 <sup>-6</sup> R			
> 10 kΩ to 100 kΩ			15 · 10 <sup>-6</sup> R			
> 100 kΩ to 1 MΩ			35 · 10 <sup>-6</sup> R			
> 1 MΩ to 10 MΩ			150 · 10 <sup>-6</sup> R			
> 10 MΩ to 100 MΩ			600 · 10 <sup>-6</sup> R			
> 100 MΩ to 1 GΩ			5 · 10 <sup>-3</sup> R			
	0,001 Ω to 0,1 Ω	Substitution procedure with normal resistance	50 · 10 <sup>-6</sup> R			
	> 0,1 Ω to 1 MΩ		20 · 10 <sup>-6</sup> R			
	> 1 MΩ to 100 MΩ		30 · 10 <sup>-6</sup> R			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Essen - Electrical Measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC resistance (areas) Measuring devices	1 Ω to < 11 Ω		$0.12 \cdot 10^{-3} R$	<i>R</i> = set value
	11 Ω to < 33 Ω		$33 \cdot 10^{-6} R$	Fluke 5520A /5522A
	33 Ω to < 110 Ω		$29 \cdot 10^{-6} R$	
	110 Ω to < 330 Ω		$28 \cdot 10^{-6} R$	
	330 Ω to < 1.1 kΩ		$28 \cdot 10^{-6} R$	
	1.1 kΩ to < 3.3 kΩ		$28 \cdot 10^{-6} R$	
	3.3 kΩ to < 11 kΩ		$28 \cdot 10^{-6} R$	
	11 kΩ to < 33 kΩ		$28 \cdot 10^{-6} R$	
	33 kΩ to < 110 kΩ		$28 \cdot 10^{-6} R$	
	110 kΩ to < 330 kΩ		$32 \cdot 10^{-6} R$	
	330 kΩ to < 1.1 MΩ		$33 \cdot 10^{-6} R$	
	1.1 MΩ to < 3.3 MΩ		$62 \cdot 10^{-6} R$	
	3.3 MΩ to < 11 MΩ		$0.13 \cdot 10^{-3} R$	
	11 MΩ to < 33 MΩ		$0.25 \cdot 10^{-3} R$	
	33 MΩ to < 110 MΩ		$0.5 \cdot 10^{-3} R$	
110 MΩ to < 330 MΩ		$3 \cdot 10^{-3} R$		
330 MΩ to < 1.1 GΩ		$15 \cdot 10^{-3} R$		
AC resistance	0,1 Ω to 2 Ω	50 Hz to 400 Hz	$10 \cdot 10^{-3} \cdot R$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Electrical Measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC power  Sources and measuring devices	1 mV to 2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV to 7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
		> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$	
	> 7 mV to 22 mV	10 Hz to 20 Hz	$80 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$80 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$65 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$75 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$75 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$95 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$0.19 \cdot 10^{-3} U$		
> 500 kHz to 1 MHz		$0.21 \cdot 10^{-3} U$		
> 22 mV to 70 mV	10 Hz to 20 Hz	$70 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$58 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$35 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$35 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$45 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$55 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$0.11 \cdot 10^{-3} U$		
	> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Electrical Measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 70 mV to 220 mV	10 Hz to 20 Hz	$39 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$35 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$25 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$25 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$28 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$42 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$85 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.1 \cdot 10^{-3} U$	
	> 220 mV to 700 mV	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$22 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$12 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$12 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$14 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$27 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$40 \cdot 10^{-6} U$	
	> 700 mV to 2.2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$11 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$11 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$22 \cdot 10^{-6} U$		
> 500 kHz to 1 MHz		$68 \cdot 10^{-6} U$		
> 2.2 V to 7 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$13 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$		
	> 500 kHz to 1 MHz	$95 \cdot 10^{-6} U$		

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Essen - Electrical Measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 7 V to 22 V	10 Hz to 20 Hz	$17 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
	> 22 V to 70 V	> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.11 \cdot 10^{-3} U$	
	> 70 V to 220 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$	
	> 220 V to 1000 V	> 50 kHz to 100 kHz	$15 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
High Voltage	> 0.7 kV to 1 kV	50 Hz	$2.5 \cdot 10^{-3} U$ 0.25 V	
	> 1 kV to 7 kV		$3.5 \cdot 10^{-3} U$ 2.0 V	



Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Electrical Measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC power Sources and measuring devices	100 $\mu$ A to 1 mA	10 Hz to 40 Hz  > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	120 $\cdot$ 10 <sup>-6</sup> / 160 $\cdot$ 10 <sup>-6</sup> / 60 $\cdot$ 10 <sup>-6</sup> /	/ = measured value
	> 1 mA to 10 mA	10 Hz to 40 Hz  > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	46 $\cdot$ 10 <sup>-6</sup> /	
Current clamps	> 10 mA to 1 A	10 Hz to 40 Hz  > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	17 $\cdot$ 10 <sup>-6</sup> /	/ = measured value
	> 1 A to 10 A	10 Hz to 40 Hz  > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	32 $\cdot$ 10 <sup>-6</sup> /	
	> 10 A to 20 A	10 Hz to 40 Hz  > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	39 $\cdot$ 10 <sup>-6</sup> /	
	1 mA to 2.2 A	40 Hz to 5 kHz	2 $\cdot$ 10 <sup>-3</sup> /	
	> 2.2 A to 20 A	40 Hz to 5 kHz	3 $\cdot$ 10 <sup>-3</sup> /	
	> 20 A to 800 A	40 Hz to 65 Hz	4 $\cdot$ 10 <sup>-3</sup> /	
	Capacity gauges	190 pF to < 400 pF	10 Hz to 10 kHz	
400 pF to < 1.1 nF	10 Hz to 10 kHz	4.5 $\cdot$ 10 <sup>-3</sup> C 8 pF		
1.1 nF to < 3.3 nF	10 Hz to 3 kHz	4.0 $\cdot$ 10 <sup>-3</sup> C 8 pF		
3.3 nF to < 11 nF	10 Hz to 1 kHz	2.5 $\cdot$ 10 <sup>-3</sup> C 8 pF		
11 nF to < 33 nF	10 Hz to 1 kHz	2.5 $\cdot$ 10 <sup>-3</sup> C 80 pF		
33 nF to < 110 nF	10 Hz to 1 kHz	2.5 $\cdot$ 10 <sup>-3</sup> C 80 pF		
110 nF to < 330 nF	10 Hz to 1 kHz	4.5 $\cdot$ 10 <sup>-3</sup> C		
330 nF to < 1.1 $\mu$ F	10 Hz to 600 Hz	4.5 $\cdot$ 10 <sup>-3</sup> C		
1.1 $\mu$ F to < 3.3 $\mu$ F	10 Hz to 300 Hz	4.5 $\cdot$ 10 <sup>-3</sup> C		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Electrical Measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Capacity gauges	3.3 $\mu$ F to < 11 $\mu$ F	10 Hz to 150 Hz	$4.5 \cdot 10^{-3} C$	With 5520A / 5522A
	11 $\mu$ F to < 33 $\mu$ F	10 Hz to 120 Hz	$6.0 \cdot 10^{-3} C$	
	33 $\mu$ F to < 110 $\mu$ F	10 Hz to 80 Hz	$6.5 \cdot 10^{-3} C$	
	110 $\mu$ F to < 330 $\mu$ F	DC to 50 Hz	$6.0 \cdot 10^{-3} C$	
	330 $\mu$ F to < 1.1 mF	DC to 20 Hz	$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	DC to 6 Hz	$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 2 Hz	$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF	DC to 0.6 Hz	$8.0 \cdot 10^{-3} C$	
	33 mF to 110 mF	DC to 0.2 Hz	$11 \cdot 10^{-3} C$	
	1 nF to 100 nF	50 Hz to 10 kHz	$1,0 \cdot 10^{-3} C$	C: measured value with normal capacitances
	> 100 nF to 1000 nF	50 Hz to 1 kHz	$1,0 \cdot 10^{-3} C$	
		> 1 kHz to 10 kHz	$2,5 \cdot 10^{-3} C$	
Frequency	1 mHz to 1 GHz		$2 \cdot 10^{-9} \cdot f_{UTJ}$	$f$ = current measured value $UTJ$ = Trigger uncertainty
Time interval	1 $\mu$ s to 1000 s		$2 \cdot 10^{-9} \cdot t$ ns	$t$ = current measured value
Speed optical	1 $\text{min}^{-1}$ up to 100.000 $\text{min}^{-1}$	with light pulse generator	$8 \cdot 10^{-6}$ but not less than 0.006 $\text{min}^{-1}$	
AC active power measuring devices		33 mV to 1000 V 45 Hz to 65 kHz $PF = 1$		$P$ = set value with Fluke 5520A/5522A $PF$ : Power factor
	109 $\mu$ W to < 11kW	33 mA to < 11 A	$1.4 \cdot 10^{-3} P$	
	363 mW up to 20 kW	11 A to 20 A	$2.0 \cdot 10^{-3} P$	
DC power Measuring devices	1 mW to 300 W		$0.5 \cdot 10^{-3} P$	
	> 300 W to 20 kW		$1.0 \cdot 10^{-3} P$	
Sources	1 mW to 300 W	Product of U and I 1	$30 \cdot 10^{-6} P$	$P$ Calculat ed power
	>300 W to 1 kW	$\text{mV} \leq U \leq 1000 \text{ V}$	$200 \cdot 10^{-6} P$	
	>1 kW to 1 MW	$100 \mu\text{A} \leq I \leq 2000 \text{ A}$	$300 \cdot 10^{-6} P$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Electrical Measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power	5 mW to 50 kW > 2,5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	$w(U_F)$ ... Uncertainty of the amplitude of the voltage fundamental $w(I_F)$ ... Uncertainty of the amplitude of the Current fundamental $w(\Phi_F)$ ... Uncertainty of the phase shift angle
AC reactive power	5 mvar to 50 kvar > 2,5 var to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	$w(U_{rms})$ ... Uncertainty of the voltage rms value $w(I_{rms})$ ... uncertainty of the rms current value
Apparent power	0.1 VA to 50 kVA > 50 VA to 120 kVA	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	
Oscilloscopes				$U$ - measured value
Vertical deflection	5 mV to 5 V 5 mV to 120 V	$R_i = 50 \Omega$ $R_i = 1 M\Omega$	$3.5 \cdot 10^{-3} U$ 35 $\mu$ V $2.4 \cdot 10^{-3} U$ 40 $\mu$ V	Square wave voltage 10 Hz to 10 kHz
Horizontal deflection	5 ns to 520 ms > 20 ms to 5 s		$3 \cdot 10^{-6} T$ 1 ns $30 \cdot 10^{-6} T$ $1.2 \cdot 10^{-4} T^2$	
Rise time	150 ps to 10 ms 250 ps to 10 ms	250 mV > 250 mV to 2.5 V	$35 \cdot 10^{-3} \cdot t_r$ 5 ps $35 \cdot 10^{-3} \cdot t_r$ 8 ps	$t_r$ = intrinsic rise time of the oscilloscope

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Dimensional Measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Length *) Cylindrical setting standards, ring gauges: Diameter *)	1 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d =$ is the measured
Plug gauges: Diameter *)	1 mm to 200 mm	Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	Diameter
Test pins: Diameter *)	0.1 mm to 30 mm	VDI/VDE/DGQ 2618 Sheet 4.2:2007 Point 3.2.2 (Opt. 1)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Thread gauges (one and multi-start cylindrical male and female threads with straight flanks, symmetrical profile) Threaded mandrels: simple Flank diameter	1.4 mm to 200 mm nominal pitch: 0.3 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.8:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Three-wire method $d =$ is the measured diameter
Threaded rings: simpler Flank diameter	3 mm to 200 mm nominal pitch: 0.5 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.9:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Two-ball method $d =$ is the measured Diameter
Threaded mandrels: simple pitch diameter	1.4 mm to 200 mm Nominal diameter	VDI/VDE/DGQ 2618 Sheet 4.8:2006	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Scanning process $d =$ is the measured
Outer diameter		Point 3.2.2 (Opt. 1)	$2 \mu\text{m}$	Diameter
Core diameter / Piercing diameter		to Point 3.2.6 (Opt. 5)	$5 \mu\text{m}$	
Gradient / pitch	0.5 mm to 8 mm		$1.5 \mu\text{m}$	
Thread profile angle $\alpha$	$> 27^\circ$		$(3 \cdot 1 / l_f)'$ , but not smaller than $6'$	$l_f =$ flank length in mm
Threaded rings: simple pitch diameter	5 mm to 200 mm Nominal diameter	VDI/VDE/DGQ 2618 Sheet 4.9:2006	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Scanning process $d =$ is the measured
Outer diameter		Point 3.2.2 (Opt. 1)	$5 \mu\text{m}$	Diameter
Core diameter / Piercing diameter		to Point 3.2.6 (Opt. 5)	$2 \mu\text{m}$	
Gradient / pitch	0.5 mm to 8 mm		$1,5 \mu\text{m}$	
Thread profile angle $\alpha$	$> 27^\circ$		$(3 \cdot 1 / l_f)'$ , but not smaller than $6'$	$l_f =$ flank length in mm

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Dimensional Measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Length <sup>*)</sup> of plane-parallel, spherical or cylindrical measuring surfaces	0,01 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 19.1:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length
Diameter	0,01 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006 Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured diameter
Feeler gauges	0,03 mm to 2,00 mm	DIN 2275:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length
Adjustment dimensions for Outside micrometers	25 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 4.4:2009	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	
Throat gauges	3 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.7:2005 Point 3.3.2 (Opt. 2)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured Diameter
Caliper for Exterior, interior and Depth measurements	0 mm until 500 mm	VDI/VDE/DGQ 2618 Sheet 9.1:2006	$30 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured Length
Depth caliper,	> 500 mm until 1000 mm	VDI/VDE/DGQ 2618 Sheet 9.2:2006	$50 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	
Height caliper		VDI/VDE/DGQ 2618 Sheet 9.3:2006		
Outside micrometers	0 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.1:2001	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Fine pointer measurement- screws	0 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 10.3:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Micrometer head screws	0 mm to 50 mm	VDI/VDE/DGQ 2618 Sheet 10.4:2008	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Depth gauges	0 mm to 300 mm	VDI/VDE/DGQ 2618 Sheet 10.5:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 2-point contact on the calibration object	13 mm to 300 mm > 300 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.7:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$ $5 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 3-line contact on the calibration object	3 mm to 100 mm	VDI/VDE/DGQ 2618 Sheet 10.8:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured diameter
Lever gauges (quick probes) for External measurements	to 200 mm	VDI/VDE/DGQ 2618 Sheet 12.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Essen - Dimensional Measurands

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Lever gauges (quick probe) for indoor measurements	2 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 13.1:2005	$7 \mu\text{m} \cdot 10^{-6} \cdot l$	/ is the measured length
Dial gauges	0 mm to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.1:2021	$3 \mu\text{m} \cdot 10^{-6} \cdot l$	mechanical dial gauges
		VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$3 \mu\text{m} \cdot 10^{-6} \cdot l$	electronic digital dial gauges
Fine pointer	0 mm to 3 mm	VDI/VDE/DGQ 2618 Sheet 11.2:2002	0.6 $\mu\text{m}$	
Feeler lever gauges	0 mm to 1,6 mm	VDI/VDE/DGQ 2618 Sheet 11.3:2002	1.0 $\mu\text{m}$	
electr. inductive Linear Encoders	to 100 mm	VDI/VDE/DGQ 2618 Sheet 14.1:2010	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	
electr. incremental Linear Encoders	to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$0.6 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l$	
Angle Squareness- deviation Flatness and Straightness deviation	to 30 $\mu\text{m}$	VDI/VDE/DGQ/DKD 2618 Sheet 7.1:2019 Point 3.2.2.2 (Opt. 2)	$2.5 \mu\text{m} \cdot 1 \cdot 10^{-6} \cdot l_z$ $4 \mu\text{m} \cdot 5 \cdot 10^{-6} \cdot l_z$	
Protractor Graduation 1° Scale interval 5'	-180° to 180° 0° to 360°	VDI/VDE/DGQ 2618 Sheet 7.2:2008	30' 1'	
Flat rulers Parallelism deviation Flatness deviation	to 500 mm	VDI/VDE/DGQ 2618 Sheet 5.1:2022	$4 \mu\text{m} \cdot 5 \cdot 10^{-6} \cdot l$	/ is the measured Length
			$2.2 \mu\text{m} \cdot 3.5 \cdot 10^{-6} \cdot l$	
Straight edge Straightness deviation	to 500 mm	VDI/VDE/DGQ 2618 Sheet 5.2:2013	$2.2 \mu\text{m} \cdot 3.5 \cdot 10^{-6} \cdot l$	/ is the measured Length
Inclinometers	-2000 $\mu\text{m}/\text{m}$ to 2000 $\mu\text{m}/\text{m}$ (-412") (412")	4_VB_00244_EN V1	1.7 $\mu\text{m}/\text{m}$ (0,35")	Max. Leg length of the KG: 500 mm

**Location Calibration Laboratory Hamburg, Meiendorfer Straße 205, 22145 Hamburg, Germany**

**Permanent Laboratory Hamburg - Electrical Measurands**

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC voltage Measuring devices	0 V 1 mV to 2,2 V > 2,2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		0.1 $\mu$ V $7 \cdot 10^{-6} U$ 1 $\mu$ V $9 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U$ = set value
DC voltage Sources	0 V 1 mV to 100 mV > 100 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		0.1 $\mu$ V $8 \cdot 10^{-6} U$ 1 $\mu$ V $11 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$	$U$ = measured value
High Voltage	> 1 kV to 10 kV		$2.5 \cdot 10^{-3} U$ 2.5 V	$U$ = measured value
DC current strength meters and sources	0 A 0.1 $\mu$ A until < 1 $\mu$ A 1 $\mu$ A until < 10 $\mu$ A 10 $\mu$ A until < 100 $\mu$ A 100 $\mu$ A until 320 mA > 320 mA until 1 A > 1 A to 10 A > 10 A > 150 A 150 A to 2000 A	Precision Open  Normal resistance and voltmeter      Current transformer	0.2 nA  $35 - 10^{-6} / 21$ pA $50 - 10^{-6} / 6$ pA $15 - 10^{-6} / 0.4$ nA $18 - 10^{-6} /$ $15 - 10^{-6} / 6$ $\mu$ A $0,2 - 10^{-3} /$ $0,3 - 10^{-3} /$ $0,3 - 10^{-3} /$	$I$ = set value
Direct current strength Current clamps	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 1000 A		$1 \cdot 10^{-3} /$ $2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Hamburg - Electrical Measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks		
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement			
DC resistance	0 Ω		50 μΩ	R = set value Fluke 5700A		
	1 Ω; 1.9 Ω		95 · 10 <sup>-6</sup> R			
	10 Ω		28 · 10 <sup>-6</sup> R			
	19 Ω		27 · 10 <sup>-6</sup> R			
	100 Ω; 190 Ω		17 · 10 <sup>-6</sup> R			
	1 kΩ		13 · 10 <sup>-6</sup> R			
	1.9 kΩ		13 · 10 <sup>-6</sup> R			
	10 kΩ		12 · 10 <sup>-6</sup> R			
	19 kΩ		12 · 10 <sup>-6</sup> R			
	100 kΩ		14 · 10 <sup>-6</sup> R			
	190 kΩ		14 · 10 <sup>-6</sup> R			
	1 MΩ		20 · 10 <sup>-6</sup> R			
	1.9 MΩ		21 · 10 <sup>-6</sup> R			
	10 MΩ		40 · 10 <sup>-6</sup> R			
	19 MΩ		48 · 10 <sup>-6</sup> R			
	100 MΩ		110 · 10 <sup>-6</sup> R			
	DC resistance	0 Ω			100 μΩ	R = measured value HP 3458A
		1 Ω to 10 Ω			16 · 10 <sup>-6</sup> R 50 μΩ	
		> 10 Ω to 100 Ω			12 · 10 <sup>-6</sup> R 500 μΩ	
		> 100 Ω to 1 kΩ			15 · 10 <sup>-6</sup> R	
> 1 kΩ to 10 kΩ			15 · 10 <sup>-6</sup> R			
> 10 kΩ to 100 kΩ			15 · 10 <sup>-6</sup> R			
> 100 kΩ to 1 MΩ			35 · 10 <sup>-6</sup> R			
> 1 MΩ to 10 MΩ			150 · 10 <sup>-6</sup> R			
> 10 MΩ to 100 MΩ			600 · 10 <sup>-6</sup> R			
> 100 MΩ to 1 GΩ			5 · 10 <sup>-3</sup> R			
DC resistance	0,001 Ω to 0,1 Ω	Substitution procedure with normal resistance	50 · 10 <sup>-6</sup> R			
	> 0,1 Ω to 1 MΩ		20 · 10 <sup>-6</sup> R			
	> 1 MΩ to 100 MΩ		30 · 10 <sup>-6</sup> R			
DC resistance	1 mΩ		0.1 · 10 <sup>-3</sup> · R			
	10 mΩ		30 · 10 <sup>-6</sup> · R			
	100 mΩ		30 · 10 <sup>-6</sup> · R			
	1 Ω		30 · 10 <sup>-6</sup> · R			
DC resistance	1 mΩ until 10 mΩ	Substitution procedure	0.1 · 10 <sup>-3</sup> · R			
	> 10 mΩ until 1 Ω	Substitution procedure	30 · 10 <sup>-6</sup> · R			
	0,1 Ω to 2 Ω	Direct procedure	50 · 10 <sup>-6</sup> · R			



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Hamburg - Electrical Measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC resistance (areas) Measuring devices	1 Ω to < 11 Ω		$0.12 \cdot 10^{-3} R$	<i>R</i> = set value Fluke 5520A /5522A
	11 Ω to < 33 Ω		$33 \cdot 10^{-6} R$	
	33 Ω to < 110 Ω		$29 \cdot 10^{-6} R$	
	110 Ω to < 330 Ω		$28 \cdot 10^{-6} R$	
	330 Ω to < 1.1 kΩ		$28 \cdot 10^{-6} R$	
	1.1 kΩ to < 3.3 kΩ		$28 \cdot 10^{-6} R$	
	3.3 kΩ to < 11 kΩ		$28 \cdot 10^{-6} R$	
	11 kΩ to < 33 kΩ		$28 \cdot 10^{-6} R$	
	33 kΩ to < 110 kΩ		$28 \cdot 10^{-6} R$	
	110 kΩ to < 330 kΩ		$32 \cdot 10^{-6} R$	
	330 kΩ to < 1.1 MΩ		$33 \cdot 10^{-6} R$	
	1.1 MΩ to < 3.3 MΩ		$62 \cdot 10^{-6} R$	
	3.3 MΩ to < 11 MΩ		$0.13 \cdot 10^{-3} R$	
	11 MΩ to < 33 MΩ		$0.25 \cdot 10^{-3} R$	
	33 MΩ to < 110 MΩ		$0.5 \cdot 10^{-3} R$	
110 MΩ to < 330 MΩ		$3 \cdot 10^{-3} R$		
330 MΩ to < 1.1 GΩ		$15 \cdot 10^{-3} R$		
AC resistance	0,1 Ω to 2 Ω	50 Hz to 400 Hz	$10 \cdot 10^{-3} \cdot R$	
	> 2 Ω to 5 Ω	50 Hz	$9 \cdot 10^{-3} \cdot R$	
	> 5 Ω to 20 Ω	50 Hz	$5 \cdot 10^{-3} \cdot R$	
	> 20 Ω to 200 Ω	50 Hz	$5 \cdot 10^{-3} \cdot R$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Hamburg - Electrical Measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	1 mV to 2.2 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$0.52 \cdot 10^{-3} U$ $0.52 \cdot 10^{-3} U$ $0.40 \cdot 10^{-3} U$ $0.40 \cdot 10^{-3} U$ $0.41 \cdot 10^{-3} U$ $0.46 \cdot 10^{-3} U$ $0.55 \cdot 10^{-3} U$ $0.60 \cdot 10^{-3} U$	<i>U</i> = measured value
	> 2.2 mV to 7 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$0.22 \cdot 10^{-3} U$ $0.22 \cdot 10^{-3} U$ $0.16 \cdot 10^{-3} U$ $0.16 \cdot 10^{-3} U$ $0.20 \cdot 10^{-3} U$ $0.22 \cdot 10^{-3} U$ $0.33 \cdot 10^{-3} U$ $0.45 \cdot 10^{-3} U$	
	> 7 mV to 22 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$80 \cdot 10^{-6} U$ $80 \cdot 10^{-6} U$ $65 \cdot 10^{-6} U$ $75 \cdot 10^{-6} U$ $75 \cdot 10^{-6} U$ $95 \cdot 10^{-6} U$ $0.19 \cdot 10^{-3} U$ $0.21 \cdot 10^{-3} U$	
	> 22 mV to 70 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$70 \cdot 10^{-6} U$ $58 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $55 \cdot 10^{-6} U$ $0,11 \cdot 10^{-3} U$ $0,13 \cdot 10^{-3} U$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Hamburg - Electrical Measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 70 mV to 220 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$39 \cdot 10^{-6} U$ $35 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $28 \cdot 10^{-6} U$ $42 \cdot 10^{-6} U$ $85 \cdot 10^{-6} U$ $0,1 \cdot 10^{-3} U$	<i>U</i> = measured value
	> 220 mV to 700 mV	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$25 \cdot 10^{-6} U$ $22 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $27 \cdot 10^{-6} U$ $40 \cdot 10^{-6} U$	
	> 700 mV to 2,2 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$20 \cdot 10^{-6} U$ $14 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $10 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $22 \cdot 10^{-6} U$ $68 \cdot 10^{-6} U$	
	> 2,2 V to 7 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$18 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $95 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**
**Permanent Laboratory Hamburg - Electrical Measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 7 V to 22 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$17 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $11 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $30 \cdot 10^{-6} U$ $0,11 \cdot 10^{-3} U$	<i>U</i> = measured value
	> 22 V to 70 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz > 100 kHz to 300 kHz > 300 kHz to 500 kHz > 500 kHz to 1 MHz	$18 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $15 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $25 \cdot 10^{-6} U$ $40 \cdot 10^{-6} U$ $0,13 \cdot 10^{-3} U$	
	> 70 V to 220 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	$19 \cdot 10^{-6} U$ $18 \cdot 10^{-6} U$ $17 \cdot 10^{-6} U$ $17 \cdot 10^{-6} U$ $32 \cdot 10^{-6} U$	
	> 220 V to 1000 V	10 Hz to 20 Hz > 20 Hz to 40 Hz > 40 Hz to 20 kHz > 20 kHz to 50 kHz > 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$ $27 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $45 \cdot 10^{-6} U$ $65 \cdot 10^{-6} U$	
High voltage	> 0,7 kV to 1 kV > 1 kV to 7 kV	50 Hz	$2,5 \cdot 10^{-3} U + 0,25 V$ $3,5 \cdot 10^{-3} U + 2,0 V$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Hamburg - Electrical Measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	100 µA to1 mA	10 Hz to40 Hz  > 40 Hz to1 kHz; > 1 kHz to10 kHz;	120 · 10 <sup>-6</sup> /  160 · 10 <sup>-6</sup> / 60 · 10 <sup>-6</sup> /	/ = measured value
	> 1 mA to10 mA	10 Hz to40 Hz  > 40 Hz to1 kHz; > 1 kHz to10 kHz;	46 · 10 <sup>-6</sup> /	
Current clamps	> 10 mA to1 A	10 Hz to40 Hz  > 40 Hz to1 kHz; > 1 kHz to10 kHz;	17 · 10 <sup>-6</sup> /	/ = measured value rt
	> 1 A to10 A	10 Hz to40 Hz  > 40 Hz to1 kHz; > 1 kHz to10 kHz;	32 · 10 <sup>-6</sup> /	
	> 10 A to20 A	10 Hz to40 Hz  > 40 Hz to1 kHz; > 1 kHz to10 kHz;	39 · 10 <sup>-6</sup> /	
	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 800 A	40 Hz to 5 kHz 40 Hz to 5 kHz 40 Hz to 65 Hz	2 · 10 <sup>-3</sup> / 3 · 10 <sup>-3</sup> / 4 · 10 <sup>-3</sup> /	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Hamburg - Electrical Measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Capacity	190 pF to < 400 pF	10 Hz to 10 kHz	$4 \cdot 10^{-3} C + 8 \text{ pF}$	With 5520A / 5522A
Measuring devices	400 pF to < 1,1 nF	10 Hz to 10 kHz	$4,5 \cdot 10^{-3} C + 8 \text{ pF}$	
	1,1 nF to < 3,3 nF	10 Hz to 3 kHz	$4,0 \cdot 10^{-3} C + 8 \text{ pF}$	
	3,3 nF to < 11 nF	10 Hz to 1 kHz	$2,5 \cdot 10^{-3} C + 8 \text{ pF}$	
	11 nF to < 33 nF	10 Hz to 1 kHz	$2,5 \cdot 10^{-3} C + 80 \text{ pF}$	
	33 nF to < 110 nF	10 Hz to 1 kHz	$2,5 \cdot 10^{-3} C + 80 \text{ pF}$	
	110 nF to < 330 nF	10 Hz to 1 kHz	$4,5 \cdot 10^{-3} C$	
	330 nF to < 1,1 μF	10 Hz to 600 Hz	$4,5 \cdot 10^{-3} C$	
	1,1 μF to < 3,3 μF	10 Hz to 300 Hz	$4,5 \cdot 10^{-3} C$	
	3,3 μF to < 11 μF	10 Hz to 150 Hz	$4,5 \cdot 10^{-3} C$	
	11 μF to < 33 μF	10 Hz to 120 Hz	$6,0 \cdot 10^{-3} C$	
	33 μF to < 110 μF	10 Hz to 80 Hz	$6,5 \cdot 10^{-3} C$	
	110 μF to < 330 μF	DC to 50 Hz	$6,0 \cdot 10^{-3} C$	
	330 μF to < 1,1 mF	DC to 20 Hz	$6,0 \cdot 10^{-3} C$	
	1,1 mF to < 3,3 mF	DC to 6 Hz	$6,0 \cdot 10^{-3} C$	
	3,3 mF to < 11 mF	DC to 2 Hz	$6,0 \cdot 10^{-3} C$	
	11 mF to < 33 mF	DC to 0.6 Hz	$8,0 \cdot 10^{-3} C$	
	33 mF to 110 mF	DC to 0.2 Hz	$11 \cdot 10^{-3} C$	
	1 nF to 100 nF	50 Hz until 10 kHz	$1,0 \cdot 10^{-3} C$	C: measured value with normal capacitances
	> 100 nF to 1000 nF	50 Hz until 1 kHz	$1,0 \cdot 10^{-3} C$	
		> 1 kHz to 10 kHz	$2,5 \cdot 10^{-3} C$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Hamburg - Electrical Measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Frequency	1 mHz to 1 GHz		$2 \cdot 10^{-9} \cdot f \cdot U_{Tf}$	$f$ = current measured value $U_{Tf}$ = Trigger uncertainty
Time interval	1 $\mu$ s to 1000 s		$2 \cdot 10^{-9} \cdot t \cdot 2$ ns	$t$ = current measured value
Speed optical	1 min <sup>-1</sup> up to 100.000 min <sup>-1</sup>	with light pulse generator	$8 \cdot 10^{-6}$ but not less than 0.006 min <sup>-1</sup>	
AC active power measuring devices	109 $\mu$ W to < 11kW	33 mV to 1000 V 45 Hz to 65 kHz PF = 1 33 mA to < 11 A	$1.4 \cdot 10^{-3} P$	$P$ = set value with Fluke 5520A/5522A PF: Power factor
	363 mW up to 20 kW	11 A to 20 A	$2.0 \cdot 10^{-3} P$	
DC power Measuring devices	1 mW to 300 W		$0.5 \cdot 10^{-3} P$	
	> 300 W to 20 kW		$1.0 \cdot 10^{-3} P$	
Sources	1 mW to 300 W	Product of U and I 1 mV	$30 \cdot 10^{-6} P$	$P$ Calculated power
	> 300 W to 1 kW	$\leq U \leq 1000$ V	$200 \cdot 10^{-6} P$	
	>1 kW up to 1 MW	$100 \mu$ A $\leq I \leq 2000$ A	$300 \cdot 10^{-6} P$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Hamburg - Electrical Measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power	5 mW up to 50 kW > 2.5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	$w(U_F)$ . .Uncertainty of the amplitude of the voltage fundamental $w(I_F)$ . .Uncertainty of the amplitude of the Current fundamental $w(\Phi_F)$ ...Uncertainty of the phase shift angle $w(U_{rms})$ ... Uncertainty of the voltage rms value $w(I_{rms})$ ... uncertainty of the rms current value
AC reactive power	5 mvar to 50 kvar > 2.5 var up to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	
Apparent power	0.1 VA up to 50 kVA > 50 VA to 120 kVA	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	
Oscilloscopes				$U$ - measured value
Vertical deflection	5 mV to 5 V 5 mV to 120 V	$R_i = 50 \Omega$ $R_i = 1 M\Omega$	$3.5 \cdot 10^{-3} U$ 35 $\mu$ V $2.4 \cdot 10^{-3} U$ 40 $\mu$ V	Square wave voltage 10 Hz to 10 kHz
Horizontal deflection	5 ns to 520 ms > 20 ms to 5 s		$3 \cdot 10^{-6} T$ 1 ns $30 \cdot 10^{-6} T$ $1.2 \cdot 10^{-4} T^2$	
Rise time	150 ps to 10 ms 250 ps to 10 ms	250 mV > 250 mV to 2.5 V	$35 \cdot 10^{-3} \tau$ 5 ps $35 \cdot 10^{-3} \tau$ 8 ps	$\tau$ = intrinsic rise time of the oscilloscope



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Location Calibration laboratory Mörfelden-Walldorf, Kurhessenstraße 11, 64546 Mörfelden-Walldorf**

**Permanent Laboratory Mörfelden-Walldorf - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC voltage Measuring devices	0 V 1 mV to 2,2 V > 2,2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		0.1 $\mu$ V $7 \cdot 10^{-6} U$ 1 $\mu$ V $9 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U$ = set value
DC voltage Sources	0 V 1 mV to 100 mV > 100 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		0.1 $\mu$ V $8 \cdot 10^{-6} U$ 1 $\mu$ V $11 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$	$U$ = measured value
High Voltage	> 1 kV to 10 kV		$2.5 \cdot 10^{-3} U$ 2.5 V	$U$ = measured value
DC current strength meters and sources	0 A 0.1 $\mu$ A until < 1 $\mu$ A 1 $\mu$ A until < 10 $\mu$ A 10 $\mu$ A until < 100 $\mu$ A 100 $\mu$ A until 320 mA > 320 mA until 1 A > 1 A to 10 A > 10 A > 150 A 150 A to 2000 A	Precision Open Normal resistance and voltmeter Current transformer	0.2 nA $35 - 10^{-6} / 21$ pA $50 - 10^{-6} / 6$ pA $15 - 10^{-6} / 0.4$ nA 18 - $10^{-6} /$ $15 - 10^{-6} / 6$ $\mu$ A 0,2 - $10^{-3} /$ 0,3 - $10^{-3} /$ 0,3 - $10^{-3} /$	$I$ = set value
Direct current strength Current clamps	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 1000 A		$1 \cdot 10^{-3} /$ $2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Mörfelden-Walldorf - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC resistance	0 Ω		50 μΩ	R = set value
Measuring devices	1 Ω; 1.9 Ω		95 · 10 <sup>-6</sup> R	Fluke 5700A
	10 Ω		28 · 10 <sup>-6</sup> R	
	19 Ω		27 · 10 <sup>-6</sup> R	
	100 Ω; 190 Ω		17 · 10 <sup>-6</sup> R	
	1 kΩ		13 · 10 <sup>-6</sup> R	
	1.9 kΩ		13 · 10 <sup>-6</sup> R	
	10 kΩ		12 · 10 <sup>-6</sup> R	
	19 kΩ		12 · 10 <sup>-6</sup> R	
	100 kΩ		14 · 10 <sup>-6</sup> R	
	190 kΩ		14 · 10 <sup>-6</sup> R	
	1 MΩ		20 · 10 <sup>-6</sup> R	
	1.9 MΩ		21 · 10 <sup>-6</sup> R	
	10 MΩ		40 · 10 <sup>-6</sup> R	
	19 MΩ		48 · 10 <sup>-6</sup> R	
	100 MΩ		110 · 10 <sup>-6</sup> R	
Sources	0 Ω		100 μΩ	R = measured value
	1 Ω to 10 Ω		16 · 10 <sup>-6</sup> R 50 μΩ	HP 3458A
	> 10 Ω to 100 Ω		12 · 10 <sup>-6</sup> R 500 μΩ	
	> 100 Ω to 1 kΩ		15 · 10 <sup>-6</sup> R	
	> 1 kΩ to 10 kΩ		15 · 10 <sup>-6</sup> R	
	> 10 kΩ to 100 kΩ		15 · 10 <sup>-6</sup> R	
	> 100 kΩ to 1 MΩ		35 · 10 <sup>-6</sup> R	
	> 1 MΩ to 10 MΩ		150 · 10 <sup>-6</sup> R	
	> 10 MΩ to 100 MΩ		600 · 10 <sup>-6</sup> R	
	> 100 MΩ to 1 GΩ		5 · 10 <sup>-3</sup> R	
	0,001 Ω to 0,1 Ω	Substitution procedure	50 · 10 <sup>-6</sup> R	
	> 0,1 Ω to 1 MΩ	with normal resistance	20 · 10 <sup>-6</sup> R	
	> 1 MΩ to 100 MΩ		30 · 10 <sup>-6</sup> R	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Mörfelden-Walldorf - Electrical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
DC resistance Measuring devices	1 Ω to < 11 Ω		$0.12 \cdot 10^{-3} R$	<i>R</i> = set value Fluke 5520A /5522A
	11 Ω to < 33 Ω		$33 \cdot 10^{-6} R$	
	33 Ω to < 110 Ω		$29 \cdot 10^{-6} R$	
	110 Ω to < 330 Ω		$28 \cdot 10^{-6} R$	
	330 Ω to < 1.1 kΩ		$28 \cdot 10^{-6} R$	
	1.1 kΩ to < 3.3 kΩ		$28 \cdot 10^{-6} R$	
	3.3 kΩ to < 11 kΩ		$28 \cdot 10^{-6} R$	
	11 kΩ to < 33 kΩ		$28 \cdot 10^{-6} R$	
	33 kΩ to < 110 kΩ		$28 \cdot 10^{-6} R$	
	110 kΩ to < 330 kΩ		$32 \cdot 10^{-6} R$	
	330 kΩ to < 1.1 MΩ		$33 \cdot 10^{-6} R$	
	1.1 MΩ to < 3.3 MΩ		$62 \cdot 10^{-6} R$	
	3.3 MΩ to < 11 MΩ		$0.13 \cdot 10^{-3} R$	
	11 MΩ to < 33 MΩ		$0.25 \cdot 10^{-3} R$	
	33 MΩ to < 110 MΩ		$0.5 \cdot 10^{-3} R$	
110 MΩ to < 330 MΩ		$3 \cdot 10^{-3} R$		
330 MΩ to < 1.1 GΩ		$15 \cdot 10^{-3} R$		
AC resistance	0,1 Ω to 2 Ω	50 Hz to 400 Hz	$10 \cdot 10^{-3} R$	
AC voltage Measuring devices and Sources	1 mV to 2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV to 7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
		> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$	
> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$			
> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Mörfelden-Walldorf - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 7 mV to 22 mV	10 Hz to 20 Hz	$80 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$80 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$65 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$75 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$75 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$95 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$0.19 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.21 \cdot 10^{-3} U$	
	> 22 mV to 70 mV	10 Hz to 20 Hz	$70 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$58 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$35 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$35 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$45 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$55 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$0.11 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
	> 70 mV to 220 mV	10 Hz to 20 Hz	$39 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$35 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$25 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$25 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$28 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$42 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$85 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.1 \cdot 10^{-3} U$	
	> 220 mV to 700 mV	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$22 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$12 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$12 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$14 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$27 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$40 \cdot 10^{-6} U$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Mörfelden-Walldorf - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage meters and sources	> 700 mV to 2.2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$11 \cdot 10^{-6} U$	
> 2.2 V to 7 V	> 700 mV to 2.2 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$13 \cdot 10^{-6} U$	
> 7 V to 22 V	> 2.2 V to 7 V	10 Hz to 20 Hz	$17 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
> 22 V to 70 V	> 7 V to 22 V	> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.11 \cdot 10^{-3} U$	
		10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$	
> 70 V to 220 V	> 22 V to 70 V	> 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
		10 Hz to 20 Hz	$19 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$18 \cdot 10^{-6} U$	
> 70 V to 220 V	> 70 V to 220 V	> 40 Hz to 20 kHz	$17 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$17 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$32 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$32 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Mörfelden-Walldorf - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC voltage meters and sources	> 220 V to 1000 V	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$27 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$45 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$45 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$65 \cdot 10^{-6} U$	
High Voltage	> 0.7 kV to 1 kV	50 Hz	$2.5 \cdot 10^{-3} U 0.25 V$	
	> 1 kV to 7 kV		$3.5 \cdot 10^{-3} U 2.0 V$	
Alternating current strength sources and measuring devices	100 $\mu$ A to 1 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$120 \cdot 10^{-6} I$ $160 \cdot 10^{-6} I$ $60 \cdot 10^{-6} I$	<i>I</i> = measured value
	> 1 mA to 10 mA	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$46 \cdot 10^{-6} I$	
Current clamps	> 10 mA to 1 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$17 \cdot 10^{-6} I$	<i>I</i> = measured value
	> 1 A to 10 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$32 \cdot 10^{-6} I$	
	> 10 A to 20 A	10 Hz to 40 Hz > 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$39 \cdot 10^{-6} I$	
	1 mA to 2.2 A > 2.2 A to 20 A > 20 A to 800 A	40 Hz to 5 kHz 40 Hz to 5 kHz 40 Hz to 65 Hz	$2 \cdot 10^{-3} I$ $3 \cdot 10^{-3} I$ $4 \cdot 10^{-3} I$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Mörfelden-Walldorf - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Capacity gauges	190 pF to < 400 pF	10 Hz to 10 kHz	$4 \cdot 10^{-3} C$ 8 pF	With 5520A / 5522A
	400 pF to < 1.1 nF	10 Hz to 10 kHz	$4.5 \cdot 10^{-3} C$ 8 pF	
	1.1 nF to < 3.3 nF	10 Hz to kHz	$4.0 \cdot 10^{-3} C$ 8 pF	
	3.3 nF to < 11 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 8 pF	
	11 nF to < 33 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	33 nF to < 110 nF	10 Hz to 1 kHz	$2.5 \cdot 10^{-3} C$ 80 pF	
	110 nF to < 330 nF	10 Hz to 1 kHz	$4.5 \cdot 10^{-3} C$	
	330 nF to < 1.1 µF	10 Hz to 600 Hz	$4.5 \cdot 10^{-3} C$	
	1.1 µF to < 3.3 µF	10 Hz to 300 Hz	$4.5 \cdot 10^{-3} C$	
	3.3 µF to < 11 µF	10 Hz to 150 Hz	$4.5 \cdot 10^{-3} C$	
	11 µF to < 33 µF	10 Hz to 120 Hz	$6.0 \cdot 10^{-3} C$	
	33 µF to < 110 µF	10 Hz to 80 Hz DC to	$6.5 \cdot 10^{-3} C$	
	110 µF to < 330 µF	50 Hz DC to Hz DC	$6.0 \cdot 10^{-3} C$	
	330 µF to < 1.1 mF	to 6 Hz DC to 2 Hz	$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	DC to 200.6 Hz	$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 0,2 Hz	$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF		$8.0 \cdot 10^{-3} C$	
	33 mF to 110 mF		$11 \cdot 10^{-3} C$	
		1 nF to 100 nF > 100 nF to 1000 nF	50 Hz to 10 kHz 50 Hz to 1 kHz > 1 kHz to 10 kHz	
Frequency	1 mHz to 1 GHz		$2 \cdot 10^{-9} \cdot f_{UTP}$	$f$ = current measured value $UTP$ = Trigger uncertainty
Time interval	1 µs to 1000 s		$2 \cdot 10^{-9} \cdot t$ 2 ns	$t$ = current measured value
Speed optical	1 min <sup>-1</sup> up to 100.000 min <sup>-1</sup>	with light pulse generator	$8 \cdot 10^{-6}$ but not less than 0.006 min <sup>-1</sup>	
AC active power measuring devices		33 mV to 1000 V 45 Hz to 65 kHz $PF = 1$		$P$ = set value with Fluke 5520A/5522A $PF$ : Power factor
	109 µW to < 11kW	33 mA to < 11 A	$1.4 \cdot 10^{-3} P$	
	363 mW up to 20 kW	11 A to 20 A	$2.0 \cdot 10^{-3} P$	
DC power Measuring devices	1 mW to 300 W		$0.5 \cdot 10^{-3} P$	
	> 300 W to 20 kW		$1.0 \cdot 10^{-3} P$	
Sources	1 mW to 300 W	Product of U and I 1 mV	$30 - 10^{-6} P$	$P$ Calculated power
	> 300 W to 1 kW	$\leq U \leq 1000 V$	$200 - 10^{-6} P$	
	> 1 kW to 1 MW	$100 \mu A \leq I \leq 2000 A$	$300 - 10^{-6} P$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Mörfelden-Walldorf - Electrical measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power	5 mW up to 50 kW  > 2.5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz 0.05 ≤ cos φ ≤ 1 0.1 A to 50 A  > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not smaller than 80 - 10 <sup>-6</sup> not smaller than 0,14 - 10 <sup>-3</sup>	w <sub>(UF)</sub> ... Uncertainty of the amplitude of the voltage fundamental w <sub>(IF)</sub> ... Uncertainty of the amplitude  of the Current fundamental w <sub>(ΦF)</sub> ... Uncertainty of the phase shift angle w <sub>(U<sub>rms</sub>)</sub> ... Uncertainty of the voltage rms value w <sub>(I<sub>rms</sub>)</sub> ... uncertainty of the rms current value
AC reactive power	5 mvar to 50 kvar  > 2.5 var up to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz 0.05 ≤ cos φ ≤ 1 0.1 A to 50 A  > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not smaller than 80 - 10 <sup>-6</sup> not smaller than 0,14 - 10 <sup>-3</sup>	
Apparent power	0.1 VA up to 50 kVA  > 50 VA to 120 kVA	1 V to 1000 V 45 Hz to 65 Hz 0.05 ≤ cos φ ≤ 1 0.1 A to 50 A  > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than 80 - 10 <sup>-6</sup> not smaller than 0.14 - 10 <sup>-3</sup>	
Oscilloscopes				U - measured value
Vertical deflection	5 mV to 5 V  5 mV to 120 V	R <sub>i</sub> = 50 Ω  R <sub>i</sub> = 1 MΩ	3.5 · 10 <sup>-3</sup> U 35 μV  2.4 · 10 <sup>-3</sup> U 40 μV	Square wave voltage 10 Hz to 10 kHz
Horizontal deflection	5 ns to 520 ms  > 20 ms to 5 s		3 · 10 <sup>-6</sup> T 1 ns  30 · 10 <sup>-6</sup> T 1.2 · 10 <sup>-4</sup> T <sup>2</sup>	
Rise time	150 ps to 10 ms  250 ps to 10 ms	250 mV  > 250 mV to 2.5 V	35 · 10 <sup>-3</sup> -tr 5 ps  35 · 10 <sup>-3</sup> -tr 8 ps	t <sub>r</sub> = intrinsic rise time of the oscilloscope



**Location Calibration Laboratory Winsen, Tönnhäuser Weg 100-106, 21423 Winsen (Luhe)**

**Permanent Laboratory Winsen - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC voltage Measuring devices	0 V 1 mV to 2,2 V > 2,2 V to 11 V > 11 V to 22 V > 22 V to 220 V > 220 V to 1000 V		0.1 $\mu$ V $7 \cdot 10^{-6} U$ 1 $\mu$ V $9 \cdot 10^{-6} U$ $8 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$ $12 \cdot 10^{-6} U$	$U$ = set value
DC voltage Sources	0 V 1 mV to 100 mV > 100 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		0.1 $\mu$ V $8 \cdot 10^{-6} U$ 1 $\mu$ V $11 \cdot 10^{-6} U$ $9 \cdot 10^{-6} U$ $13 \cdot 10^{-6} U$ $16 \cdot 10^{-6} U$	$U$ = measured value
High Voltage	> 1 kV to 10 kV		$2.5 \cdot 10^{-3} U$ 2.5 V	$U$ = measured value
DC current strength meters and sources	0 A 0.1 $\mu$ A until < 1 $\mu$ A 1 $\mu$ A until < 10 $\mu$ A 10 $\mu$ A until < 100 $\mu$ A 100 $\mu$ A until 320 mA > 320 mA until 1 A > 1 A to 10 A > 10 A > 150 A 150 A to 2000 A	Precision Open Normal resistance and voltmeter Current transformer	0.2 nA $35 \cdot 10^{-6} / 21$ pA $50 \cdot 10^{-6} / 6$ pA $15 \cdot 10^{-6} / 0.4$ nA $18 \cdot 10^{-6} /$ $15 \cdot 10^{-6} / 6$ $\mu$ A $0,2 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$ $0,3 \cdot 10^{-3} /$	$I$ = set value
Direct current strength Current clamps	1 mA to 2,2 A > 2,2 A to 20 A > 20 A to 1000 A		$1 \cdot 10^{-3} /$ $2 \cdot 10^{-3} /$ $3 \cdot 10^{-3} /$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Winsen - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC resistance	0 Ω		50 μΩ	R = set value
Measuring devices	1 Ω; 1.9 Ω		95 · 10 <sup>-6</sup> R	Fluke 5700A
	10 Ω		28 · 10 <sup>-6</sup> R	
	19 Ω		27 · 10 <sup>-6</sup> R	
	100 Ω; 190 Ω		17 · 10 <sup>-6</sup> R	
	1 kΩ		13 · 10 <sup>-6</sup> R	
	1.9 kΩ		13 · 10 <sup>-6</sup> R	
	10 kΩ		12 · 10 <sup>-6</sup> R	
	19 kΩ		12 · 10 <sup>-6</sup> R	
	100 kΩ		14 · 10 <sup>-6</sup> R	
	190 kΩ		14 · 10 <sup>-6</sup> R	
	1 MΩ		20 · 10 <sup>-6</sup> R	
	1.9 MΩ		21 · 10 <sup>-6</sup> R	
	10 MΩ		40 · 10 <sup>-6</sup> R	
	19 MΩ		48 · 10 <sup>-6</sup> R	
	100 MΩ		110 · 10 <sup>-6</sup> R	
Sources	0 Ω		100 μΩ	R = measured value
	1 Ω to 10 Ω		16 · 10 <sup>-6</sup> R 50 μΩ	HP 3458A
	> 10 Ω to 100 Ω		12 · 10 <sup>-6</sup> R 500 μΩ	
	> 100 Ω to 1 kΩ		15 · 10 <sup>-6</sup> R	
	> 1 kΩ to 10 kΩ		15 · 10 <sup>-6</sup> R	
	> 10 kΩ to 100 kΩ		15 · 10 <sup>-6</sup> R	
	> 100 kΩ to 1 MΩ		35 · 10 <sup>-6</sup> R	
	> 1 MΩ to 10 MΩ		150 · 10 <sup>-6</sup> R	
	> 10 MΩ to 100 MΩ		600 · 10 <sup>-6</sup> R	
	> 100 MΩ to 1 GΩ		5 · 10 <sup>-3</sup> R	
	0,001 Ω to 0,1 Ω	Substitution procedure	50 · 10 <sup>-6</sup> R	
	> 0,1 Ω to 1 MΩ	with normal resistance	20 · 10 <sup>-6</sup> R	
	> 1 MΩ to 100 MΩ		30 · 10 <sup>-6</sup> R	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Winsen - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
DC resistance	1 Ω to < 11 Ω		$0.12 \cdot 10^{-3} R$	<i>R</i> = set value Fluke 5520A /5522A
Measuring devices	11 Ω to < 33 Ω		$33 \cdot 10^{-6} R$	
	33 Ω to < 110 Ω		$29 \cdot 10^{-6} R$	
	110 Ω to < 330 Ω		$28 \cdot 10^{-6} R$	
	330 Ω to < 1.1 kΩ		$28 \cdot 10^{-6} R$	
	1.1 kΩ to < 3.3 kΩ		$28 \cdot 10^{-6} R$	
	3.3 kΩ to < 11 kΩ		$28 \cdot 10^{-6} R$	
	11 kΩ to < 33 kΩ		$28 \cdot 10^{-6} R$	
	33 kΩ to < 110 kΩ		$28 \cdot 10^{-6} R$	
	110 kΩ to < 330 kΩ		$32 \cdot 10^{-6} R$	
	330 kΩ to < 1.1 MΩ		$33 \cdot 10^{-6} R$	
	1.1 MΩ to < 3.3 MΩ		$62 \cdot 10^{-6} R$	
	3.3 MΩ to < 11 MΩ		$0.13 \cdot 10^{-3} R$	
	11 MΩ to < 33 MΩ		$0.25 \cdot 10^{-3} R$	
	33 MΩ to < 110 MΩ		$0.5 \cdot 10^{-3} R$	
	110 MΩ to < 330 MΩ		$3 \cdot 10^{-3} R$	
	330 MΩ to < 1.1 GΩ		$15 \cdot 10^{-3} R$	
AC resistance	0,1 Ω to 2 Ω	50 Hz to 400 Hz	$10 \cdot 10^{-3} R$	
AC voltage	1 mV to 2.2 mV	10 Hz to 20 Hz	$0.52 \cdot 10^{-3} U$	<i>U</i> = measured value
Measuring devices and Sources		> 20 Hz to 40 Hz	$0.52 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.40 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.40 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.41 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.46 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.55 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.60 \cdot 10^{-3} U$	
	> 2.2 mV to 7 mV	10 Hz to 20 Hz	$0.22 \cdot 10^{-3} U$	
		> 20 Hz to 40 Hz	$0.22 \cdot 10^{-3} U$	
		> 40 Hz to 20 kHz	$0.16 \cdot 10^{-3} U$	
		> 20 kHz to 50 kHz	$0.16 \cdot 10^{-3} U$	
		> 50 kHz to 100 kHz	$0.20 \cdot 10^{-3} U$	
		> 100 kHz to 300 kHz	$0.22 \cdot 10^{-3} U$	
		> 300 kHz to 500 kHz	$0.33 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.45 \cdot 10^{-3} U$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Winsen - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and Sources	> 7 mV to 22 mV	10 Hz to 20 Hz	$80 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$80 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$65 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$75 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$75 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$95 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$0.19 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.21 \cdot 10^{-3} U$	
	> 22 mV to 70 mV	10 Hz to 20 Hz	$70 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$58 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$35 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$35 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$45 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$55 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$0.11 \cdot 10^{-3} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
	> 70 mV to 220 mV	10 Hz to 20 Hz	$39 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$35 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$25 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$25 \cdot 10^{-6} U$	
> 50 kHz to 100 kHz		$28 \cdot 10^{-6} U$		
> 100 kHz to 300 kHz		$42 \cdot 10^{-6} U$		
> 300 kHz to 500 kHz		$85 \cdot 10^{-6} U$		
> 500 kHz to 1 MHz		$0.1 \cdot 10^{-3} U$		
> 220 mV to 700 mV	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$		
	> 20 Hz to 40 Hz	$22 \cdot 10^{-6} U$		
	> 40 Hz to 20 kHz	$12 \cdot 10^{-6} U$		
	> 20 kHz to 50 kHz	$12 \cdot 10^{-6} U$		
	> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$		
	> 100 kHz to 300 kHz	$14 \cdot 10^{-6} U$		
	> 300 kHz to 500 kHz	$27 \cdot 10^{-6} U$		
	> 500 kHz to 1 MHz	$40 \cdot 10^{-6} U$		

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Winsen - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and Sources	> 700 mV to 2.2 V	10 Hz to 20 Hz	$20 \cdot 10^{-6} U$	<i>U</i> = measured value
		> 20 Hz to 40 Hz	$14 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$10 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$10 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$11 \cdot 10^{-6} U$	
> 2.2 V to 7 V	> 700 mV to 2.2 V	10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$12 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$13 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$13 \cdot 10^{-6} U$	
> 7 V to 22 V	> 2.2 V to 7 V	10 Hz to 20 Hz	$17 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$11 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$11 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$11 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
> 22 V to 70 V	> 7 V to 22 V	> 300 kHz to 500 kHz	$30 \cdot 10^{-6} U$	
		> 500 kHz to 1MHz	$0.11 \cdot 10^{-3} U$	
		10 Hz to 20 Hz	$18 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$16 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$15 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$15 \cdot 10^{-6} U$	
> 70 V to 220 V	> 22 V to 70 V	> 50 kHz to 100 kHz	$25 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$25 \cdot 10^{-6} U$	
		> 300 kHz to 500 kHz	$40 \cdot 10^{-6} U$	
		> 500 kHz to 1 MHz	$0.13 \cdot 10^{-3} U$	
		10 Hz to 20 Hz	$19 \cdot 10^{-6} U$	
		> 20 Hz to 40 Hz	$18 \cdot 10^{-6} U$	
> 70 V to 220 V	> 70 V to 220 V	> 40 Hz to 20 kHz	$17 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$17 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$32 \cdot 10^{-6} U$	
		> 100 kHz to 300 kHz	$32 \cdot 10^{-6} U$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Winsen - Electrical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
AC voltage  Measuring devices and Sources	> 220 V to 1000 V	10 Hz to 20 Hz	$25 \cdot 10^{-6} U$	U = measured value
		> 20 Hz to 40 Hz	$27 \cdot 10^{-6} U$	
		> 40 Hz to 20 kHz	$45 \cdot 10^{-6} U$	
		> 20 kHz to 50 kHz	$45 \cdot 10^{-6} U$	
		> 50 kHz to 100 kHz	$65 \cdot 10^{-6} U$	
High Voltage	> 0.7 kV to 1 kV	50 Hz	$2.5 \cdot 10^{-3} U 0.25 V$	
	> 1 kV to 7 kV		$3.5 \cdot 10^{-3} U 2.0 V$	
Alternating current strength sources and measuring devices	100 $\mu$ A to 1 mA	10 Hz to 40 Hz	$120 \cdot 10^{-6} I$	I = measured value
		> 40 Hz to 1 kHz; > 1 kHz to 10 kHz;	$160 \cdot 10^{-6} I$ $60 \cdot 10^{-6} I$	
Current clamps	> 1 mA to 10 mA	10 Hz to 40 Hz	$46 \cdot 10^{-6} I$	
		> 40 Hz to 1 kHz; > 1 kHz to 10 kHz;		
		> 10 mA to 1 A	$17 \cdot 10^{-6} I$	
	> 1 A to 10 A	$32 \cdot 10^{-6} I$		
	> 10 A to 20 A	$39 \cdot 10^{-6} I$		
Current clamps	1 mA to 2.2 A	40 Hz to 5 kHz	$2 \cdot 10^{-3} I$	I = measured value
	> 2.2 A to 20 A	40 Hz to 5 kHz	$3 \cdot 10^{-3} I$	
	> 20 A to 800 A	40 Hz to 65 Hz	$4 \cdot 10^{-3} I$	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Winsen - Electrical measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Extended uncertainty of measurement	Remarks
	Measuring range	Measuring conditions / Procedures			
Capacity gauges	190 pF to < 400 pF	10 Hz to 10 kHz		$4 \cdot 10^{-3} C$ 8 pF	With 5520A / 5522A
	400 pF to < 1.1 nF	10 Hz to 10 kHz		$4.5 \cdot 10^{-3} C$ 8 pF	
	1.1 nF to < 3.3 nF	10 Hz to kHz		$4.0 \cdot 10^{-3} C$ 8 pF	
	3.3 nF to < 11 nF	10 Hz to 1 kHz		$2.5 \cdot 10^{-3} C$ 8 pF	
	11 nF to < 33 nF	10 Hz to 1 kHz		$2.5 \cdot 10^{-3} C$ 80 pF	
	33 nF to < 110 nF	10 Hz to 1 kHz		$2.5 \cdot 10^{-3} C$ 80 pF	
	110 nF to < 330 nF	10 Hz to 1 kHz		$4.5 \cdot 10^{-3} C$	
	330 nF to < 1.1 µF	10 Hz to 600 Hz		$4.5 \cdot 10^{-3} C$	
	1.1 µF to < 3.3 µF	10 Hz to 300 Hz		$4.5 \cdot 10^{-3} C$	
	3.3 µF to < 11 µF	10 Hz to 150 Hz		$4.5 \cdot 10^{-3} C$	
	11 µF to < 33 µF	10 Hz to 120 Hz		$6.0 \cdot 10^{-3} C$	
	33 µF to < 110 µF	10 Hz to 80 Hz		$6.5 \cdot 10^{-3} C$	
	110 µF to < 330 µF	DC to 50 Hz		$6.0 \cdot 10^{-3} C$	
	330 µF to < 1.1 mF	DC to Hz		$6.0 \cdot 10^{-3} C$	
	1.1 mF to < 3.3 mF	DC to 6 Hz		$6.0 \cdot 10^{-3} C$	
	3.3 mF to < 11 mF	DC to 2 Hz		$6.0 \cdot 10^{-3} C$	
	11 mF to < 33 mF	DC to 200.6 Hz		$8.0 \cdot 10^{-3} C$	
33 mF to 110 mF	DC to 0,2 Hz		$11 \cdot 10^{-3} C$		
	1 nF to 100 nF > 100 nF to 1000 nF	50 Hz to 10 kHz 50 Hz to 1 kHz > 1 kHz to 10 kHz	$1,0 - 10^{-3} C$ $1,0 - 10^{-3} C$ $2,5 - 10^{-3} C$	C: measured value with normal capacitances	
Frequency	1 mHz to 1 GHz		$2 \cdot 10^{-9} \cdot f_{UTP}$	$f$ = current measured value $UTP$ = Trigger uncertainty	
Time interval	1 µs to 1000 s		$2 \cdot 10^{-9} \cdot t$ 2 ns	$t$ = current measured value	
Speed optical	1 min <sup>-1</sup> up to 100.000 min <sup>-1</sup>	with light pulse generator	$8 \cdot 10^{-6}$ but not less than 0.006 min <sup>-1</sup>		
AC active power measuring devices		33 mV to 1000 V 45 Hz to 65 kHz $PF = 1$		$P$ = set value with Fluke 5520A/5522A $PF$ : Power factor	
	109 µW to < 11kW	33 mA to < 11 A	$1.4 \cdot 10^{-3} P$		
	363 mW up to 20 kW	11 A to 20 A	$2.0 \cdot 10^{-3} P$		
DC power Measuring devices	1 mW to 300 W		$0.5 \cdot 10^{-3} P$		
	> 300 W to 20 kW		$1.0 \cdot 10^{-3} P$		
Sources	1 mW to 300 W > 300 W to 1 kW > 1 kW to 1 MW	Product of $U$ and $I$ $1 \text{ mV} \leq U \leq 1000 \text{ V}$ $100 \text{ µA} \leq I \leq 2000 \text{ A}$	$30 - 10^{-6} P$ $200 - 10^{-6} P$ $300 - 10^{-6} P$	$P$ : calculated power	

Annex to the partial accreditation certificate D-K-15070-01-01

Permanent Laboratory Winsen - Electrical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
AC active power	5 mW to 50 kW > 2,5 W to 120 kW	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	$w(U_F)$ ... Uncertainty of the amplitude of the voltage fundamental $w(I_F)$ ... Uncertainty of the amplitude of the Current fundamental $w(\Phi_F)$ ... Uncertainty of the phase shift angle
AC reactive power	5 mvar to 50 kvar > 2,5 var to 120 kvar	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	$w(U_{rnc})$ ... Uncertainty of the voltage rms value $w(I_{rnc})$ ... uncertainty of the rms current value
Apparent power	0.1 VA to 50 kVA > 50 VA to 120 kVA	1 V to 1000 V 45 Hz to 65 Hz $0.05 \leq \cos \varphi \leq 1$ 0.1 A to 50 A > 50 A to 120 A	$2\sqrt{w(U_F)^2 + w(I_F)^2 + w(\Phi_F)^2}$ not less than $80 \cdot 10^{-6}$ not smaller than $0.14 \cdot 10^{-3}$	
Oscilloscopes				$U$ - measured value
Vertical deflection	5 mV to 5 V 5 mV to 120 V	$R_i = 50 \Omega$ $R_i = 1 M\Omega$	$3.5 \cdot 10^{-3} U$ 35 $\mu$ V $2.4 \cdot 10^{-3} U$ 40 $\mu$ V	Square wave voltage 10 Hz to 10 kHz
Horizontal deflection	5 ns to 520 ms > 20 ms to 5 s		$3 \cdot 10^{-6} T$ 1 ns $30 \cdot 10^{-6} T$ 1.2 $\cdot 10^{-4} T^2$	
Rise time	150 ps to 10 ms 250 ps to 10 ms	250 mV > 250 mV to 2.5 V	$35 \cdot 10^{-3} \tau_r$ 5 ps $35 \cdot 10^{-3} \tau_r$ 8 ps	$\tau_r$ = intrinsic rise time of the oscilloscope



**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Winsen - Dimensional measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Length *) Cylindrical setting standards, ring gauges: Diameter	1 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ = is the measured Diameter
Plug gauges: Diameter	1 mm to 200 mm	Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Test pins: Diameter	0,1 mm to 30 mm	VDI/VDE/DGQ 2618 Sheet 4.2:2007 Point 3.2.2 (Opt. 1)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	
Thread gauges (one and multi-start cylindrical male and female threads with straight flanks, symmetrical profile) Threaded mandrels: simple Flank diameter	1.4 mm to 200 mm nominal pitch: 0.3 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.8:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Three-wire method $d$ = is the measured diameter
Threaded rings: simpler Flank diameter	3 mm to 200 mm nominal pitch: 0.5 mm to 6 mm	VDI/VDE/DGQ 2618 Sheet 4.9:2006 Point 3.2.2 (Opt. 1)	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	Two-ball method $d$ = is the measured Diameter
Length of plane-parallel, spherical or cylindrical Measuring surfaces	0,01 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 19.1:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	$l$ is the measured Length
Diameter	0,01 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.1:2006 Point 3.3.4 (Opt. 3), Point 3.3.5 (Opt. 4)	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	$d$ is the measured Diameter
Feeler gauges	0,03 mm to 2,00 mm	DIN 2275:2014	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	$l$ is the measured length
Adjustment dimensions for Outside micrometers	25 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 4.4:2009	$1.5 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot l$	

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Permanent Laboratory Winsen - Dimensional measurands**

**Calibration and Measurement Capabilities (CMC)**

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Throat gauges	3 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 4.7:2005 Point 3.3.2 (Opt. 2)	$0.8 \mu\text{m} \cdot 2 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured Diameter
Caliper for outside, inside and Depth measurements	0 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 9.1:2006	$30 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length
Depth caliper,	>500mm to 1000 mm	VDI/VDE/DGQ 2618 Sheet 9.2:2006	$50 \mu\text{m} \cdot 30 \cdot 10^{-6} \cdot l$	
Height caliper		VDI/VDE/DGQ 2618 Sheet 9.3:2006		
Outside micrometers	0 mm to 500 mm	VDI/VDE/DGQ 2618 Sheet 10.1:2001	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Fine pointer measuring screws	0 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 10.3:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Micrometer head screws	0 mm to 50 mm	VDI/VDE/DGQ 2618 Sheet 10.4:2008	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Depth gauges	0 mm to 300 mm	VDI/VDE/DGQ 2618 Sheet 10.5:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 2-point contact on the calibration object	13 mm until 300 mm >300mm until 500 mm	VDI/VDE/DGQ 2618 Sheet 10.7:2010	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$ $5 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Inside micrometers with 3-line contact on the calibration object	3 mm to 100 mm	VDI/VDE/DGQ 2618 Sheet 10.8:2002	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot d$	<i>d</i> is the measured diameter
Lever gauges (quick probes) for External measurements	to 200 mm	VDI/VDE/DGQ 2618 Sheet 12.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length
Lever gauges (quick probe) for indoor measurements	2 mm to 200 mm	VDI/VDE/DGQ 2618 Sheet 13.1:2005	$7 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	
Dial gauges	0 mm to 100 mm	VDI/VDE/DGQ/DKD 2618 Sheet 11.1:2021	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	mechanical dial gauges
		VDI/VDE/DGQ/DKD 2618 Sheet 11.4:2020	$3 \mu\text{m} \cdot 10 \cdot 10^{-6} \cdot l$	electronic digital dial gauges
Fine pointer	0 mm to 3 mm	VDI/VDE/DGQ 2618 Sheet 11.2:2002	0.6 $\mu\text{m}$	
Feeler lever gauges	0 mm to 1,6 mm	VDI/VDE/DGQ 2618 Sheet 11.3:2002	1.0 $\mu\text{m}$	
Tape measures and scales				
Tape measures	0 m until 100 m	4_VB_00237_EN V1	$50 \mu\text{m} \cdot 20 \cdot 10^{-6} \cdot l$	<i>l</i> is the measured length
Standards	0 m until 3 m			

**Annex to the partial accreditation certificate D-K-15070-01-01**

**Abbreviations used:**

CMC	Calibration and measurement capabilities
DIN	German Institute for Standardization R.A.
DKD	Guideline of the German Calibration Service (DKD), published by the Federal Physical-Technical Institute (PTB)
DKD-R	Guideline of the German Calibration Service (DKD), published by the Federal Physical-Technical Institute (PTB)
DGQ	German Society for Quality R.A.
VDE	Association for Electrical, Electronic & Information Technologies R.A.
VDI	Association of German Engineers R.A.
VDI/VDE/DGQ 2618	VDI guideline series for test equipment monitoring
VB	Self-developed calibration method of the laboratory

## German Accreditation Body

### Annex to the partial accreditation certificate D-K-15070-01-02 according to DIN EN ISO/IEC 17025:2018

**Valid from:** 30.05.2023

**Date of issue:** 30.05.2023

This document attachment is part of the accreditation certificate D-K-15070-01-00.

Partial Accreditation Certificate Holder:

**Testo Industrial Services GmbH**  
**Gewerbestraße 3, 79199 Kirchzarten,**  
**Germany**

**The German original version**  
**„Anlage zur Akkreditierungsurkunde D-K-15070-01-02 nach DIN EN ISO/IEC 17025:2018“**  
**is valid.**

The calibration laboratory meets the requirements according to DIN EN ISO/IEC 17025:2018 to perform the conformity assessment activities listed in this annex. The calibration laboratory shall comply with additional legal and normative requirements, where applicable, including those in relevant sectoral programs, provided that they are explicitly confirmed below.

The requirements for the management system in DIN EN ISO/IEC 17025 are written in a language relevant for calibration laboratories and are overall in accordance with the principles of DIN EN ISO 9001.

Calibrations at the sites:

**Calibration laboratory Kirchzarten, Gewerbestraße 3, 79199 Kirchzarten**  
**Calibration laboratory Kirchzarten, Erich-Rieder Straße 4, 79199 Kirchzarten**  
**Calibration laboratory Munich, Nikolaus-Otto-Straße 2, 85221 Dachau**  
**Calibration laboratory Essen, Alte Landstraße 3c, 45329 Essen**  
**Calibration laboratory Hamburg, Meiendorfer Straße 205, 22145 Hamburg**  
**Calibration laboratory Mörfelden-Walldorf, Kurhessenstraße 11,**  
**64546 Mörfelden-Walldorf**  
**Calibration laboratory Winsen, Tönnhäuser Weg 100-106, 21423 Winsen (Luhe)**

*This deed annex is valid only together with the deed issued in writing and reflects the status at the date of issue. The current status of valid and monitored accreditation can be found in the database of accredited bodies of the German Accreditation Body ([www.dakks.de](http://www.dakks.de))*

Calibration in the fields:

**Thermodynamic measurands**

**Temperature measurands**

- Resistance thermometer <sup>a)</sup>
- Thermocouples, thermocouples <sup>a)</sup>
- Radiation Thermometer
- Temperature fixed point cells
- Temperature block calibrators <sup>a)</sup>
- Temperature indicators and simulators <sup>a), b)</sup>
- Climate cabinets (temperature) <sup>c)</sup>
- Temperature transmitter, data logger <sup>a)</sup>

**Humidity measurement variables**

- Measuring instruments for relative humidity <sup>a)</sup>
- Measuring instruments for absolute humidity <sup>a)</sup>
- Climate cabinets (humidity) <sup>c)</sup>

**Chemical and medical measurands Chemical analyses and reference materials**

- Measuring instruments for electrolytic

**Conductivity**

- pH value
- Gas mixtures

**Mechanical measured variables**

- Force
- Scales <sup>c)</sup>
- Pressure <sup>a)</sup>
- Torque <sup>a)</sup>
- Acceleration flow

**measurement**

**variables**

- Flow velocity of gases
- Volume of flowing gases
- Mass of flowing gases
- Volume of flowing liquids
- Mass of flowing liquids Acoustic

**measurands**

a) also on-site calibration

b) also mobile laboratory

c) On-site calibration only

Within the measurands/calibration items marked with <sup>\*</sup>), the calibration laboratory is permitted to apply the standards/calibration guidelines listed here with different editions without requiring prior information and approval by DAkkS. The calibration laboratory has an up-to-date list of all standards/calibration guidelines in the flexible accreditation area.

## Calibration and measurement capabilities

<b>Location Calibration laboratory Gewerbestraße 3, 79199 Kirchzarten .....</b>	<b>4</b>
Permanent laboratory Calibration laboratory Gewerbestraße 3 - Mechanical measurands .....	4
Permanent Laboratory Calibration Laboratory Gewerbestraße 3 - Acoustics.....	6
Permanent laboratory Calibration laboratory Gewerbestraße 3 - Thermodynamic measurands.....	7
Permanent laboratory calibration laboratory Gewerbestraße 3 - flow measured variables .....	12
Permanent laboratory Calibration laboratory Gewerbestraße 3 - Chemical and medical measurands.....	12
<b>On-site calibration.....</b>	<b>13</b>
On-site calibration - Thermodynamic measurands.....	13
On-site calibration - Mechanical measurands .....	16
On-site calibration - flow measured variables.....	18
<b>Mobile laboratory .....</b>	<b>19</b>
Mobile laboratory - Thermodynamic measurands .....	19
Mobile laboratory - Mechanical measurands.....	20
<b>Location Calibration Laboratory Kirchzarten, Erich-Rieder Straße 4, 79199 Kirchzarten.....</b>	<b>21</b>
Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Thermodynamic measurands .....	21
Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Mechanical measurands .....	21
<b>Calibration Laboratory Munich, Nikolaus-Otto-Straße 2, 85221 Dachau .....</b>	<b>22</b>
Permanent Laboratory Munich - Mechanical measurands .....	22
Permanent Laboratory Munich - Flow Measured Values .....	23
Permanent Laboratory Munich - Thermodynamic Measurands.....	25
<b>Calibration Laboratory Essen, Alte Landstraße 3c, 45329 Essen.....</b>	<b>26</b>
Permanent Laboratory Essen - Thermodynamic Measurands.....	26
<b>Calibration Laboratory Hamburg, Meiendorfer Straße 205, 22145 Hamburg .....</b>	<b>26</b>
Permanent Laboratory Hamburg - Thermodynamic Measurands.....	26
<b>Calibration Laboratory Mörfelden-Walldorf, Kurhessenstraße 11, 64546 Mörfelden-Walldorf.....</b>	<b>27</b>
Permanent Laboratory Mörfelden-Walldorf - Thermodynamic Measurands .....	27
<b>Calibration Laboratory Winsen, Tönnhäuser Weg 100-106, 21423 Winsen (Luhe).....</b>	<b>29</b>
Permanent Laboratory Winsen - Thermodynamic Measurands.....	29
Permanent Laboratory Winsen - Mechanical measurands.....	29
<b>Abbreviations used .....</b>	<b>30</b>

**Location Calibration laboratory Gewerbestraße 3, 79199 Kirchzarten**

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Mechanical measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
<b>Acceleration</b> vibration transducer, vibration measuring device	0.1 m/s <sup>2</sup> to 20 m/s <sup>2</sup>	Sinus excitation DKD-R 3-1 Sheet 3:2020		Complex transfer coefficient (magnitude / phase). Transducer mass up to 0.9 kg, displacement amplitude up to 100 mm
		0.2 Hz to < 0.4 Hz	2,5 % / 1,6 °	
		0.4 Hz to < 1 Hz	1,5 % / 1,6 °	
		1 Hz to < 16 Hz	0,8 % / 0,8 °	
		16 Hz	0,55 % / 0,6 °	
		> 16 Hz to 63 Hz	0,8 % / 0,8 °	
		> 63 Hz to 160 Hz	1,0 % / 1,1 °	
	1 m/s <sup>2</sup> to 200 m/s <sup>2</sup>	Sinus excitation DKD-R 3-1 Sheet 3:2020		Complex transfer coefficient (magnitude / phase). Transducer mass up to 0.2 kg, displacement amplitude up to 8 mm
		5 Hz to < 10 Hz	1,5 % / 1,5 °	
		10 Hz to < 20 Hz	0,8 % / 0,8 °	
		20 Hz to 1 kHz	0,6 % / 0,6 °	
		> 1 kHz to 5 kHz	0,8 % / 0,8 °	
		> 5 kHz to 10 kHz	2,0 % / 1,5 °	
		> 10 kHz to 15 kHz	2,5 % / 2,5 °	
	> 15 kHz to 20 kHz	3,0 % / 3,0 °		
	1 m/s <sup>2</sup> to 500 m/s <sup>2</sup>	Sinus excitation DKD-R 3-1 Sheet 3:2020		Complex transfer coefficient (magnitude / phase). Transducer mass up to 0.5 kg, displacement amplitude up to 10 mm
		3 Hz to < 5 Hz	1,6 % / 1,1 °	
		5 Hz to < 20 Hz	1,1 % / 1,1 °	
		20 Hz to < 80 Hz	0,8 % / 0,8 °	
		80 Hz	0,55 % / 0,6 °	
		> 80 Hz to 1 kHz	0,8 % / 0,8 °	
> 1 kHz to 5 kHz		1,3 % / 1,1 °		
> 5 kHz to 10 kHz	2,3 % / 1,1 °			
Vibration calibrator	0.1 m/s <sup>2</sup> to 100 m/s <sup>2</sup>	Sinus excitation DIN ISO 16063-44:2019		
		10 Hz to < 20 Hz	0,8 %	
		20 Hz to 1 kHz	0,6 %	
		> 1 kHz to 5 kHz	0,8 %	
		> 5 kHz to 10 kHz	2,0 %	

Permanent laboratory Calibration laboratory Gewerbestraße 3 - Mechanical measurands

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
<b>Print *)</b> Negative and positive overpressure $p_e$	-1 bar to -0.03 bar	DKD-R 6-1: 2014	$10 \mu\text{bar}^{5 \cdot 10^{-5}} - p_e$	$p_e =$ measured value pressure medium: Gas	
	> -0.03 bar to < -1 mbar		$50 \mu\text{bar}^{1 \cdot 10^{-4}} - p_e$		
	-1 mbar to < 0 mbar		4 $\mu\text{bar}$		
	0 bar		0.6 $\mu\text{bar}$		
	> 0 mbar to < 0.2 mbar		4 $\mu\text{bar}$		
	0.2 mbar to 3.6 mbar		0.7 $\mu\text{bar}$		
	> 3.6 mbar to 0.2 bar		$2 \mu\text{bar}^{1 \cdot 10^{-4}} - p_e$		
	> 0.2 bar to 2 bar		$30 \mu\text{bar} \cdot 2^{5 \cdot 10^{-5}} - p_e$		
	> 2 bar to 20 bar		$0.05 \text{ mbar} \cdot 2^{5 \cdot 10^{-5}} - p_e$		
	> 20 bar to 100 bar		$0.5 \text{ mbar} \cdot 2^{5 \cdot 10^{-5}} - p_e$		
	> 100 bar to 400 bar		$6 \text{ mbar} \cdot 3^{5 \cdot 10^{-5}} - p_e$		
	0 bar		$7^{10^{-5}} - p_e$ at least 7.5 mbar		Print medium: Oil
	2 bar to 1200 bar				
Absolute pressure $p_{\text{abs}}$	0.01 bar to 2 bar	DKD-R 6-1: 2014  $p_{\text{abs}} = p_e \cdot p_{\text{amb}}$	$15 \mu\text{bar} \cdot 2^{5 \cdot 10^{-5}} - p_{\text{abs}}$	$p_{\text{abs}} =$ measured value pressure medium gas  The measurement uncertainty of the residual gas measurement must be taken into account.	
	> 2 bar to 20 bar		$170 \mu\text{bar} \cdot 2^{5 \cdot 10^{-5}} - p_{\text{abs}}$		
	> 20 bar to 101 bar		$0.6 \text{ mbar} \cdot 2^{5 \cdot 10^{-5}} - p_{\text{abs}}$		
	> 101 bar to 401 bar		$7 \text{ mbar} \cdot 3^{5 \cdot 10^{-5}} - p_{\text{abs}}$	The measurement uncertainty of the barometer must be taken into account	
	1 bar		$7^{10^{-5}} - p_{\text{abs}}$ at least 7.5 mbar		$p_{\text{abs}} =$ measured value pressure medium: Oil Measurement uncertainty of the barometer must be taken into account
	3 bar to 1201 bar				



**Permanent Laboratory Calibration Laboratory Gewerbestraße 3 - Acoustics**

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Acoustics</b> Measuring microphone/  Free-field open- circuit or free- field operating transmission dimension	-60 dB to 20 dB (referred to 1 V/Pa)	DIN EN 61094-8:2013 Substitution process in a low-reflection chamber with ½"- reference microphone. 125 Hz to 8 kHz > 8 kHz to 20 kHz	0.35 dB 0.50 dB	
Measuring microphone/  Pressure-idle or pressure-operating transmission dimension	-60 dB to 20 dB (referred to 1 V/Pa)  250 Hz / 114 dB 1000 Hz / 94 dB 1000 Hz / 114 dB	DIN EN IEC 60942:2018 Calibration with reference standard Pistonphone  Calibrator  Calibrator	0.2 dB	
	-60 dB to 20 dB (referred to 1 V/Pa)	DIN EN 61094-5:2016 Comparative measurement with SQ-4.2 electro-acoustic coupler 31.5 Hz to 5 kHz > 5 kHz to 16 kHz	0.25 dB 0.50 dB	Only 1/2" - microphones
Sound level meter/  Sound pressure level display (free field)	250 Hz / 114 dB 1000 Hz / 94 dB 1000 Hz / 114 dB	DIN EN 61672-3:2017 Calibration with reference standard Pistonfon calibrator Calibrator	0.2 dB	
		DIN EN 61672-3:2017 Comparative measurement with SQ-4.2 electro-acoustic coupler 31.5 Hz to 10 kHz > 10 kHz to 16 kHz	0.35 dB 0.60 dB	Only 1/2" - microphones
	74 dB to 94 dB (referred to 20 µPa)	DIN EN 61672-3:2017 Substitution process in a low-reflection chamber with ½"- Reference microphone 125 Hz to < 250 Hz 250 Hz to 8 kHz > 8 kHz to 20 kHz	0.50 dB 0.40 dB 0.60 dB	
Sound calibrator/  sound pressure level	70 dB to 130 dB	DIN EN IEC 60942:2018 Calibration with ½" reference microphone 250 Hz or 1 kHz	0.15 dB	
	Frequency: 250 Hz or 1 kHz		0.1 Hz	

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature</b> Fixed point cells	-189,3442 °C	G-ITS-90, Part 2.3:2021 Argon triple point	4.0 mK	Comparison with reference fixed-point cell using normal resistance thermometers
	-38,8344 °C	G-ITS-90, Part 2.4:2021 Mercury triple point	1.0 mK	
	0,01 °C	G-ITS-90, Part 2.2:2018 Water triple point	0.5 mK	
	29,7646 °C	G-ITS-90, Part 2.4:2021 Gallium melting point	0.8 mK	
	156,5985 °C	G-ITS-90, Part 2.4:2021 Indium solidification point	2.5 mK	
	231,928 °C	G-ITS-90, Part 2.4:2021 Tin solidification point	1.5 mK	
	419,527 °C	G-ITS-90, Part 2.4:2021 Zinc solidification point	2.0 mK	
	660,323 °C	G-ITS-90, Part 2.4:2021 Aluminum- solidification point	7.0 mK	
Standard platinum resistance thermometers (SPRT), direct reading thermometers and temperature transmitters with resistance sensor (SPRT)	-196 °C to -189.3442 °C	EURAMET Technical Guide No. 1:2017	8.0 mK	Extrapolation
	-189,3442 °C	G-ITS-90, Part 2.3:2021 Argon triple point	4.0 mK	Calibration at temperature fixed points
	-38,8344 °C	G-ITS-90, Part 2.4:2021 Mercury triple point	1.5 mK	
	0,01 °C	G-ITS-90, Part 2.2:2018 Water triple point	0.5 mK	
	29,7646 °C	G-ITS-90, Part 2.4:2021 Gallium melting point	1.0 mK	
	156,5985 °C	G-ITS-90, Part 2.4:2021 Indium solidification point	2.5 mK	
	231,928 °C	G-ITS-90, Part 2.4:2021 Tin solidification point	2.5 mK	
	419,527 °C	G-ITS-90, Part 2.4:2021 Zinc solidification point	2.5 mK	
	660,323 °C	G-ITS-90, Part 2.4:2021 Aluminum- solidification point	7.0 mK	
	-189,3442 °C to 0.01 °C	G-ITS-90, Part 5:2021 Fixed points: Ar, Hg, TPW	6.0 mK	
	-38.8344 °C to 29.7646 °C	G-ITS-90, Part 5:2021 Fixed points: Hg, TPW, Ga	2.0 mK	
	0 °C to 156,5985 °C	G-ITS-90, Part 5:2021 Fixed points: TPW, In	3.5 mK	
	0 °C to 231.928 °C	G-ITS-90, Part 5:2021 Fixed points: TPW, In, Sn	3.5 mK	
	0 °C to 419.527 °C	G-ITS-90, Part 5:2021 Fixed points: TPW, Sn, Zn	4.0 mK	
0 °C to 660.323 °C	G-ITS-90, Part 5:2021 Fixpkte.: TPW, Sn, Zn, Al	8.0 mK		

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Resistance thermometers, direct reading thermometers with resistance sensor *)	-196 °C	DKD-R 5-1:2018 in liquid nitrogen	15 mK	Comparison with standard resistance thermometers
	0,00 °C	DKD-R 5-1:2018 Ice Point	5.0 mK	
	-120 °C to 200 °C	DKD-R 5-1:2018 in liquid bath	10 mK	
	> 200 °C to 300 °C		15 mK	
	> 300 °C to 420 °C		20 mK	
	> 420 °C to 500 °C		50 mK	
> 500 °C to 660 °C	DKD-R 5-1:2018 in the tube furnace	0,20 K		
Temperature transmitter with resistance sensor *)	-40 °C to 200 °C	DKD-R 5-1:2018 in liquid bath	15 mK	Comparison with standard resistance thermometers
	> 200 °C to 500 °C		25 mK	
direct reading thermometers, temperature transmitters and data loggers with resistance sensor *)	-40 °C to < 0 °C	DKD-R 5-1:2018 in the climatic chamber	0,30 K	Comparison with standard resistance thermometers
	0 °C to 50 °C		0,15 K	
	> 50 °C to 80 °C		0,25 K	
	> 80 °C to 120 °C		0,40 K	
	> 120 °C to 180 °C		0,90 K	
Precious metal thermocouples *)	0,01 °C	DKD-R 5-3 Water triple point	0,4 K	Calibration at temperature fixed points of the ITS 90
	231,928 °C	DKD-R 5-3 Tin solidification point	0,4 K	
	419,527 °C	DKD-R 5-3 Zinc solidification point	0,4 K	
	660,323 °C	DKD-R 5-3 Aluminum solidification point	0,4 K	
	961,78 °C	DKD-R 5-3 Silver solidification point	0,5 K	
	0 °C to 1000 °C	DKD-R 5-3 with DKD-R 5-6 at temperature fixed points	0,6 K	Calibration at fixed temperature points with characteristic curve determination  The measurement uncertainty refers to the characteristic curve in the specified range
Precious metal thermocouples, direct reading thermometers with attached precious metal thermocouple sensor *)	-40 °C to 500 °C	DKD-R 5-3:2018 in liquid bath	0,5 K	Comparison with standard resistance thermometers
	> 500 °C to 1000 °C	DKD-R 5-3:2018 in the tube furnace	0,8 K	Comparison with standard thermocouples
	> 1000 °C to 1200 °C	DKD-R 5-3:2018 in the ball furnace	1,6 K	

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Thermodynamic measurands**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Non-precious metal thermocouples, direct reading with connected non-precious metal thermocouple sensor *)	-196 °C	DKD-R 5-3:2018 in liquid nitrogen	0,5 K	Comparison with standard resistance thermometers
	-80 °C to 200 °C	DKD-R 5-3:2018 in liquid bath	0,2 K	
	200 °C to 400 °C		0,4 K	
	> 400 °C to 500 °C		0,5 K	
	> 500 °C to 1000 °C	DKD-R 5-3:2018 in the tube furnace	1,0 K	Comparison with standard thermocouples
Temperature transmitter and data logger with thermocouple sensor *)	-80 °C to 200 °C	DKD-R 5-3:2018 Liquid bath	0,5 K	Comparison with standard resistance thermometers
	> 200 °C to 500 °C		1,0 K	
	> 500 °C to 1000 °C	DKD-R 5-3:2018 Tube furnace	2,0 K	Comparison with standard thermocouples
Liquid glass thermometer *)	-80 °C to < 0 °C	PTB test rule volume 2:2003	20 mK	Comparison with standard resistance thermometers
	0 °C to 200 °C		10 mK	
Circulated thermostats and baths	-80 °C to < 200 °C	3-APD-0-0155-EN: 2023-01	10 mK	Comparison with standard resistance thermometers
	200 °C to 300 °C		15 mK	
Temperature block calibrators *)	-90 °C to 125 °C	DKD-R 5-4:2018	0,04 K	Comparison with standard resistance thermometers
	> 125 °C to 150 °C		0,05 K	
	> 150 °C to 300 °C		0,25 K	
	> 300 °C to 650 °C		0,50 K	
	> 650 °C to 800 °C		2,5 K	Comparison with standard thermocouples
	> 800 °C to 1000 °C		4 K	
Surface temperature sensor	50 °C to 100 °C	3-APD-0-0016-EN: 2023-01	0,8 K	t = measured value in °C
	> 100 °C to 500 °C		0.008 K - t/°C	
Radiation thermometer	-18 °C to 60 °C	3-APD-0-0018-EN: 2023-01 Spectral range: 8 µm to 14 µm	0,6 K	Calibration with liquid flushed cavity radiator
	> 60 °C to 100 °C		0,9 K	
	> 100 °C to 350 °C		1,2 K	
Temperature simulators for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display devices for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Temperature indicators and simulators for Precious metal thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature indicators and simulators for Non-precious metal thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	
<b>Frost and dew point temperature</b> Dew point mirror, -transmitter, -hygrometer	-32 °C to < -25 °C	3-APD-0-0036-EN: 2023-01 1-Temperature 2-Pressure Humidity Generator	90 mK	Primary generator
	-25 °C to < 0 °C	3-APD-0-0035-EN: 2023-01 1-Temp. 1/ 2-Pressure Humidity Generator	35 mK	
	0 °C to < 70 °C		30 mK	
	70 °C to < 90 °C		40 mK	
	90 °C to 95 °C		45 mK	
	-20 °C to 50 °C	3-APD-0-0037-EN: 2023-01 in the climatic chamber	0,2 K	Comparison with dew point hygrometer
> 50 °C to 70 °C	0,25 K			

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Relative humidity</b> Dew point mirror	2 % to 98 %	3-APD-0-0035-EN: 2023-01	0.1 % 0.003 - <i>rH</i>	<i>rH</i> = measured value
Electrical psychrometer	2 % to 98 %	1-Temp.- 1-/ 2-Pressure- Humidity generator with temperature chamber 3 °C to 98 °C Frost point ≥ -25 °C	0.3 % 0.007 - <i>rH</i>	Uncertainty of measurement expressed as absolute value of relative humidity
Hygrometers, data loggers, transmitters *)	2 % to 98 %	DKD-R 5-8:2019 1-Temp.- 1-/ 2-Pressure- Humidity generator with temperature chamber 3 °C to 98 °C Frost point ≥ -25 °C	0.2 % 0.003 - <i>rH</i>	
	5 % to 30 %	DKD-R 5-8:2019 in the climatic chamber temperature range: -18 °C to 0 °C Frost point ≥ -32 °C	2,0 %	References: Dew point mirror and resistance thermometer Uncertainty of measurement expressed as absolute value of relative humidity
	> 30 % to 60 %		3,9 %	
	> 60% to 95%		6,2 %	
	5 % to 30 %	DKD-R 5-8:2019 in the climatic chamber temperature range: > 0 °C to 25 ° Frost point ≥ -32 °C	1,0 %	
	> 30 % to 60 %		1,8 %	
	> 60% to 95%		3,3 %	
	5 % to 30 %	DKD-R 5-8:2019 in the climatic chamber temperature range: > 25 °C to 50 °C Frost point ≥ -32 °C	0,6 %	
	> 30 % to 60 %		1,1 %	
	> 60% to 95%		1,8 %	
	5 % to 30 %	DKD-R 5-8:2019 in the climatic chamber temperature range: > 50 °C to 80 °C Frost point ≥ -32 °C	0,8 %	
	> 30 % to 60 %		1,5 %	
	> 60% to 95%		2,4 %	

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Flow measured variables**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Gas flow rate</b> Flow velocity from Gases	0.1 m/s to 68 m/s	Measurement with low-turbulence free jet	0.5 %; but not less than 0.01 m/s	
Volume flow rated $V/dt$ of flowing gases	15 m <sup>3</sup> /h to 2000 m <sup>3</sup> /h	3-APD-0-0055-EN: 2023-01	1.5 %; but not less than 0.3 m <sup>3</sup> /h	
Mass flow rate $dm/dt$ of flowing gases	15 kg/h to 2000 kg/h	Air under ambient conditions	1.5 %; but not less than 0.3 m <sup>3</sup> /h	

**Permanent laboratory Calibration laboratory Gewerbestraße 3 - Chemical and medical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Electrolytic conductivity</b> conductivity meters and facilities	1.3 $\mu$ S/cm up < 5 $\mu$ S/cm	3-APD-0-0265-EN: 2023-01	1,5 %	discrete values
	5 $\mu$ S/cm to < 100 $\mu$ S/cm		0,7 %	
	100 $\mu$ S/cm to < 706 $\mu$ S/cm		0,5 %	
	706 $\mu$ S/cm up 100 mS/cm		0,3 %	
<b>pH - value</b> pH meters and facilities	1.68 pH to 10 pH	3-APD-0-0266-EN: 2023-01	0.03 pH	
<b>Exhaust gas / flue gas measuring instruments</b> Gas concentration Oxygen <sub>O2</sub> Carbon monoxide CO Carbon dioxide <sub>CO2</sub> Nitric Oxide NO Nitrogen dioxide <sub>NO2</sub> Sulfur dioxide <sub>SO2</sub> Sulfur- hydrogen <sub>H2S</sub> Methane <sub>CH4</sub>	0,0 % vol 1,4 % vol 2,5 % vol 5,0 % vol	3-APD-0-0169-EN: 2023-01	0,02 % vol 0.035 % vol 0.060 % vol 0,12 % vol	ppm vol = 10 <sup>-6</sup> - m <sup>3</sup> /m <sup>3</sup> % vol = 10 <sup>-2</sup> - m <sup>3</sup> /m <sup>3</sup>
	80 ppm vol 100 ppm vol 300 ppm vol 400 ppm vol 700 ppm vol 5000 ppm vol		3 ppm vol 3 ppm vol 7.5 ppm vol 10 ppm vol 17.5 ppm vol 125 ppm vol	
	0,0 % vol 0,1 % vol 0,5 % vol 17 % vol 38,5 % vol		0,03 % vol 0,03 % vol 0,03 % vol 0,41 % vol 0,9 % vol	
	150 ppm 300 ppm		3.8 ppm vol 7.5 ppm vol	
	100 ppm		3.0 ppm vol	
	100 ppm		3.0 ppm vol	
	200 ppm		7.0 ppm vol	
	5000 ppm		120 ppm vol	

## On-site calibration

### On-site calibration - Thermodynamic measurands

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			Remarks
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	
Temperature resistance thermometers, direct reading thermometers with resistance sensor *)	0,00 °C	DKD-R 5-1:2018 Ice Point	10 mK	Comparison with standard resistance thermometers
	-80 °C to 200 °C	DKD-R 5-1:2018 in liquid bath	10 mK	
	> 200 °C to 300 °C	DKD-R 5-1:2018 in the block calibrator	0,5 K	
	> 300 °C to 660 °C	DKD-R 5-1:2018 in the calibration furnace	3,0 K	Comparison with standard thermocouples
Temperature trans- mitter and data logger with resistance sensor *)	-80 °C to 200 °C	DKD-R 5-1:2018 in liquid bath	25 mK	Comparison with standard resistance thermometers
	> 200 °C to 300 °C		0,5 K	
	> 300 °C to 500 °C	DKD-R 5-1:2018 in the block calibrator	3,0 K	
direct reading thermometers, temperature transmitters and data loggers with resistance sensor *)	-40 °C to < 0 °C	DKD-R 5-1:2018 in the climatic chamber	0,30 K	Comparison with standard resistance thermometers
	0 °C to 50 °C		0,15 K	
	> 50 °C to 80 °C		0,25 K	
	> 80 °C to 120 °C		0,40 K	
	> 120 °C to 180 °C		0,90 K	
	0 °C to 70 °C	DKD-R 5-1:2018 1-temperature 2-pressure humidity generator with temperature chamber	0,05 K	Comparison with resistance thermometer
Non-precious metal thermocouples, direct reading thermometers with non- precious metal thermocouple sensor *)	-80 °C to 200 °C	DKD-R 5-3:2018 in liquid bath	0,2 K	Comparison with standard resistance thermometers
	> 200 °C to 300 °C	DKD-R 5-3:2018 in the block calibrator	0,5 K	
	> 300 °C to 1000 °C	DKD-R 5-3:2018 in the calibration furnace	3,0 K	Comparison with standard thermocouples
Temperature trans- mitter with thermo- element sensor *)	-80 °C to 200 °C	DKD-R 5-3:2018 in liquid bath	0,3 K	Comparison with standard resistance thermometers
	> 200 °C to 1000 °C	DKD-R 5-3:2018 in the calibration furnace	3,5 K	Comparison with standard thermocouples
Circulating thermostats, precision baths	-80 °C to < 200 °C	3-APD-0-0155-EN: 2023-01	10 mK	Comparison with standard resistanc thermometers
	200 °C to 300 °C		15 mK	



**On-site calibration - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Block calibrators *)	-90 °C to 125 °C	DKD-R 5-4:2018	0,04 K	Comparison with standard resistance thermometers
	> 125 °C to 150 °C		0,05 K	
	> 150 °C to 300 °C		0,25 K	
	> 300 °C to 650 °C		0,5	
	> 650 °C to 800 °C		2,5 K	Comparison with standard thermocouples
	> 800 °C to 1000 °C		4 K	
Temperature simulators for resistance-thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display devices for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for precious metal thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature indicators and simulators for non-precious metal thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	
Measuring locations in climatic chambers with circulating air *)	-90 °C to 0 °C	DKD-R 5-7:2018 Method C	0,3 K	Measuring medium: Air  Comparison with resistance thermometers
	> 0 °C to 100 °C		0,2 K	
	> 100 °C to 200 °C		0,3 K	
	> 200 °C to 350 °C		0,5 K	
Measuring locations in climatic chambers without circulating air *)	-90 °C to 0 °C	DKD-R 5-7:2018 Method C	0,5 K	Measuring medium: Air  Comparison with resistance thermometers
	> 0 °C to 100 °C		0,3 K	
	> 100 °C to 200 °C		0,5 K	
	> 200 °C to 350 °C		0,8 K	
Air conditioners with circulating air *)	-90 °C to 0 °C	DKD-R 5-7:2018 Method A and B	0,5 K	Measuring medium: Air  Comparison with resistance thermometers
	> 0 °C to 100 °C		0,3 K	
	> 100 °C to 200 °C		0,5 K	
	> 200 °C to 350 °C		0,8 K	
Air conditioners without recirculation *)	-90 °C to 0 °C	DKD-R 5-7:2018 Method A and B	0,8 K	Measuring medium: Air  Comparison with resistance thermometers
	> 0 °C to 100 °C		0,5 K	
	> 100 °C to 200 °C		0,8 K	
	> 200 °C to 350 °C		1,2 K	

**Annex to the partial accreditation certificate D-K-15070-01-02**

**On-site calibration - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Dew point temperature</b> Dew point meters, -hygrometer	-25 °C to 70 °C	3-APD-0-0164-EN: 2023-01 1-Temperature 2-Pressure Humidity Generator	0,09 K	Comparison with c point mirror
<b>Relative humidity</b> measuring points in climatic chambers with circulating air *)	5 % to 30 %	DKD-R 5-7:2018 Method C Temperature range: -10 °C to 95 °C	0,3 %	Measuring medium Humidity reference calculated from de point and air temperature Uncertainty of measurement expressed as absol value of relative humidity
	> 30 % to 60 %		0,4 %	
	> 60% to 98%		0,6 %	
Air conditioners with circulating air *)	5 % to 30 %	DKD-R 5-7:2018 Method A and B Temperature range: -10 °C to 95 °C	0,4 %	Uncertainty of measurement expressed as absol value of relative humidity
	> 30 % to 60 %		0,6 %	
	> 60% to 98%		0,8 %	
<b>Relative humidity</b> hygrometers, data loggers, transmitters *)	10 % to 95 %	DKD-R 5-8:2019 1-Temperature 2-Pressure Humidity Generator Chamber temperature: 0 °C to 70 °C	0,6 %	References: Dew p mirror and resistar thermo- meter Uncertainty of measurement expressed as absol value of relative humidity
Hygrometers, data loggers, transmitters *)	10 % to 95 %	DKD-R 5-8:2019 Humidity generator with limited useful volume Chamber temperature: 0 °C to 40 °C	0,9 %	References: Dew p mirror and resistar thermo- meter Uncertainty of measurement expressed as absol value of relative humidity
	10 % to 95 %	DKD-R 5-8:2019 Humidity generator with limited useful volume Chamber temperature: > 40 °C to 70 °C	2,0 %	

**Annex to the partial accreditation certificate D-K-15070-01-02**

**On-site calibration - Mechanical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Scales</b> *) non- automatic electronic scales	0 kg to 10 kg	EURAMET cg-18 Version 4.0 Calibration at the installation site	1 - 10 <sup>-6</sup>	with weights of classE <sub>2</sub>
	> 10 kg up to 80 kg		5 - 10 <sup>-6</sup>	
<b>Torque</b> *) Calibration equipment	0,2 N-m to 1000 N-m	DKD-R 10-8:2020	2 - 10 <sup>-3</sup>	
Torque - transducer, - sensors, - measuring chains	0.4 N-m to < 10 N-m	DIN 51309:20225		
	10 N-m to 5000 N-m		0,5 - 10 <sup>-3</sup>	
Hand operated torque screwing tools	10 N-m to 1000 N-m	DIN EN ISO 6789- 2:2017	1 %	
<b>Print</b> *) Negative and positive overpressure $p_e$	-1 bar to -0.03 bar	DKD-R 6-1: 2014	12 $\mu$ bar 5.2 <sup>-10<sup>-5</sup></sup> - $p_e$	$p_e$ = measured value pressure medium: Gas
	> -0.03 bar to < -1 mbar		50 $\mu$ bar <sup>1-10<sup>-4</sup></sup> - $p_e$	
	-1 mbar to < 0 mbar		4 $\mu$ bar	
	0 bar		2 $\mu$ bar	
	> 0 mbar to < 0.2 mbar		4 $\mu$ bar	
	0.2 mbar to 3.6 mbar		0.7 $\mu$ bar	
	> 3.6 mbar to 0.2 bar		2 $\mu$ bar <sup>1-10<sup>-4</sup></sup> - $p_e$	
	> 0.2 bar to 2 bar		30 $\mu$ bar 2.7 <sup>-10<sup>-5</sup></sup> - $p_e$	
	> 2 bar to 20 bar		75 $\mu$ bar 2.7 <sup>-10<sup>-5</sup></sup> - $p_e$	
	> 20 bar to 100 bar		0.5 mbar 3.0 <sup>-10<sup>-5</sup></sup> - $p_e$	
	> 100 bar to 400 bar		7 mbar 3.5 <sup>-10<sup>-5</sup></sup> - $p_e$	
	0 bar			
	2 bar to 1200 bar		7,2 - 10 <sup>-5</sup> - $p_e$ at least 7.5 mbar	

Annex to the partial accreditation certificate D-K-15070-01-02

On-site calibration - Mechanical measurands

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
Absolute pressure $p_{abs}$	0,01 bar to 2 bar	DKD-R 6-1: 2014  $p_{abs} = p_e p_{amb}$	$17 \mu\text{bar} \cdot 2 \cdot 10^{-5} - p_{abs}$	$p_{abs}$ = measured value pressure medium gas The measurement uncertainty of the residual gas measurement must be taken into account.	
	> 2 bar to 20 bar		$180 \mu\text{bar} \cdot 2 \cdot 10^{-5} - p_{abs}$		
	> 20 bar to 101 bar		$0.5 \text{ mbar} \cdot 3 \cdot 10^{-5} - p_{abs}$	The measurement uncertainty of the barometer must be taken into account	
	> 101 bar to 401 bar		$7 \text{ mbar} \cdot 3 \cdot 10^{-5} - p_{abs}$		
	1 bar			$10^{-5} - p_{abs}$ at least 7.5 mbar	$p_{abs}$ = measured value Print medium: Oil Measurement uncertainty of the barometer must be taken into account
	3 bar to 1201 bar				

**Annex to the partial accreditation certificate D-K-15070-01-02**

**On-site calibration - flow measured variables**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Flow rate of liquids</b> Volume flow $dV/dt$ of flowing liquids	0.01 L/min to 250 L/min	3-APD-0-0251-EN: 2023-01  Volumetric comparison measurement (references: Turbines, gear counters, Coriolis; MID,...)	0,15 %	Measuring instruments with analog output, frequency output and visual display
	Mass flow rate $dm/dt$ of flowing liquids	Density from 700 kg/m <sup>3</sup> to 1100 kg/m <sup>3</sup>  Viscosity from 0.8 mm <sup>2</sup> /s to 1600 mm <sup>2</sup> /s	0,20 %	
<b>Gas flow rate</b> Volume flow $dV/dt$ of flowing gases	0.001 L/min to 250 L/min	3-APD-0-0251-EN: 2023-01  Volumetric comparison measurement (references: Coriolis, piston calibrator,...)	0,50 %	Measuring instruments with analog output, frequency output, visual display  in the standard state $p_N=1013.25$ mbar $T_N=0$ °C
	Mass flow rate $dm/dt$ of flowing gases	1.3 mg/min up to 312 g/min  Calibration medium: Compressed air (with compressed air qualities according to DIN ISO 8573-1; clean and oil-free air with max. 55 % rH) at room temperature up to max. 10 bar overpressure	0,50 %	
Volume flow $dV/dt$ from streaming Gases	5 mL/min up to 250 L/min	3-APD-0-0251-EN: 2023-01  Volumetric comparison measurement (References: Coriolis, piston calibrator, ...)  Calibration medium: $N_2, CO_2, Ar, CH_4, He$ (purity > 99.99 % by volume ); Room temperature up to max. 10 bar overpressure	0,50 %	Measuring instruments with analog output, frequency output, visual display  In the standard state: $p_N =$ 1013.25 mbar $T_N =$ 0 °C and $\rho_{N_2} =$ 1.250 kg/m <sup>3</sup> , $\rho_{NCO_2} =$ 1.976 kg/m <sup>3</sup> , $\rho_{NAr} =$ 1.783 kg/m <sup>3</sup> , $\rho_{NCH_4} =$ 0.717 kg/m <sup>3</sup> $\rho_{NHe} = 0.1785$ kg/m <sup>3</sup>

## Mobile laboratory

### Mobile laboratory - Thermodynamic measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature measurement variables</b> Temperature simulators for Resistance-thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display devices for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for Precious metal-thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature-display devices and simulators for non-precious metal thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	

**Mobile laboratory - Mechanical measurands**

Calibration and Measurement Capabilities (CMC)					
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
<b>Print *)</b> Negative and positive overpressure $p_e$	-1 bar to -0.03 bar	DKD-R 6-1: 2014	$12 \mu\text{bar} \quad 5 \cdot 10^{-5} - p_e$	$p_e$ = measured value pressure medium: Gas	
	> -0.03 bar to < -1 mbar		$50 \mu\text{bar}^{1-10^{-4}} - p_e$		
	-1 mbar to < 0 mbar		4 $\mu\text{bar}$		
	0 bar		2 $\mu\text{bar}$		
	> 0 mbar to < 0.2 mbar		4 $\mu\text{bar}$		
	0.2 mbar to 3.6 mbar		1 $\mu\text{bar}$		
	> 3.6 mbar to 0.2 bar		$2 \mu\text{bar}^{1-10^{-4}} - p_e$		
	> 0.2 bar to 2 bar		$30 \mu\text{bar} \quad 2 \cdot 10^{-5} - p_e$		
	> 2 bar to 20 bar		$75 \mu\text{bar} \quad 2 \cdot 10^{-5} - p_e$		
	> 20 bar to 100 bar		$0.5 \text{ mbar} \quad 3 \cdot 10^{-5} - p_e$		
	> 100 bar to 400 bar		$7 \text{ mbar} \quad 3 \cdot 10^{-5} - p_e$		
	0 bar		$7,2 \cdot 10^{-5} - p_e$ at least 7.5 mbar		Print medium: Oil
	2 bar to 1200 bar				
	Absolute pressure $p_{\text{abs}}$		0.01 bar to 2 bar	DKD-R 6-1: 2014  $p_{\text{abs}} = p_e \quad p_{\text{amb}}$	$17 \mu\text{bar} \quad 2 \cdot 10^{-5} - p_{\text{abs}}$
> 2 bar to 20 bar		$180 \mu\text{bar} \quad 2 \cdot 10^{-5} - p_{\text{abs}}$			
> 20 bar to 101 bar		$0.5 \text{ mbar} \quad 3 \cdot 10^{-5} - p_{\text{abs}}$			
> 101 bar to 401 bar		$7 \text{ mbar} \quad 3 \cdot 10^{-5} - p_{\text{abs}}$			
1 bar		$7 \cdot 10^{-5} - p_{\text{abs}}$ at least 7.5 mbar	$p_{\text{abs}}$ = measured value pressure medium: Oil Measurement uncertainty of the barometer must be taken into account		
3 bar to 1201 bar					

**Location Calibration laboratory Kirchzarten, Erich-Rieder Straße 4, 79199 Kirchzarten**
**Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature measurement variables</b> Temperature simulators for Resistance-thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display-devices for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for  Precious metal-thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature indicators and simulators for Base metal-thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	

**Permanent laboratory Kirchzarten, Erich-Rieder Straße 4 - Mechanical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Torque *)</b> manually operated Torque Screwdriving tools	0.01 N-m to 0.2 Nm	DIN EN ISO 6789-2:2017  DIN EN ISO 6789-2:2017	1 - 10 <sup>-2</sup>	
indicating rotary moment wrenches Calibra- for torque tools	> 0.2 N-m to 1 kNm	DKD-R 3-7:2018  DKD-R 10-8:2020	2 - 10 <sup>-3</sup>	
Torque - transducer, - sensors, - measuring chains	10 N-m to 5 kNm	DIN 51309:2022	0,5 - 10 <sup>-3</sup>	
<b>Force *)</b> Force gauges, -transducer	10 N to 250 kN	DKD-R 3-3:2018 DIN EN ISO 376:2011	5 - 10 <sup>-4</sup>	



**Calibration Laboratory Munich, Nikolaus-Otto-Strasse 2, 85221 Dachau**

**Permanent Laboratory Munich - Mechanical measurands**

Calibration and Measurement Capabilities (CMC)

Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks	
<b>Acceleration *)</b> Vibration transducer, vibration measuring device	0.1 m/s <sup>2</sup> to 20 m/s <sup>2</sup>	Sinus excitation DKD-R 3-1 Sheet 3:2020		Complex transfer coefficient (magnitude / phase). Transducer mass up to 0.9 kg, displacement amplitude up to 100 mm	
		0.2 Hz to < 0.4 Hz	2,5 % / 1,6 °		
		0.4 Hz to < 1 Hz	1,5 % / 1,6 °		
		1 Hz to < 16 Hz	0,8 % / 0,8 °		
		16 Hz	0,55 % / 0,6 °		
		> 16 Hz to 63 Hz	0,8 % / 0,8 °		
	> 63 Hz to 160 Hz	1,0 % / 1,1 °			
	1 m/s <sup>2</sup> to 200 m/s <sup>2</sup>	Sinus excitation DKD-R 3-1 Sheet 3:2020			Complex transfer coefficient (magnitude / phase). Transducer mass up to 0.2 kg, displacement amplitude up to 8 mm
		5 Hz to < 10 Hz	1,5 % / 1,5 °		
		10 Hz to < 20 Hz	0,8 % / 0,8 °		
		20 Hz to 1 kHz	0,6 % / 0,6 °		
		> 1 kHz to 5 kHz	0,8 % / 0,8 °		
		> 5 kHz to 10 kHz	2,0 % / 1,5 °		
		> 10 kHz to 15 kHz	2,5 % / 2,5 °		
> 15 kHz to 20 kHz	3,0 % / 3,0 °				
Vibration calibrator	1 m/s <sup>2</sup> to 100 m/s <sup>2</sup>	Sine excitation DIN ISO 16063- 44:2019			
		10 Hz to < 20 Hz	0,8 %		
		20 Hz to 1 kHz	0,6 %		
		> 1 kHz to 5 kHz	0,8 %		
		> 5 kHz to 10 kHz	2,0 %		
<b>Torque *)</b> hand operated torque screwing tools	0,2 N-m to 1000 Nm	DIN EN ISO 6789- 2:2017	2 - 10 <sup>-3</sup>		
		indicating rotary moment wrenches			DKD-R 3-7:2018
		Calibration equipment for turning moment tools			DKD-R 10-8:2020
		Torque transducers, torque meters			DIN 51309:2005
<b>Force *)</b> Tensile force, compressive force, force gauges, force transducers	10 N to 100 kN	DKD-R 3-3:2018	1 - 10 <sup>-3</sup>		

Annex to the partial accreditation certificate D-K-15070-01-02

**Permanent Laboratory Munich - Mechanical measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Print *)</b> Positive and negative overpressure $p_e$	-1 bar to < -0.1 bar	DKD-R 6-1:2014	$0.2 \text{ mbar } 1^{-10^{-4}} - p_e$	$p_e =$ measured value pressure medium gas
	-0.1 bar to < 0 bar		$50 \mu\text{bar } 1^{-10^{-4}} - p_e$	
	0 bar		10 $\mu\text{bar}$	
	> 0 bar to 0.1 bar		$50 \mu\text{bar } 1^{-10^{-4}} - p_e$	
	> 0.1 bar to 10 bar		$0.2 \text{ mbar } 1^{-10^{-4}} - p_e$	
	> 10 bar to 250 bar		$2^{-10^{-4}} - p_e$	
Absolute pressure $p_{abs}$	0.03 bar to 10 bar	DKD-R 6-1:2014 $p_{abs} = p_e p_{amb}$	$0.2 \text{ mbar } 1^{-10^{-4}} - p_{abs}$	$p_{abs} =$ measured value pressure medium: Gas The measurement uncertainty of the barometer must be taken into account
	> 10 bar to 251 bar		$2^{-10^{-4}} - p_{abs}$	

**Permanent Laboratory Munich - Flow measurement variables**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Flow rate of liquids</b> Volume flow rate $dV/dt$ of flowing liquids Mass flow rate $dm/dt$ of flowing Liquids	1,2 m <sup>3</sup> /h to 340 m <sup>3</sup> /h	Comparative measurement with reference flow meters  Calibration medium: Water  3-APD-0-0171-EN: 2023-06	0,1 %	
	1200 kg/h to 3.4 - 10 <sup>5</sup> kg/h			
Volume flow $dV/dt$ of flowing Liquids	0.1 mL/min to 15 mL/min	Volumetric measurement Piston calibrator  Liquids with a Density from 700kg/m <sup>3</sup> to 1100kg/m <sup>3</sup>	0,08 %	Measuring instruments with analog output, frequency output, visual display
	0.8 mL/min to 40 L/min			
	10 mL/min to 300 L/min			
Mass flow rate $dm/dt$ of flowing liquids	1 L/min to 1200 L/min	Viscosity from 0.3 mm <sup>2</sup> /s to 1600 mm <sup>2</sup> /s  3-APD-0-0090-EN: 2023-05	0,05 %	
	0.1 g/min to 15 kg/min			
	0.6 g/min to 32 kg/min			
	8 g/min up to 240 kg/min			
	0.8 kg/min to 1000 kg/min		0,12 %	
			0,09 %	

**Permanent Laboratory Munich - Flow measurement variables**

Measurand/ calibration item	Calibration and Measurement Capabilities (CMC)			
	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Volume $V$ from streaming Liquids	25 mL to 2,5 L	3-APD-0-0090-EN: 2023-05 Flow rates not less than 0.5 mL/min	0,08 %	Measuring instruments with analog output, frequency output, visual display
	190 mL to 19 L	3-APD-0-0090-EN: 2023-05 Flow rates not less than 1 mL/min		
	410 mL to 41 L	3-APD-0-0090-EN: 2023-05 Flow rates not less than 10 mL/min		
<b>Gas flow rate</b> Volume flow $dV/dt$ of flowing Gases	1 mL/min to < 3mL/min	3-APD-0-0088-EN: 2023-01 Laminar flow elements dry air (dew point < -15°C)	0,40 %	Measuring instruments with analog output, frequency output, visual display  In normal condition $p_N$ =1013.25 mbar $T_N = 0$ °C
	3 mL/min to 1000 L/min		0,33 %	
	8 L/min to 15000 L/min	3-APD-0-0088-EN: 2023-01 Critical nozzles dry air (dew point < -15°C)	0,24 %	
Volume flow rate $dV/dt$ of flowing gases	5 mL/min to 250 L/min	3-APD-0-0251-EN: 2023-01 Volumetric comparison measurement (References: Coriolis, piston calibrator, ...) Calibration medium: $N_2, CO_2, Ar, CH_4, He$ (purity > 99.99 vol. %); Room temperature up to max. 10 bar overpressure	0,50 %	Measuring instruments with analog output, frequency output, visual display  In the standard state: $p_N =$ 1013.25 mbar $T_N$ = 0 °C and $\rho_{NN_2} = 1.250 \text{ kg/m}^3$ $\rho_{NCO_2} =$ 1.976 $\text{kg/m}^3$ , $\rho_{NAr} =$ 1.783 $\text{kg/m}^3$ , $\rho_{NCH_4} =$ 0.717 $\text{kg/m}^3$ $\rho_{NHe} =$ 0.1785 $\text{kg/m}^3$
Mass flow rate $dm/dt$ of flowing gases	1.3 g/min to < 3.9 g/min	3-APD-0-0088-EN: 2023-01 Laminar flow elements  dry air (dew point < -15°C)	0,42 %	
	3.9 g/min to < 1300 g/min		0,36 %	
	10 g/min to 15000 g/min	3-APD-0-0088-EN: 2023-01 Critical nozzles dry air (dew point < -15°C)	0,24 %	

**Permanent Laboratory Munich - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature measurement variables</b> Temperature simulators for Resistance-thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display-devices for Resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for Precious metal thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature indicators and simulators for Base metal-thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	

**Calibration Laboratory Essen, Alte Landstraße 3c, 45329 Essen, Germany**

**Permanent Laboratory Essen - Thermodynamic Measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature measurement variables</b> Temperature simulators for Resistance-thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display devices for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for  Precious metal-thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature-display devices and simulators for non-precious metal thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	

**Calibration Laboratory Hamburg, Meiendorfer Straße 205, 22145 Hamburg, Germany**

**Permanent Laboratory Hamburg - Thermodynamic Measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature measurement variables</b> Temperature simulators for resistive thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display devices for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for precious metal thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature indicators and simulators for non-precious metal thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	

**Calibration laboratory Mörfelden-Walldorf, Kurhessenstraße 11, 64546 Mörfelden-Walldorf**

**Permanent Laboratory Mörfelden-Walldorf - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature measurement variables</b> Resistance thermometers, direct reading thermometers with resistance sensor *)	-40 °C to 200 °C	DKD-R 5-1:2018 in liquid bath	10 mK	Comparison with standard resistance thermometers
Temperature transmitter and data logger with resistance sensor *)	-40 °C to 200 °C	DKD-R 5-1:2018 in liquid bath	15 mK	
direct reading thermometers, temperature transmitters and data loggers with resistance sensor *)	-40 °C to < 0 °C	DKD-R 5-1:2018 in the climatic chamber	0,30 K	
	0 °C to 50 °C		0,15 K	
	> 50 °C to 80 °C		0,25 K	
	> 80 °C to 120 °C		0,40 K	
	> 120 °C to 180 °C		0,90 K	
Precious metal- Thermocouples, direct reading Thermometer with Precious metal- Thermocouple sensor *)	-40 °C to 200 °C	DKD-R 5-3:2018 in liquid bath	0,5 K	Comparison with normal Resistance-thermometers
Base metal- Thermocouples, direct reading Thermometer with Base metal- Thermocouple sensor *)	-40 °C to 200 °C	DKD-R 5-3:2018 in liquid bath	0,2 K	
Temperature transmitter and data logger with thermocouple sensor *)	-40 °C to 200 °C	DKD-R 5-3:2018 in liquid bath	0,3 K	Comparison with standard resistance thermometers
Circulating thermostats, precision baths	-40 °C to 200 °C	3-APD-0-0155-EN: 2023-01	10 mK	Comparison with standard resistance thermometers
Temperature block calibrators *)	-40 °C to 150 °C	DKD-R 5-4:2018	0,05 K	Comparison with standard resistance thermometers

**Permanent Laboratory Mörfelden-Walldorf - Thermodynamic measurands**

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
Radiation thermometer	-18 °C to 60 °C	3-APD-0-0018-EN: 2023-01 Spectral range: 8 µm to 14 µm	0,6 K	Calibration with liquid flushed cavity radiator
	> 60 °C to 100 °C		0,9 K	
	> 100 °C to 350 °C		1,2 K	
Temperature simulators for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display- devices for resistance thermometers *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for  Precious metal- thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584- 1:2014
Temperature- display devices and simulators for non-precious metal thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	
<b>Frost / Dew point temperature</b> Dew point mirror, -transmitter, -hygrometer	-32 °C to -25 °C	3-APD-0-0036-EN: 2023-01 1-temperature 2-pressure humidity generator	90 mK	
	> -25 °C to -10 °C		60 mK	
	> -10 °C to 40 °C		90 mK	
	> 40 °C to 70 °C		0,12 K	
<b>Relative humidity</b>  Hygrometer,  Transmitter	10 % to 20 %	DKD-R 5-8:2019 1-Temp.- 2-Pressure Humidity generator -10 °C to < 0°C	0,7 %	Measurement uncertainty as Absolute value of the relative humidity
	> 20 % to 40 %		1,3 %	
	> 40 % to 85 %		2,1 %	
	10 % to 40 %	DKD-R 5-8:2019 1-temp. 2-pressure humidity generator 0 °C to 20 °C	0,3 %	
	> 40 % to 80 %		0,6 %	
	> 80 % to 95 %		0,7 %	
	10 % to 40 %	DKD-R 5-8:2019 1-Temp.- 2-Pressure Humidity generator > 20 °C to 70 °C	0,3 %	
	> 40 % to 80 %		0,5 %	
> 80 % to 95 %		0,6 %		

## Calibration Laboratory Winsen, Tönnhäuser Weg 100-106, 21423 Winsen (Luhe)

### Permanent Laboratory Winsen - Thermodynamic measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Temperature measurement variables</b> Temperature simulators for Resistance-thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,016 K	Characteristic according to DIN EN 60751:2009
Temperature display devices for resistance thermometer *)	-200 °C to 850 °C	DKD-R 5-5:2018	0,03 K	
Temperature indicators and simulators for  Precious metal-thermocouples *)	-200 °C to 1750 °C	DKD-R 5-5: 2018	0,1 K	Characteristic according to DIN EN 60584-1:2014
Temperature-display devices and simulators for non-precious metal thermocouples *)	-200 °C to 1300 °C	DKD-R 5-5: 2018	0,05 K	

### Permanent Laboratory Winsen - Mechanical measurands

Calibration and Measurement Capabilities (CMC)				
Measurand/ calibration item	Measuring range	Measuring conditions / Procedures	Extended uncertainty of measurement	Remarks
<b>Print *)</b>  Positive and negative Overpressure $p_e$	-1 bar to 10 bar  > 10 bar to 250 bar	DKD-R 6-1:2014	0.2 mbar $1 \cdot 10^{-4} - p_e$	$p_e$ = measured value Pressure medium gas
Absolute pressure $p_{abs}$	0,03 bar to 11 bar  > 11 bar to 251 bar		DKD-R 6-1:2014 $p_{abs} = p_e p_{amb}$	



**Abbreviations used**

APD	In-house developed calibration procedure by Testo Industrial Services GmbH
CMC	Calibration and measurement capabilities
DIN	German Institute for Standardization R.A.
DKD-R	Guideline of the German Calibration Service (DKD), published by the Federal Physical-Technical Institute (PTB)
EURAMET	European Association of National Metrology Institutes
G-ITS-90, Part 2.2	BIPM-Guide to the Realization of the ITS-90, Triple Point of Water
G-ITS-90, Part 2.3	BIPM-Guide to the Realization of the ITS-90, Cryogenic Fixed Points
G-ITS-90, Part 2.4	BIPM-Guide to the Realization of the ITS-90, Metal Fixed Points for Contact Thermometry
G-ITS-90, Part 5	BIPM-Guide to the Realization of the ITS-90, Platinum Resistance Thermometry